A-A-54081
15 February 1990

COMMERCIAL ITEM DESCRIPTION

SYRINGE, IRRIGATING, DENTAL, Sterile (Disposable)

The General Services Administration has authorized the use of this Commerical Item Description as a replacement for Military Specification MIL-S-42050, which is cancelled.

This commercial item description covers a sterile, plastic and disposable endodontic syringe suitable for flushing, irrigating and sterilizing root canals with medicated solutions.

Salient characteristics:

Material: All material shall comply with the requirements of the Federal Food, Drug, and Cosmetic Act. The portions of the unit, that comes in contact with the fluid to be injected, shall be compatible with all known endodontic medications. Also, the plastic components of the syringe shall comply with the U.S.P. requirements for Class III plastic containers.

Barrel. The syringe barrels shall be fabricated from a rigid, transparent, virgin, nontoxic, high-grade polypropylene plastic. Selected regrind free from graduation filler and any other adulterants may be used. The barrel shall be considered to be transparent if the scale can be seen in daylight, by the naked eye, when viewed through the empty barrel from the side opposite to the scale.

Alcohol resistance of graduation marks. The graduated markings shall withstand five vigorous rubbing cycles of a 70% isopropyl alcohol moistened cotton swab. A cycle shall be one up and one down stroke along the length of the barrel. The markings shall remain legible.

Plunger. The plunger rod shall be fabricated from a rigid polypropylene or polystyrene plastic.

Plunger tip. The plunger tip shall be fabricated from fresh stock, pharmaceutical grade, nonblooming, nonleaching, nontoxic, age-resistant, inert, natural rubber or neoprene.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Inertness to medication. The plunger tip shall show no evidence of a reaction with a high pH buffer solution when tested as specified herein.

Needle. The needle shall consist of a cannula, a plastic hub and adhesive.

Cannula. The cannula shall be fabricated from either seamless drawn or welded and drawn, Class 3, type 304 austenitic stainless steel, as specified in ASTM Designation F899, for composition only.

Hub. The hub shall be fabricated from a suitable plastic compound containing no elutable toxic materials. The hub shall be compatible with all known endodontic medications.

Design, dimensions and construction. Design, design and construction shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24513. The syringe shall be a sterile, disposable endodontic syringe. The syringe shall consist of a plunger, barrel, and an endodontic needle (23 B. W. Gage), suitable for introducing endodontic medicated solutions to flush, irrigate, and sterilize root canals.

Barrel. The syringe barrel shall consist of a tube closed at one end, except for an orfice, and open at the other end to receive the plunger. At the closed end, there shall be formed a tapered tip, which shall have a 0.060 inch taper and Luer-lock skirt, which shall be in accordance with Figure 7 of ANSI/HIMA MD70.1, to receive a standard endodontic needle. The tip shall be formed concentric to the axis of the barrel.

The wall at the open end of the barrel shall extend radially outward to form a flange of suitable size and shape to provide finger grip and anti-roll features. The inside of the barrel shall have a very fine, uniform surface, substantially free from pits, bubbles, mold marks, flash, or other defects that would affect the fluid tightness and smoothness of its fit with the plunger. The inside of the barrel shall be lightly coated with a pharmaceutical grade silicone lubricant.

The inside lubricant shall be such as to provide an accurate fit with the plunger tip, resulting in a fluid tight fit, and yet permitting the plunger to move freely throughout its excursion in the barrel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Ungraded portion of the barrel, at the proximal end, shall be of sufficient length to permit the plunger tip to be aligned with the top graduation mark, while the barrel is encompassing the entire rubber tip, but shall not be greater than one-third of the overall length of the syringe barrel.

Graduation and other markings. The barrel shall be calibrated to 6 ml (cc) with main graduation at each 1 ml (cc), subgraduations at each 0.2 ml (cc) and numbered at each 1 ml (cc). The outside surface of the barrel shall be marked with a graduation scale and the name or registered trademark of the manufacturer. The graduation scale shall be applied to the barrel in such a manner as to be approximately in line with the flattened side of the flange. The zero (0) mark of the scale shall register precisely with the distal ring of the plunger tip, when the plunger is pushed to the bottom of the barrel. The scale and numbering shall be in cubic centimeters (cc). Main graduation marks shall be at least 33-1/3% longer than subgraduation marks. No number is required at the zero (0) mark. The markings shall be legible and of a dark, contrasting color, easily readable.

Plunger. The plunger shall be either hollow or solid. The distal end of the plunger shall be shaped to permit attachment of a plunger tip. The proximal end of the plunger shall extend well out of the barrel, and shall be shaped to form a knob of thickness and diameter appropriate to the size of the syringe, and shaped to allow a suitable finger grip for usage. The distal end of each plunger with its attached plunger tip shall accurately fit its individual barrel so as to make the syringe successfully withstand the fluid tightness requirements, as specified herein, permit the plunger and the plunger tip to draw liquids, and yet move freely and smoothly throughout its excursion in the barrel. The plunger tip shall be securely attached to the plunger rod so as not to become loose or detached when in use. The tip shall be shaped to closely fit the distal end of the barrel so as to remove all fluid from the inside of the barrel when the plunger is pushed all the way in. In addition, the tip shall be designed, e.g., beveled, in order to permit easy and efficient removal of air bubbles from the fluid path. The plunger tip shall be treated with a pharmaceutical grade silicone lubricant. An excess of lubricant that detracts from appearance, interferes with visibility of plunger tip ring, or can be cause to form pools or droplets, or can be washed out with the solution being used, shall be cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Graduation markings, denoting correct, accurate capacity, shall register precisely with the first ring at the plunger tip distal end, touching the sides of the barrel. The tip shall be of a permanent, suitable color so that the first ring at the plunger tip distal end, touching the sides of the barrel, shall be clearly visible and shall accurately indicate through the barrel wall. Each plunger shall have a suitable solid thumbrest at its proximal end. The thumbrest shall be perpendicular to the centerline of the plunger and may be flat or slightly curved. As an alternate, the thumbrest may have a center hole which shall not exceed 3/8 inch in diameter.

Plunger tip resiliency. The rubber plunger tip, when tested as specified herein, shall not crack, crumble or become fissured. Furthermore, after one-minute recovery time, the plunger tip shall again be be circular in appearance. An obviously oval shape indicating a lack of adequate recovery is cause for rejection.

Needle.

Cannula. The cannula shall be a straight tube with a special notch at the distal end. The point shall be ground flat. Both outer and inner edges of the point shall be free from burrs and slivers. The proximal end of the cannula shall extend well into the plastic hub in such a manner as to provide a straight continuous fluid path from the hub cavity. The cannula shall not deviate from the central longitudinal axis of the needle by more than 3 degrees. The exterior of the cannula shall be clean, smooth and free from pits, deep furrows and die marks.

Corrosion resistance of cannula. When tested as specified herein, the cannula shall show no appreciable corrosion.

Hub. Shall be a Luer-Lock hub to provide means for facilitating attachment to, and removal from, a syringe so designed and constructed to form a leak-proof connection with a Luer-Lock connection constructed in accordance with Figure 7 of ANSI/HIMA MD 70.1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

The hubs shall be tested for diameter and depth of the hub cavity as specified herein. Hubs which do not fit the plug gage within the "GO" - "NO GO" limits as specified herein, shall be considered not acceptable. The fit of the Luer taper of the plastic hubs on the plug gage need not be tested with the feeler gage present. When the hub is fitted onto the plug gage as specified, there shall be no movement of the hub on the gage when the hub is pressed lightly from side to side in all directions. The hub and cannula shall be securely assembled.

Accuracy. When tested as specified herein, syringes shall deliver plus or minus 5% of the capacity indicated at each graduation mark tested.

Fluid tightness. The syringe shall be tested for fluid tightness between the plunger and barrel, and between the barrel tip and needle hub as, specified below.

Suction. There shall be no air leakage as revealed by the suction test as specified herein.

Plunger fit. The plunger shall make a snug fit in the barrel to provide leak-tight operation. At no part of the travel of the plunger shall the plunger fit be so loose as to permit movement of the plunger by its own weight when tested as specified herein.

Stiffness. The maximum deflection of the cannula shall not exceed 0.033 inch, as specified herein.

Security to breakage. There shall be no breakage of the needle as specified herein.

Sterility. The syringe and needle shall be tested for sterility in accordance with the method specified in the U.S.P. or A.A.M.I. guidelines.

Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile syringes, as applicable, from the containers and preparing them for use. In addition, information as to methods for destroying syringes, as applicable, to prevent their re-use shall also be shown.

Workmanship. The syringe shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing fifty syringes, as specified, constitutes one unit.

Unit package. Each syringe shall be packaged in a sealed peel-open type container or in a rigid plastic tamper proof container capable of protecting the needle cannula and maintaining sterility of the contents, unless the package is damaged or opened. The container shall be so designed to facilitate dispensing of the syringe by sterile techniques.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the plastic and rubber, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination and test characteristics. The unit of product for inspection, examination and test purposes shall be one syringe (end item). Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-2, AQL 1.0 (percent defective), except sterilization test acceptance number shall be zero (0).

Examination. The syringes shall be examined to determine compliance with all requirements contained in this specification. Examination, however, shall not be restricted to defects listed below. Any syringes in the sample containing one or more defects shall be rejected and, if the number of defective syringes in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Defects

Needle missing or dislodged.
Cannula lumen not free of burrs, slivers, foreign matter or obstructions.
Cannula exterior not free from foreign matter, pits, tool marks, burrs, scale or corrosion and not smoothly finished.
Cannula not straight or free of cracks.
Point blunt or bent, 1/ not sharp, not symmetrical or properly formed, or not free of burrs, feathered edges, or other detrimental defects (examined under 7X magnification).
Plunger rod, plunger tip, or both missing.
Plunger or plunger tip tip frozen in barrel.
Inside of barrel not coated or contains excessive lubricant.
Barrel tip orfice missing or blocked.
Barrel not clear, plunger tip not clearly visible making accurate alignment with graduation markings difficult or impossible.
Barrel (including tip with orfice) not free from pits, bubbles, cracks, breaks, crazing, tackiness, or loose or detachable flash.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Inside surface of barrel not fine uniform surface, free from pits, bubbles, mold marks, flash, or other defects that would affect fluid tightness and smoothness of its fit with the plunger.
Plunger tip not smooth, free from flash, bubbles, breaks, cracks, or other defects affecting its fit with the plunger.
Plunger tip detaches from plunger rod when pushed back and forth in syringe barrel.
Plunger does not operate smoothly in the barrel.
Plunger not free from cracks, breaks, crazing or tackiness.
Bore of syringe tip not free from flash which obstructs bore of tip below the minimum area.
Deep scratches (not interfering with reading of dosage) on outside surface of barrel (except on tapered portion of the tip).
Discoloration of outside surface of the hub.
Manufacturer's name or registered trademark missing, incomplete or not legible.
Graduation scale not approximately in line with flattened side of flange.
Instructions missing, incomplete or not legible.

1/ A bent point is defined as any deviation of the point from the longitudinal axis of the needle when viewed at 7X magnification from the side of the needle and in a line parallel to the plane of the bevel.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Diameter and depth of hub cavity. Hub cavity shall be tested on the female Luer taper gage in accordance with plug gage of Figure 5 of ANSI/HIMA MD70.1. (This gage shall be as shown in Figure 5, except that it shall be constructed with a suitable fluid path to permit performance of the fluid tightness test as specified). The hub shall be slipped onto the gage with very light pressure. The fit of the hub on the gage shall be tested by means of a feeler gage in accordance with Figure 6 and 8 of ANSI/HIMA MD70.1. The fit of the Luer taper, of the plastic hubs on the plug gage need not be tested with the feeler gage present. All hubs shall provide a firm, non-wobble fit when testing the Luer taper of the hubs on the plug gage. The flange on the Luer hub shall be properly recessed from the proximal edge and of dimensions to provide a full, firm, leak-proof fit of the hub and barrel Luer tip (see Figure 3 of ANSI/HIMA MD70.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Accuracy test. Syringes shall be tested for accuracy of delivery at each main graduation mark using water and a calibrated burette, or as an alternate, gravimetrically taking into account the temperature and density of water. The first ring from the distal end of the plunger tip shall be aligned with each graduation mark tested. One-cc syringes shall not be tested for accuracy at the first 15% of the scale.

Chemical analysis. Cannula material shall be subject to chemical analysis in accordance with method 111 or 112 of FED-STD-151.

Corrosion-resistance of cannula. The cannula shall be immersed in 10 percent citric acid at room temperature for 5 hours and, after removal from the acid, shall be boiled in distilled water for 30 minutes, and then shall remain immersed for 48 hours in the distilled water. At the end of the 48-hour period, the cannula shall be allowed to dry by evaporation, and shall then be examined for evidence of corrosion. A glass vessel shall be used throughout this test.

Plunger tip inertness. The rubber plunger tip shall be removed from the plunger rod, wiped with a clean, dry cloth to remove silicone lubricant, and immersed in a test tube under a volume of pH 10.5 buffer solution equal to the capacity of the syringe. The exposure shall be for one hour at 68^0 to 77^0 F, (20^0 to 25^0 C). There shall be no color or any other visible change of the buffer solution. A control shall be used during the exposure consisting of the same amount of buffer solution in a test tube which does not contain a plunger tip.

Fluid tightness test. The fluid tightness shall consist of the pressure test and the suction test described in the following paragraphs:

Pressure test. The water pressure test system shall have a water pressure gage and a medication injection plug. The needle of the syringe shall be inserted into the medication injection plug and the water pressure brought up to 45 psig. The plunger shall then be withdrawn to the 50% point of the graduated capacity where it shall be allowed to remain for 10 seconds. The plunger shall then be withdrawn to the 75% point and held there for another 10 seconds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

The plunger shall then be withdrawn to the 100% point for a final 10 seconds. During each 10-second period there shall not be enough water to pass the connection between the needle hub and Luer tip of the barrel or needle-barrel junction and between the plunger tip and barrel of the syringe to produce a falling drop. After each successive test, any water that may have passed shall be wiped with a dry cloth and the next test applied. The needle of the syringe shall be withdrawn from the medication injection plug and the water ejected at the rate of 1/2 cc/second until the plunger reaches the zero mark. The needle shall not leak or separate at its junction with the barrel.

Suction test. An amount of water equal to more than one-half the capacity shall be drawn into the syringe through the endodontic needle. The syringe shall then be held vertically, needle up, to allow air to rise above the water and be expelled. Tapping the barrel of the syringe will facilitate removal of trapped air. The plunger shall be pressed inward to the 50% mark. A soft rubber stopper shall then be used to seal the needle. The plunger shall then be carefully withdrawn, at a rate of about one inch/10 seconds, to the full mark. After the plunger has reached the full mark, there shall be no more than three bubbles coming by the distal ring for at least 10 seconds. Should there be more than three air bubbles passing the distal ring during the 10-second period, there shall be an additional period of observation of 15 seconds during which there shall be no bubbles passing the distal ring. Upon release of the plunger, it shall return to the 50% mark.

Air bubbles which appear on the distal surface of the plunger tip during the withdrawal of the plunger shall be disregarded. Some air bubbles frequently appear under conditions of reduced pressure because of phenomena such as decreased solubility of air in water, interfacial tension, and aggregation of minute bubbles into larger ones.

Plunger tip resiliency. The unmounted rubber plunger tip shall be free of visible lubricant. If necessary wipe dry with a cloth. The outside diameter shall be measured to the nearest 1/16 inch. The plunger tip shall then be compressed laterally to 7/16 to 1/2 of the outside diameter for 15 seconds. This shall be done in a vise the jaws of which have smooth faces. The compressed plunger tip shall be observed for cracks, crumbling or fissures. The plunger tip is removed from the vise and allowed to recover for 60 seconds after which it is examined for circularity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Plunger fit. The syringe shall be held vertically with the Luer-tip exposed. The plunger shall be worked back and forth twice, for the full length of the barrel. The plunger shall be withdrawn from zero, in increments of 1/10 the capacity, and observed for 10 seconds at each level. The Luer tip can be up or down in this test. Movement of the plunger by its own weight is cause for rejection. The same test shall be used for syringes with permanently attached needles except that 20 seconds observation time shall be used.

Resistance to breakage test. The needle hub shall be held in a fixed position, and a force shall be applied to the cannula, at a distance from the hub of 5/8 inch, in order to bend the cannula in a clockwise direction through an arc of 25 degrees from the original or neutral position. The operation shall then be reversed (i.e., bent through an arc of 25 degrees from the neutral position in a counter-clockwise position. This complete operation (or cycle) shall be repeated 20 times.

Security of staking. A straight pull of 5-pounds shall be applied between the cannula and hub for a period of 1-minute.

Stiffness test. The cannula shall be placed across two test edges (0.002 inch radius maximum) separated by a distance of 0.625 plus or minus 0.003 inch. A load of 40 plus or minus 1/4 ounces shall be applied at the center of the span and the resulting maximum deflection of the cannula shall be measured using a gage with graduations of at least 0.001 inch. The force shall be applied through an edge with a maximum radius of 0.036 inch, and centered to within $\pm$ 0.003 inch.

Needles with plastic hubs. Needles furnished with plastic hubs shall be tested for safety as follows:

Cover 20 needles with 30 ml of sterile, pyrogen-free saline test solution (USP) and warm at 37^0 C for 24 hours. Decant the saline and test, undiluted, for both acute toxicity and tissue sensitivity. Determine acute toxicity by injecting 1.0 ml of the eluate intravenously into each of 10 nonfasted, healthy 28 to 32 gram mice. Observe the animals at 4, 24, and 72 hours after injection for any deaths or discernible symptoms of toxicity. Determine tissue reaction by injecting 0.20 ml of the eluate into each of the 10 sites on the back of a rabbit; the injections shall be made in such a manner as to raise a distinct bleb. Observe the injection for any evidence of redness, swelling, or hardness. Trauma shall be ruled out with suitable saline control sites.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-01-036-0723	SYRINGE, IRRIGATING, DENTAL, Disposable, 6 ml, 50s: Sterile.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

48978

PROJECT NO. 6520-2296
LOCATION: NCR - DISK 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in mitting comments and suggestions for improvements. All users of military standardization documents are invited to provide ...gestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54081	Syringe, Irrigating, Dental, Sterile (Disposable)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITIONS ARE OBSOLETE

TO DETACH THIS FORM, CUT ALONG THIS LINE.

Source: https://assist.dla.mil - Downloaded: 2026-04-26T09:28Z
Check the source to verify that this is the current version before use.