A-A-50692
11 October 1990

COMMERCIAL ITEM DESCRIPTION

OSCILLOGRAPH, PORTABLE

The General Services Administration has authorized the use of this Commercial Item Description.

Abstract. This Commercial Item Description (CID), decribes a Oscillograph Recorder. This CID is meant as a minimum requirement for the Oscillograph Recorder in which only those manufacturers that meet or surpass the following requirements are qualified per this CID.

Salient characteristics. The equipment shall be capable of operation and be bid sample tested within the accuracies, limits, and specifications herein.

1. Classfication. Equipment covered by this CID shall be commercially available equipment modified to the extent necessary to meet the following description. The equipment shall be type III, class 5, style E and color R in accordance with MIL-T-28800.

2. Order of precedence. In the event of a conflict between the text of this CID and the references cited herein (except for associated detail specifications, specification sheets or MS standards) the text of this CID shall take precedence. Nothing in this CID, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. Safety and environmental. The equipment shall meet all safety and environmental requirements as specified in MIL-T-28800 for the classification as stated herein.

4. Electrical power sources. The equipment shall operate from nominal commercial, military, and shipboard power sources of 100, 120, 220, 240 volts ac (+10%) at line frequencies of 50 Hz to 60 Hz. Power consumption shall not exceed 180 watts. Transient state conditions shall meet MIL-T-28800 requirements for the classification specified herein.

5. Temperature requirements.

   a. Non-operating temperature. The equipment shall meet its performance and accuracy requirements in an non-operating environment of -10°C to +50°C.

   b. Operating temperature. The equipment shall meet its performance and accuracy requirements in an operating environmental of 0°C to 40°C.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any data which may improve this document should be sent to: Support Division, SA-ALC/MMDEM, Kelly AFB TX 78241-5000.

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:45Z
Check the source to verify that this is the current version before use.

6. Mean Time Between Failures (MTBF). The equipment Mean Time Between Failures (MTBF) shall be 2,000 hours. The lower test MTBF shall be 2,000 hours and the upper test MTBF shall be 4,000 hours.

7. Pre-conditioning requirement. Each deliverable equipment shall be capable of being subjected to the environmental screening effects of burn in and thermal cycling, with thermal cycling being accomplished in the sample testing phase.

8. Calibration and maintenance adjustments. The design of the system shall provide for readily accessible calibration adjustments and maintenance adjustments. These adjustments shall be provided by variable value components which are adjustable by the use of simple means. The calibration by substitution of selected components or parts is unacceptable unless specifically approved. The calibration adjustments, wherever possible, shall be accessible without removal of the instrument case or modules. The calibration interval shall be a period of one year or greater based on an operating time of 2000 hours per year.

9. Performance characteristics.

Number of Channels	2
Channel Width	40 or 50mm
Writing Method	Thermal or Ink
Inputs	Differential and balanced to ground
Input Impedance	≥1Mohm for both differetial and balanced to ground
Input Over Voltage Protection	Capable of withstanding ± 100 V on the mV range ± 500 V on V range
Frequency Response (+5%, -10%)	
40mm pk-pk	dc to at least 35 Hz
20mm pk-pk	dc to at least 50 Hz
10mm pk-pk	dc to at least 70 Hz
Chart Speeds	Shall cover the range of 100mm/sec to 10mm/min
Full Scale Input Ranges	Shall cover the range of 50mV to 200 V
Approximate Dimensions	8.0in. x 8.0in x 17in. or smaller.
Paper Take Up Spool	Stores written paper internally. Provides environmental protection for written paper.
Approximate Weight	28lbs. or less.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:45Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within specified tolerances using conversion tables contained in the latest revision of Federal Standard No 376, and all other requirements of this Commercial Item Description are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements. In accordance with section 23.403 of the Federal Acquisition Regulations, the Government's policy is to acquire items composed of the highest percentage of recovered materials practicable, consistent with maintaining a satisfactory level of competition without adversely affecting performance requirements or exposing the supplier's employees to undue hazards from the recovered materials.

Preservation, packaging, packing, labeling, and marking. The preservation, packaging, packing, labeling, and marking shall be as specified in the contract or order.

Notes.

MIL-T-28800 is available from the Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia, PA 19111-5094.

Review activity:
Air Force - 99

Preparing activity:
Air Force - 82

(Project 6625-F985)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:45Z
Check the source to verify that this is the current version before use.