A-A-30066B
November 27, 1991
SUPERSEDING
A-A-30066A
July 26, 1989

COMMERCIAL ITEM DESCRIPTION

STOCKINGS, ANTI-EMBOLISM

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes an anti-embolism stocking which is worn by patients as an aid in promoting venous return while bedridden.

Salient characteristics. Stocking, Anti-Embolism, white, with closed heel and toe, foot inspection point, graduated compression from ankle to top, compression at ankle of 16mm/hg to 30mm/hg, 1 inch (2.54 cm) cuff at the top.

STYLE A - Knee length

Size 1 - Medium, regular length for use on 12 to 15 inch (30 cm to 38 cm) calf

Size 2 - Large, regular length for use on 15 to 17-1/2 inch (38 cm to 44 cm) calf

STYLE B - Thigh length

Size 1 - Small, regular length for use on calf up to 12 inches (30 cm), 29 to 33 inch (74 cm to 84 cm) length

Size 2 - Medium, short length, for use on 12 to 15 inch (30 cm to 38 cm) calf, up to 29 inch (74 cm) length

Size 3 - Medium, regular length, for use on 12 to 15 inch (30 cm to 38 cm) calf, 29 to 33 inch (74 cm to 84 cm) length

Size 4 - Medium, long length, for use on 12 to 15 inch (30 cm to 38 cm) calf, over 33 inch (84 cm) length

Size 5 - Large, short length, for use on 15 to 17-1/2 inch (38 cm to 44 cm) calf, up to 29 inch (74 cm) length

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Size 6 - Large, regular length, for use on 15 to 17-1/2 inch (38 cm to 44 cm) calf, 29 to 33 inch (74 cm to 84 cm) length

The leg length shall be measured from the gluteal furrow (intra-gluteal crease) to the bottom of the heel.

The stocking shall be designed to provide uniformly graduated compression, decreasing from ankle to knee and from lower thigh to upper thigh. The stocking shall provide pressure relief at the knee. The stockings shall exert proper compression to the legs as prophylaxis for thombo-embolism.

The stocking shall be seamless with a full foot and shall contain an inspection toe opening. As an alternate, a clear-view inspection window may be substituted for the inspection toe opening.

The stocking shall be knit with stretch type nylon yarns using a combination of "S" and "Z" twist torque yarns to produce a balanced fabric. The spandex elastic yarn shall be at least single wrapped with continuous multi-filament nylon stretch yarn.

The color of the stocking shall be white. The top of the stocking shall be color coded to indicate length: white for regular length, yellow for short length, and blue for long length.

Workmanship. The anti-embolism stocking shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Style B - Package (PG). One package containing six pair of anti-embolism stockings of one size only, along with laundry instructions, as specified, constitutes one unit. Packaging for Style A will appear in the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 and 2.5 (percent defective) and an inspection level of S-1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

Style B -	Style A -
Size 1 - 6515-01-229-2640	Size 1 - 6515-00-016-8230
Size 2 - 6515-01-229-2639	Size 2 - 6515-00-016-8229
Size 3 - 6515-01-229-2638	
Size 4 - 6515-01-229-7709	
Size 5 - 6515-00-307-5597	
Size 6 - 6515-00-227-9414	

This document contains the requirements of Interim Commercial Item Description A-A-0051670 (DM).

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type, class, style and size

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

Project No. 6515-5191

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.