A-A-50143
January 6, 1988

COMMERCIAL ITEM DESCRIPTION

SCALE, DIGITAL, PLATFORM

The General Services Administration has authorized the use of this commercial item description.

1. Intended use. This commercial item description covers platform digital scales used for weighing bulk and heavy products.

2. CLASSIFICATION

2.1 Classification. The digital platform scales shall be of the following types and capacities (see 6.1):

Type I - 500 pound capacity, portable
Type II - 5,000 pound capacity
Model 1 Low profile, floor mounted with two access ramps
Model 2 - Floor level pit installed

3. SALIENT CHARACTERISTICS

3.1 Design and construction of type I scale. The type I scale shall be a floor model having a 500 pound capacity. The weighing range shall go from 0 to 500 pounds in 0.05 graduations. The platform, column, and digital indicator shall be constructed of stainless steel, 11 gauge minimum. The digital indicator and weighing load cell shall be waterproof. The digital indicator shall be mounted atop the scale column, be a single display with a minimum five digit readout, and shall have a backup battery pack that shall be capable of operating for 8 hours when power is not available. The scale shall have four casters, two of which shall be equipped with locking mechanisms. The scale shall have the capability to accurately operate in an environment of 14° to 113°F and at 10 to 95 percent relative humidity, noncondensing. The load cell shall be capable of handling one-hundred percent overload without causing damage to the weighing mechanism. The digital indicator keyboard shall have provisions for: automatic zero tracking; tare, recall, and clear tare buttons; and tape printer. The scale shall be furnished with a 12-foot minimum electric cord. The approximate dimensions of the scale shall be: platform - 30 inches long by 24 inches wide by 4 inches high, column - 39 inches high, and digital indicator - 11 inches high by 14 inches wide by 6 inches deep.

FSC 6670

ISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:39Z
Check the source to verify that this is the current version before use.

3.2 Design and construction of type II scale. The type II model 1 scale shall be a low profile, floor mounted model with two access ramps. The type II model 2 scale shall be a floor level pit installed model as specified (see 6.1). The scales shall have a 5,000 pound capacity and the weight range shall go from 0 to 5,000 pounds in 1 pound graduations. The platform and ramps shall be constructed of heavy duty carbon steel and shall have skid-resistant raised steel patterns on the surface. The scales shall be 3 to 6 inches high. The platform shall be available in 3 by 3, 4 by 4, 4 by 6, and 5 by 5 approximate feet dimensions. The type II model 1 scale shall be equipped with two 36-inch long minimum access ramps. The width of the ramps shall be compatible with the specified width of the platform. The load cell shall be capable of handling a fifty percent overload without causing damage to the weighing mechanism. The digital indicator shall be remote from the weighing platform. The digital indicator shall be either a plastic enclosed, dust tight, desk mounted model or a stainless steel enclosed, dust-tight, wall mounted model. The digital indicator shall be a single display, five digit minimum and shall have the same capabilities and keyboard programmability as the type I scale indicator. The scale shall be furnished with a 50-foot electrical cable that shall be connected from the platform to the digital indicator.

3.3 Installation. The installation and electrical tie-in of the type II scale shall be determined by the ordering agency.

3.4 Warranty. The scales shall be guaranteed against manufacturing defects. Operating instructions shall be included with each scale.

3.5 Electrical. The scales shall operate on 120 volts, 60 hertz, single phase power source domestically and 220 volts, 50 hertz, single phase for overseas operations. The power consumption shall not exceed 25 watts.

3.6 Standards compliance. The scales shall comply with the National Bureau of Standards Handbook NBS-H44.

4. QUALITY ASSURANCE

4.1 Certification. The contractor shall certify that the product offered meets the salient characteristics of this description, conforms to the producer's own drawings, specifications, standards, and quality assurance practices, and is the same product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to the first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

5. PACKAGING

5.1 Packaging. The digital scales shall be packaged, packed, and marked in accordance with the domestic or overseas requirements of ASTM D 3951, as applicable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:39Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data.

a. Title, number, and date of this document.
b. Type and model required (see 2.1).
c. Electrical power requirements (see 3.5).
d. Type II platform dimensions (see 3.2).
e. Type II digital indicator model (see 3.2).
f. Type II installation and tie-in requirement (see 3.3).

6.2 Source for nongovernment association documents.

ASTM D 3951 - Standard Practice for Commercial Packaging

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

NBS Handbook 44 - Specifications and Tolerances for Commercial Weighing and Measuring Devices

(Application for copies should be addressed to the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.)

MILITARY INTERESTS:

Custodian

Army - GL

Review Activity

Army - TS

CIVIL AGENCY COORDINATING ACTIVITY:

GSA - FCGC

PREPARING ACTIVITY:

Army - GL

Project No. 6670-0174

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:39Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:39Z
Check the source to verify that this is the current version before use.