MILITARY SPECIFICATION

INDICATORS, TEMPERATURE, ELECTRICAL RESISTANCE
MULTIFUNCTION, -50° TO +200°C

This amendment forms a part of Military Specification MIL-I-27674B(USAF), dated 25 October 1973, and is approved for use by the Department of the Air Force and is available for use by all Departments and Agencies of the Department of Defense.

PAGE 1

Add the following preamble immediately under the title:

"This specification is approved for use by the Department of the Air Force and is available for use by all Departments and Agencies of the Department of Defense."

Add the DD Form 1426 note on the bottom center of the page:

"Beneficial comments (recommendations, additions, deletions) and any pertinent date which may be of use in improving this document should be addressed to: the Engineering Division, San Antonio Air Logistics Center/ MMEDO, Kelly AFB, TX 78241 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter."

PAGE 2

Delete the parenthetical source paragraph at the end of 2.1 and substitute:

"(Copies of specifications standards, handbooks, drawings, and publications required by manufacturers in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)"

2.2: Delete and substitute:

"2.2 Other publications. The following document(s) form a part of this specification to the extent specified herein. The issues of the documents which are indicated as DoD adopted shall be the issue listed in the current DoDISS and the supplement thereto, if applicable."

3.1: Delete and substitute:

"3.1 First article. The indicator furnished under this specification shall be a product which has been tested and passed the test requirements in MIL-I-7070."

PAGE 6

4.4.1: Delete and substitute:

"4.4.1 Preproduction tests."

FSC 6685

Source: https://assist.dla.mil -- Downloaded: 202640-20T12:40Z
Check the source to verify that this is the current version before use.

PAGE 7

Table I: Delete and replace with new Table I.

PAGE 10

4.4.3.10: First sentence, line one; delete "qualification" and substitute "preproduction".

PAGE 11

6.4: Qualification. Delete in its entirety.

PAGE 12

6.5: Renumber paragraph to "6.4".

Concluding material: Delete in its entirety and substitute:

"Custodian:
Air Force - 99

Review activity:
DLA - GS"

Preparing activity:
Air Force - 82

(Project 6685-F648)

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T12:40Z
Check the source to verify that this is the current version before use.

TABLE I. Test tolerances

Test Point	Test Resistance	Scale Error at Room Temperature	Reduced Voltage	Scale Error at Low Temperature	Scale Error at High Temperature	Scale Error after Vibration
°C	ohms	Tol ±°C	Tol ±°C	Tol ±°C	Tol ±°C	Tol ±°C
-50	74.24	5.0 (5.0)	3.0 (3.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
-30	80.56	5.0 (5.0)	3.0 (3.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
0	90.38	5.0 (5.0)	3.0 (3.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
+30	100.91	5.0 (5.0)	3.0 (3.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
80	120.36	5.0 (5.0)	2.0 (2.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
120	137.78	5.0 (5.0)	2.0 (2.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
150	151.91	5.0 (5.0)	2.0 (2.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
200	177.95	5.0 (5.0)	3.0 (3.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)

Test Point	Test Resistance	Open and Short Circuit	Bulb Lead Resistance	EXTREME TEMPERATURE EXPOSURE		
				Scale Error at Low Temperature	Room Temperature After Low Temperature	Room Temperature After High Temperature
°C	ohms	Tol ±°C	Tol ±°C	Tol ±°C	Tol ±°C	Tol ±°C
-50	74.24	7.5 (7.5)	2.0 (2.0)	9.0 (7.5)	7.5 (7.5)	7.5 (7.5)
-30	80.56	7.5 (7.5)	2.0 (2.0)	8.0 (7.5)	7.5 (7.5)	7.5 (7.5)
0	90.38	7.5 (7.5)	2.0 (2.0)	8.0 (7.5)	7.5 (7.5)	7.5 (7.5)
+30	100.91	7.5 (7.5)	2.0 (2.0)	8.0 (7.5)	7.5 (7.5)	7.5 (7.5)
80	120.36	7.5 (7.5)	2.0 (2.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
120	137.78	7.5 (7.5)	2.0 (2.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
150	151.91	7.5 (7.5)	2.0 (2.0)	7.5 (7.5)	7.5 (7.5)	7.5 (7.5)
200	177.95	7.5 (7.5)	2.0 (2.0)	9.0 (7.5)	7.5 (7.5)	7.5 (7.5)

Source: https://assist.dla.mil -- Downloaded: 2026-02-20T12:40Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T12:40Z
Check the source to verify that this is the current version before use.