MIL-V-2/37B
9 September 1970
SUPERSEDING
MIL-V-2/37A
12 March 1962

MILITARY SPECIFICATION SHEET

VALVE, CYLINDER, GAS:

NITROUS OXIDE, OUTLET 1321 (MEDICAL)

The complete requirements for procuring the valve described herein shall consist of this document and the issue in effect of specification MIL-V-2.

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

TYPE DESIGNATION: V37-1321-2

STYLE	CLASS	COMPOSITION	INLET SIZE	OUTLET APPLICATION	HANDWHEEL	SAFETY DEVICE
III	36	A	6	1321	Required	4
IV	24	A	6	1321	Required	4

SPECIAL REQUIREMENTS:

1. The valve shall be chromium plated.
2. Mark the valve body: NITROUS OXIDE MED.
3. A valve outlet cap shall be supplied.

OPTIONAL REQUIREMENTS:

INTENDED USE:

The valves covered by this specification sheet are intended for use with nitrous oxide cylinders in medical applications. Typical capacities are 1280, 2000 and 3200 gallon cylinders.

NOTES:

1. Type designation will include any style valve from the chart above.
2. Special requirements will be mandatory for all valve procurements.
3. Optional requirements will be specified for each valve procurement.

FSC 8120

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:40Z
Check the source to verify that this is the current version before use.

SUPERSESSION DATA:

The outlet for this application was 321 in previous editions of this specification.

Custodians:

Army - ME
Navy - SH
Air Force - 68

Review activities:

Army - MI, MU
Navy - AS

User activities:

Army - AV
Navy - YD, MC

Preparing activity:

Army - ME

Project No. 8120-0150

Source: https://assist.dla.mil -- Downloaded: 202610-15T15:40Z
Check the source to verify that this is the current version before use.