NOT MEASUREMENT SENSITIVE

MIL-DTL-2/37C
9 June 1995
SUPERSEDING
MIL-V-2/37B
9 September 1970

DETAIL SPECIFICATION SHEET

VALVE, CYLINDER, GAS :

NITROUS OXIDE, OUTLET 326 (MEDICAL)

Pressures through 3000 PSIG (20,680 KPa) at 120 degrees F (48.9 degrees C)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The complete requirements for acquiring the valve described herein shall consist of this specification and the latest issue of MIL-DTL-2.

TYPE DESIGNATION: V37-326-4

Style	Class	Composition	Inlet Size	Outlet Application	Handwheel	Safety Device
III	30	A	6	326	Required	1
IV	30	A	6	326	Required	1

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:40Z
Check the source to verify that this is the current version before use.

SPECIAL REQUIREMENTS :

1. The valve shall be chromium plated.
2. Mark the valve body: NITROUS OXIDE MED.
3. A valve outlet cap shall be supplied.

INTENDED USE:

The valves covered by this specification sheet are intended for use with nitrous oxide cylinders in medical applications. Typical capacities are 1280, 2000 and 3000 gallon cylinders.

NOTES:

1. Type designation will include any style valve from the chart above.
2. Special requirements will be mandatory for all valve procurements.
3. Optional requirements will be specified for each valve procurement.

Custodians:
Army - AT
Navy - SH
Air Force - 68

Preparing activity:
Air Force - 68

(Project 8120-1025-26)

Review activities:
Army - AV
Navy - AS, MC, YD1
DLA - GS

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:40Z
Check the source to verify that this is the current version before use.