MIL-V-2/38C
9 September 1970
SUPERSEDING
MIL-V-2/38B
19 November 1964

MILITARY SPECIFICATION SHEET

VALVE, CYLINDER, GAS:

NITROUS OXIDE, OUTLET 911 (PIN-INDEX-MEDICAL)

The complete requirements for procuring the valve described herein shall consist of this document and the issue in effect of specification MIL-V-2.

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

TYPE DESIGNATION: V38-911-2

STYLE	CLASS	COMPOSITION	INLET SIZE	OUTLET APPLICATION	HANDWHEEL	SAFETY DEVICE
I	30	D	4	911	No	4

SPECIAL REQUIREMENTS:

1. The valve shall be chromium plated.
2. Mark the valve body: NITROUS OXIDE MED.

OPTIONAL REQUIREMENTS:

1. Wrench or key (required) (not required).

INTENDED USE:

The valves covered by this specification sheet are intended for use with nitrous oxide cylinders which require pin-index outlets. Typical capacities are 100 and 250 gallon cylinders.

NOTES:

1. Type designation will include any style valve from the chart above.
2. Special requirements will be mandatory for all valve procurements.
3. Optional requirements will be specified for each valve procurement.

FSC 8120

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:40Z
Check the source to verify that this is the current version before use.

Custodians:

Army - ME
Navy - SH
Air Force - 68

Review activities:

Army - MI, MU
Navy - AS

User activities:

Army - AV
Navy - YD, MC

Preparing activity:

Army - ME

Project No. 8120-0252

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:40Z
Check the source to verify that this is the current version before use.