NOT MEASUREMENT SENSITIVE

MIL-DTL-2/43D
25 September 2002
SUPERSEDING
MIL-DTL-2/43C
9 June 1995

DETAIL SPECIFICATION SHEET

VALVE, CYLINDER, GAS: PHOSGENE, (TON CONTAINER), OUTLET 660
(PRESSURES THROUGH 3000 psig (20680 kPa) AT 120 ºF (48.9 ºC))

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The requirements for acquiring the product described herein shall consist of this specification sheet and the issue of MIL-DTL-2 listed in that issue of the Department of Defense Index of Specifications and Standards (DoDISS) specified in the solicitation.

REQUIREMENTS:

TYPE DESIGNATION: V43-660-12

TABLE I. Valve requirements.

Style	Class	Composition	Inlet size	Outlet application	Handwheel	PRD[1]
I	30	C	6	660	No	Prohibited

[1] Pressure Relief Device

SPECIAL REQUIREMENTS:

1. The valve shall be forged from aluminum-silicon-bronze in conformance with alloy B in drawing number 112 of the Chlorine Institute.
2. The stem shall be nickel-copper conforming to ASTM B 164.
3. A valve outlet cap shall be supplied.
4. Mark the valve body: PHOSGENE.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:41Z
Check the source to verify that this is the current version before use.

INTENDED USE: The valves covered by this specification sheet are intended for use with 1-ton phosgene containers.

NOTES:

1. Special requirements will be mandatory for all valve procurements.

CHANGES FROM PREVIOUS ISSUE: Marginal notations are not used in this revision to identify changes with respect to the previous issue due to the extent of the changes.

Custodians:
Army - AT
Navy - SH
Air Force - 68

Preparing Activity:
DLA - GS3

(Project 8120-1047-43)

Review Activities:
Army - AV
Navy - AS, MC, YD

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:41Z
Check the source to verify that this is the current version before use.