NOT MEASUREMENT SENSITIVE

MIL-DTL-2/60A
w/Amendment 1
29 September 2006
SUPERSEDING
MIL-DTL-2/60A
25 September 2002

DETAIL SPECIFICATION SHEET

VALVE, CYLINDER, GAS: OXYGEN-HELIUM MIXTURE (HELIUM OVER 80 PERCENT), OUTLET 500, MEDICAL (PRESSURES THROUGH 3000 psig (20680 kPa) AT 120 ºF (48.9 ºC))

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The requirements for acquiring the product described herein shall consist of this specification sheet, MIL-DTL-2, and QPL-2.

REQUIREMENTS:

TYPE DESIGNATION: V60-500-2

TABLE I. Valve requirements.

Style	Class	Composition	Inlet size	Outlet application	Handwheel	PRD[1]
IIB	30	A	6	500	Yes	4
III	30	A	6	500	Yes	4
IV	30	A	6	500	Yes	4

[1] Pressure Relief Device

SPECIAL REQUIREMENTS:

1. The valve shall be chromium plated.
2. Mark the valve body: OXYGEN-HELIUM MED.
3. A valve outlet cap shall be supplied.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:44Z
Check the source to verify that this is the current version before use.

INTENDED USE: The valves covered by this specification sheet are intended for use with cylinders for oxygen-helium mixtures in medical applications where helium is over 80 percent of the mixture.

NOTES:

1. Type designation will include any style valve from table I.
2. Special requirements will be mandatory for all valve procurements.

AMENDMENT NOTATIONS: The margins of this specification are marked with vertical lines to indicate where modifications from this amendment were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations.

Custodians:
Army - AT
Navy - SH
Air Force - 68

Preparing Activity:
DLA - GS7

(Project 8120-2006-045)

Review Activities:
Army - AV
Navy - AS, MC

NOTE: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST database at http://assist.daps.dla.mil/.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:44Z
Check the source to verify that this is the current version before use.