NOT MEASUREMENT SENSITIVE

MIL-DTL-2/61
9 June 1995

DETAIL SPECIFICATION SHEET

VALVE, CYLINDER, GAS :

OXYGEN-HELIUM MIXTURE (HELIUM NOT OVER 80 PERCENT),
OUTLET 930, (PIN-INDEX) MEDICAL

Pressures through 3000 PSIG (20,680 KPa) at 120 degrees F (48.9 degrees C)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The complete requirements for acquiring the valve described herein shall consist of this specification and the latest issue of MIL-DTL-2.

TYPE DESIGNATION: V61-930-2

Style	Class	Composition	Inlet Size	Outlet Application	Handwheel	Safety Device
I	30	D	4	930	No	4

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:44Z
Check the source to verify that this is the current version before use.

SPECIAL REQUIREMENTS :

1. The valve shall be chromium plated.
2. Mark the valve body: OXYGEN-HELIUM MED.

OPTIONAL REQUIREMENTS:

1. Wrench or key (required) (not required).

INTENDED USE:

The valves covered by this specification sheet are intended for use with cylinders for mixtures of oxygen and helium where helium is not over 80 percent of the mixture and a pin-index outlet is required.

NOTES:

1. Type designation will include any style valve from the chart above.
2. Special requirements will be mandatory for all valve procurements.
3. Optional requirements will be specified for each valve procurement.

Custodians:
Army - AT
Navy - SH
Air Force - 68

Review activities:
Army - AV
Navy - AS, MC, YD1
DLA - GS

Preparing activity:
Air Force - 68

(Project 8120-1025-46)

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:44Z
Check the source to verify that this is the current version before use.