DDD-G-148D
October 15, 1976
SUPERSEDING
Int. Fed. Spec. DDD-G-00148C
April 1, 1964
Fed. Spec. DDD-G-148B
January 30, 1958

FEDERAL SPECIFICATION

GAUZE, ABSORBENT, STERILE, AND GAUZE, ABSORBENT IODOFORM IMPREGNATED (PACKING STRIPS)

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers gauze, absorbent, sterile and gauze, absorbent iodoform impregnated (packing strips) (See 6.2).

1.2 Classification.

1.2.1 Types and sizes. Gauze packing strips shall be of the types and sizes as specified (See 6.1).

Type I - Plain, sterile
Type II - Medicated, sterile

Size 1 - 1/4 inch wide
Size 2 - 1/2 inch wide
Size 3 - 1 inch wide
Size 4 - 2 inches wide

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

Federal Specifications:

PPP-B-566	-	Boxes, Folding, Paperboard.
PPP-B-585	-	Boxes, Wood, Wirebound.
PPP-B-601	-	Boxes, Wood, Cleated-Plywood.
PPP-B-621	-	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	-	Boxes, Shipping, Fiberboard.
PPP-B-676	-	Boxes, Set-Up.

Federal Standards:

Fed. Std. No. 123	-	Marking for Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal specifications required by activities outside the Federal Government for bidding purposes are available without charge from the business service centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Ft. Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established disbribution points in their agencies.)

Military Specifications:

MIL-L-10547	-	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Military Standards:

MIL-STD-105	-	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	-	Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

(Copies of Military Specifications and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless a specific issue is identified, the issues in effect on date of invitation for bids or request for proposals shall apply.

U.S. Pharmacopeial Convention, Inc:

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA.)

The National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA.)

3. REQUIREMENTS

3.1 Material. Shall be sterile, absorbent gauze packing strips for type I and shall be sterile, absorbent gauze packing strips, uniformly impregnated with iodoform for type II.

3.1.1 Suitability. For type I, size 1, shall be suitable for use as packing strips and as drainage wicks in the treatment of draining wounds. For type II, size 1, shall be suitable for use in oral surgery, surgery of the nose and sinuses, nasal hemastosis, etc., and as drainage wicks in the treatments of boils, abcesses, fistulas, and other draining wounds. For type II, size 4, shall be suitable for use as hemostatic packing, in gynecological surgery, and as an antiseptic dressing.

3.1.2 Iodoform content. For type II, shall contain not less than 3.7 percent iodoform, when determined as specified in 4.3.2.2.

3.1.3 Weave. The gauze shall be of plain ribbon loom weave.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

3.1.4 Edges. The gauze shall have selvage edges free from loose threads and ravelings.

3.1.5 Length. The length of the gauze strips, in one piece, when determined as specified in 4.3.2.3, shall average not less than 5 yards, and no sample tested shall be less than 4 yards, 32 inches in length.

3.1.6 Width. The width of the gauze strips, when determined as specified in 4.3.2.4, shall average not less than 1/4 inch ± 1/16 inch for size 1, not less than 1/2 inch ± 1/8 inch for size 2, not less than 1 inch ± 3/16 inch for size 3, and not less than 2 inches ± 1/4 inch for size 4.

3.1.7 Sterility. The gauze shall be subjected to a sterilization process which shall render the gauze sterile, and shall be tested in accordance with 4.3.2.1. Type I shall be sterilized by steam sterilization, and thoroughly vacuum dried after sterilization.

3.1.7.1 Certificate of sterility. A certificate of sterility, listing sterility lot or control number, and stating that samples from each sterilizer charge have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate shall include the following: (a) Date of sterilization, (b) Method of sterilization, (c) Specific USP test method used, (d) National stock number, (e) Item identification, and (f) Contract number.

3.1.8 Uniformity of impregnation. For type II, the distribution of the iodoform shall be even and uniform throughout the gauze. There shall be no area free of medication upon visual examination.

3.2 Ingredients.

3.2.1 Gauze. The gauze used in the manufacture of the finished gauze shall be bleached, absorbent cotton, plain ribbon loom weave with selvage edges. The gauze shall also comply with the requirements specified herein.

3.2.1.1 United States Pharmacopeia requirements. The gauze shall comply with the following USP requirements for absorbent gauze:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

Absorbency, except a 5 gram sample shall not exceed 30 seconds when tested before sterilization in accordance with USP test method for cotton.
Ignited Residue, Acid or Alkali, and Dextrin or Starch in water extract, except the limits of the weight of the residue shall not exceed 0.50% and the limits of the ignited residue shall not exceed 0.15%.
Residue on Ignition, except the limits shall not exceed 0.50%
Fatty Matter
Alcohol soluble dyes

3.2.1.2 Physical requirements. Shall comply with the requirements listed below:

Sizes	No. of warp threads per full width	Threads per inch fill
1	20 to 24	25 to 31
2	28 to 33	25 to 31
3	46 to 50	25 to 31
4	94 to 100	25 to 31

3.2.2 Medication. For type II, the medication used in the manufacture shall be iodoform, and shall comply with the tests, standards, and requirements of the NF XIII, which was official from 1 September 1970 to 30 June 1975.

3.2.3 Other ingredients. All other ingredients used in the manufacture of the gauze shall be of USP or NF quality or, if not included in either of these compendia, the ingredients shall be of the highest pharmaceutical grade for use on a preparation to be employed on the human membranes.

3.3 Workmanship. The strips shall be free from all defects which will affect the appearance of the strips or may impair their serviceability.

3.4 Age at time of delivery to the Government. Not more than 6 months shall have elapsed from the date of manufacture to the date of delivery to the Government.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the items to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.1.3 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of gauze produced from one formulation, subjected to the same manufacturing operation, and placed into final containers.

4.1.3.1 Sterility lot. A sterility lot is a lot or portion of a lot

(a) which is sterilized in the same sterilizer load, for which are terminally sterilized;

(b) which is filled aseptically into final containers through a sterilizing apparatus or assembly in a period of time not to exceed one working day or shift, for products filled aseptically without terminal sterilization; or

(c) which is subjected to some other sterilization process which shall be acceptable to the procuring activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

4.2 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

4.2.1 Examination of the end item. An examination shall be made of each sterility lot of the finished gauze for visual defects. The defects found during this examination shall be classified in accordance with the defects listed in 4.2.1.1 The sample unit for this examination shall be one jar. The sample size shall be selected in accordance with inspection level S-1 of MIL-STD-105. The AQL for major defects shall be 1.0 (percent defective), the AQL for minor defects shall be 2.5 (percent defective).

4.2.1.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects found in table I.

TABLE I. Classification of defects.

Categories	Defects 1/
Major:	
101	Finished gauze not free of loose threads and ravelings.
102	Distribution of iodoform not uniform (type II only).
103	Weave and edges not as specified.
104	Cap liner not as specified.
105	Cap seals not intact.
Minor:	
201	Exterior of jar not free from foreign matter.
202	Exterior of jar not free from medication (type II only).

1/ Examination is not restricted to the classified possible defects listed above.

4.3 Tests.

4.3.1 Testing of components. The gauze shall be tested as specified in 3.2.1.1 and 3.2.1.2. Each test shall be performed at least twice (2) per lot. If one or more individual tests fail, the lot shall be rejected. The iodoform used in type II shall be tested in accordance with the NF XIII. Sufficient quantity of each component shall be sampled from each lot or control of material to accomplish all determinations required by the test methods.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

4.3.1.1 Thread count. The warp and filling threads shall be counted as specified in the USP monograph for absorbent gauze.

4.3.1.2 Samples. If (1) the quality assurance representative submits samples to the laboratory at the Defense Personnel Support Center, or (2) the bidder (offeror) is required by the terms of the Schedule or otherwise, to submit samples to said laboratory, a quantity of approximately 100 grams of unmedicated gauze used in the manufacture of the lot of medicated gauze, accompanied by a certificate of analysis listing all test results, shall be submitted with the finished product to the Technical and Quality Assurance Division, Directorate of Medical Materiel, Defense Personnel Support Center, ATTN: Quality Assurance Branch, 2800 South 20th Street, Philadelphia, PA 19101.

4.3.2 Testing of end item. The end item shall be tested as specified below. Each sterility lot shall be tested for sterility. Each lot shall be tested for dimensional characteristics. For type II, each lot shall be tested for Iodoform content. Compliance with all test requirements shall be mandatory.

4.3.2.1 Sterility test. The sterility test shall be performed as described under "Sterility Tests for Purified Cotton, Gauze, Surgical Dressings, and Related Material," in the latest edition of the USP, including any supplements or revisions thereto.

4.3.2.2 Iodoform content (type II). Extract approximately 5 yards of the medicated gauze with 200 ml of alcohol. Add 50 ml of 0.1N silver nitrate and 10 ml of nitric acid and boil under a reflex condenser for one hour. Allow the solution to cool and add 1 ml of ferric ammonium sulfate, T.S., and titrate the excess silver nitrate with 0.1N ammonium thiocynate to a pink end point. The gauze used in the calculation of the iodoform content shall be dried to content weight to determine the weight of the extracted and dried gauze. Calculate the percent iodoform as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

$$M = (A - B) \times 13.12$$

$$\text{Percent Iodoform} = \frac{M}{W + M} \times 100$$

M = weight of iodoform found
A1/ = Volume of 0.1N silver nitrate volumetric solution (ml)
B1/ = Volume of 0.1N ammonium thiocynate volumetric Solution (ml) to titrate the excess silver nitrate
13.12 = Equivalent weight of iodoform
W = Weight of extracted and dried gauze

1/ Make all necessary adjustments in calculation to correct for normalities of solutions.

For military activities, as an alternate method for the iodoform content, an applicable, accurate, reproducible method may be used provided it is approved and acceptable to the Defense Personnel Support Center.

4.3.2.3 Length. The gauze shall be laid out on a smooth horizontal surface without tension. The length shall be measured parallel to the edge using a yardstick, metal tape, or any other suitable measuring device graduated in a maximum of 1/4 inch. The length shall be determined on randomly selected samples of each lot; inspection level shall be S-1 and AQL of 2.5 (percent defective). The same samples may be used to determine the width.

4.3.2.4 Width. The gauze width shall be measured using a ruler, metal tape, or any other suitable measuring device graduated in a maximum of 1/16 inch. The width shall be determined on randomly selected samples of each lot; inspection level shall be S-1 and AQL of 2.5 (percent defective). The same samples may be used to determine the length. The width of the sample shall be determined by taking a minimum of three measurements per 5 yard length from the beginning, middle, and end of each strip and averaging the measurements.

4.4 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of section 5. Defects shall be scored as specified in table II. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 expressed in terms of percent defective.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

TABLE II. Examination of preparation for delivery. 1/

Examination	Defect
Containers	Not as specified.
Contents	Not as specified.
Markings	Omitted; incorrect; illegible; improper size, location, sequence, or method of application.
Materials	Component missing or damaged.
Workmanship	Bulging or distortion of containers. Cushioning inadequate, improper or missing

1/ The above examination is only applicable to the unit, intermediate and exterior containers.

4.4.1 Examination of closure and reinforcement for containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure, waterproofing or banding, shall be in accordance with the appendix to that specification.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (See 6.1).

TABLE III. Packaging quantities.

Type	Size	Unit package	Intermediate package	Exterior container
I	1 thru 4	1 unit	12 units	144 units
II	1 thru 3	1 unit (12 jars)	Not required	12 units
II	4	1 unit (12 jars)	Not required	8 units

5.1.1 Level A.

5.1.1.1 Immediate container. Gauze of one type and size only shall be supplied pleat-folded in a commercially available jar with a screw-cap and seal acceptable to the Food and Drug Administration for the product contained therein. The container shall be of such compatibility that neither the container nor the contents are altered in any way by each other. The container shall be capable of maintaining sterility of contents unless damaged or opened. In addition, the container shall be supplied with an outer tamperproof seal.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

5.1.2.2 Unit package.

5.1.2.2.1 Type I. At the option of the contractor, a unit package may be furnished for each jar. If a unit package is furnished, each jar shall be packaged in a box constructed in accordance with PPP-B-566 or PPP-B-676. Closure of the box shall be adequate to prevent accidental opening.

5.1.1.2.2 Type II. Twelve jars (one package) containing gauze of one size only shall be packaged in a box constructed in accordance with PPP-B-566, "E" flute, or PPP-B-636, type CF, class domestic. The box shall include full-height, half-slotted style partitions so arranged that each jar is contained in an individual cell. Partitions shall be fabricated of the same material as the box. As an alternate, the partitions may be fabricated from 0.03^ inch kraft lined chipboard. When jars are packaged in individual boxes or when plastic jars are furnished, partitions shall not be required. Closure of the box shall be adequate to prevent accidental opening.

5.1.1.3 Intermediate package. Twelve jars containing type I gauze of one size only shall be packaged as specified in 5.1.1.2.2.

5.1.2.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (See 6.1).

5.2.1 Level A. The number of units of one type and size gauze as specified in table III shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. The number of units of one type and size gauze as specified in table II shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and king commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with the special marking for the method of sterilization specified in 5.3.2.1 and Fed. Std. No. 123.

5.3.2 Military activities.

5.3.2.1 Immediate container. Each immediate container shall be labeled in accordance with the requirements of the Federal Food, Drug and Cosmetic Act and shall be that usually furnished by the supplier as commercial practice. The lot (control) number and legend "STERILE", and "STERILITY GUARANTEED UNLESS JAR IS DAMAGED OR OPENED" or similar statement shall be shcwn.

5.3.2.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Sterility lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization G-Gas, R-Radiation, C-Cold, S-Steam or D-Dry Heat sterilization shall be shown following the level of packaging and date of packaging. Example" "A 11/75-G". Final sterilization shall be on or after indicated packaging date.

NOTE: When a unit package is not furnished, the National Stock Number and the date and method of sterilization shall be shown on the immediate container. Strip labeling may be used.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

5.3.2.3 Intermediate package. Each intermediate package shall be marked in accordance with MIL-STD-129. Sterility lot (control) number, contract or purchase order number, name of contractor and the method of sterilization as specified in 5.3.2.2 shall be shown.

5.3.2.4 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. The sterility lot (control) number and the method of sterilization as specified in 5.3.2.2 shall also be shown.

6. NOTES

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

(a) Title, number and date of this specification.
(b) Type and size required (See 1.2.1).
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).
(d) The National Stock Number.

6.2 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification	Specification Designation
6510-00-533-2150	GAUZE, ABSORBENT, 1/4 inch by 5 yards	Type I, Size 1
6510-00-559-3162	GAUZE, ABSORBENT, 1 inch by 5 yards	Type I, Size 3
6510-00-559-3163	GAUZE, ABSORBENT, 2 inches by 5 yards	Type I, Size 4
6510-00-559-3172	GAUZE, ABSORBENT, 1/2 inch by 5 yards	Type I, Size 2
6510-01-003-7696	GAUZE, ABSORBENT, IODOFORM IMPREGNATED, 2 inches by 5 yards, 12s	Type II, Size 4
6510-01-003-7697	GAUZE, ABSORBENT, IODOFORM IMPREGNATED, 1/4 inch by 5 yards, 12s	Type II, Size 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

MILITARY CUSTODIANS:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Civil Agency Coordinating Activities:
GSA-FSS
HEW-FDS, NIH
VA-DMS

Project No. 6510-0595

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See section 2 of this specification to obtain extra copies and other documents referenced herein. Price 25 cents each.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers' vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE: DDD-G-148D - Gauze, Absorbent, Sterile, and Gauze, Absorbent Iodoform Impregnated (Packing Strips)

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.