MIL-C-19246C
30 November 1964
Superseding
MIL-C-19246B(Weps)
9 August l960

MILITARY SPECIFICATION

CONNECTOR, OXYGEN MASK HOSE, TYPE MC-3A

This specification is mandatory for use by all Departments and Agencies of the Department Of Defense.

1. SCOPE

1.1 This specification covers type MC-3A oxygen mask home connector for connecting the oxygen mark hose to the oxygen regulator home.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of inviation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-636	Boxes, Fiberboard
PPP-T-76	Tape; Pressure Sensitive Adhesive Paper, Water Resistant, (for Carton Sealing)

Military

MIL-P-116	Preservation, Methods of
MIL-E-5272	Environment Testing, Aeronautical and Associated Equipment, General Specification for

FSC 1660

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military (Continued)

MIL-F-22191	Films, Transparent, Flexible, Heat Sealing, for Packaging Applications
MIL-O-27210	Oxygen, Aviator's Breathing, Liquid and Gas

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MIL-STD-130	Identification Marking of U.S. Military Property
MS22016	Connector, Oxygen Mask Hose, Type MC-3A
MS22058	Connector, Oxygen Hose to Regulator
MS33586	Metals, Definition of Dissimilar

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2. Other publications - The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Society far Testing Materials Standard -

D1149	Method of Test for Accelerated Ozone Cracking of Vulcanized Rubber

(Application for copies should be addressed to the American Society for Testing Materials, 1916 Race Street, Philadelphia 3, Pennsylvania.)

Source: https://assist.dla.mil -- Downloaded: 202610-18T12:38Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Qualification - The connector furnished under this specification shall be a product which has been inspected, and passed the qualification inspections specified herein, and has been listed on or approved for listing on the applicable qualified products list.

3.2 First Article - The connector furnished under this specification shall be a product which has been inspected and passed the first article inspection specified herein.

3.3 Materials - Materials shall conform to applicable specifications and shall be as specified herein and on applicable drawings. Materials which are not covered by specifications or which are not specifically described herein, shall be of the best quality, of the lightest practicable weight, and suitable for the purpose intended.

3.3.1 Nonmetallic materials - Nonmetallic materiala, except for the poppet material as shown on MS22016, shall be of an ozone-resistant composition.

3.3.1.1 Age - Elastomer components shall not be more than 12 months old from the date of manufacture to the date of delivery to any Government service or to any airframe or accessory manufacrer.

3.3.2 Metal Parts - All metal parts shall be of a corrosion-resistant material or treated in a manner to render them adequately resistant to corrosion.

3.3.2.1 Dissimilar metals - Unless suitably protected against electrolytic corrosion, dissimilar metals shall not be used in intimate contact with each other. Dissimilar metals are defined in MS33586.

3.3.3. Protective treatment - When materials are used in the construction of the connector assembly that are subject to deterioration when exposed to environmental conditions likely to occur during service usage, they shall be protected against such deterioration in a manner that will in no way prevent compliance with the performance requirements of this specification. Protective coating which might crack, chip, or scale during normal service life or under extremes of environmental conditions shall not be used.

3.4 Design and construction - The design and construction of the connector assembly shall be in accordance with MS22016.

Source: https://assist.dla.mil -- Downloaded: 202610-18T12:38Z
Check the source to verify that this is the current version before use.

3.4.1 Connector assemblies - Connector assemblies shall have a valve which will make it possible for the user to breathe normally up through. the connector when the valve is properly inserted into a connector confoming to MS22058. When disconnected the valve shall permit inhalation. but a noticeable resistance shall be introduced to indicate that disconnection has occurred. A spring-loaded relief valve shall be located in the connector body and so adjusted as to allow oxygen from the bailout cylinder to escape whenever the mask pressure exceeds a pressure equal to 12 inches of water more than ambient pressure. The relief valve shall vent small flows before the mask pressure reaches 12 inches of water more than ambient pressure.

3.4.2 Bailout attachment - The side bailout attachment shall contain a check valve as shown on MS22016.

3.5 Performance -

3.5.1 Disconnection force - The connector when tested as specified in 4.7.2, shall disconnect between 12 and 20 pounds.

3.5.2 Leakage (room temperature and extreme temperature) - Leakage of the connector assembly, when tested as specified in 4.7.3, 4.7.9.1 and 4.7.9.2, shall not exceed 0.01 liters per minute.

3.5.3 Relief valve setting - The connector when tested as specified in 4.7.4, shall beet the following requirements:

Flow (1pm)	Developed pressure (inches of water)
2.5	less than 12
12	between 11 and 25
50	between 12 and 18

3.5.4 Pressure drop through MC-3A connector - The pressure drop through the connector, when tested as specified in 4.7.5, shall not be less than 4 inches nor more than 6 inches of water.

3.5.5 Pressure drop through MC-3A connector with MS22058 connector - The pressure drop through the assembly, when tested as specified in 4.7.6, shall not exceed the following values:

Flow (lpm)	Pressure drop (inches of water)
135	1.50
90	.66
70	.40
50	.20
30	.07

Source: https://assist.dla.mil -- Downloaded: 202610-18T12:38Z
Check the source to verify that this is the current version before use.

3.5.6 Proof pressure - The connector, when tested as specified in 4.7.7, shall not show any evidence of structural damage.

3.5.7 Bailout flow - The connector, when tested as specified in 4.7.8, shall permit a minimum flow of 9 liters per minute.

3.5.8 Endurance - The connector, when tested as specified in 4.7.10, shall disconnect between 12 and 20 pounds.

3.5.9 Drop test - The connector, when tested as specified in 4.7.11, shall operate satisfactorily.

3.5.10 Salt spray - The connector, when tested as specified in 4.7.12, shall not show any evedence of corrosion.

3.5.11 Disassembly and examination - the connector, when tested as specified in 4.7.13, shall not show evidence of excessive deterioration or wear due to testing which might effect the connector's performance.

3.5.12 Ozone-resistance - The elastomer components, when tested as specified in 4.7.14, shall not bloom and shall not show any evidence of checking, cracking, or other damage.

3.6 Interchangeability - All parts having the same manufacturer's part number shall be functionally and dimensionally interchangeable.

3.7 Weight - The weight of the complete connector shall not exceed 0.20 pounds.

3.8 Identification of product - The connectors shall be marked for identification in accordance with MIL-STD-130.

3.9 Workmanship - The connectors shall be uniform in quality, and shall be free from irregularities, defects, or foreigh matter which could adversely affect safety, performance, reliability or durability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection - Unless otherwise specified, in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

4.2 Classification of inspection - The examination and testing of the connectors shall be classified as follows:

(a) Qualification inspection - Qualification inspection consists of examinations and tests performed on samples submitted for approval as qualified products,

(b) First article inspection - First article inspection consists of examinations and tests performed on samples which are representative of the production item after award of a contract to determine that the production item meets the requirements of this specification.

(c) Quality conformance inspection - Quality conformance inspection consists of examinations and tests performed on individual products or lots to determine conformance of the products or lots with the requirements set forth in this specification.

4.3 Qualification inspection - Qualification inspection shall consist of all the examinations and tests of this specification,

4.3.1 Qualification samples - Qualification samples shall consist of five complete connector assemblies and three elastomer test slabs, 6 by 6 by 0.075 inches, composed of the identical ozone-resistant composition and cure that was used in manufacturing the connector assemblies. The qualification samples shall be forwarded to the Supply Officer, Naval Air Engineering Center, Philadelphia, Pennsylvania 19112, Attention: Director, Aerospace Crew Equipment Laboratory. The samples shall be plainly identified by securely attached durable tags marked with the following information:

Sample submitted by (name) (date) for qualification inspection in accordance with the requirements of MIL-C-l9246C and number under authorization {reference authorizing letter and number, see 8.4).

4.4 First article inspection - First article inspection shall consist of the examinations and tests deemed necessary by the testing activity.

4.4.1 First article samples -

4.4.1.1 Applicable to the Bureau of Naval Weapons - Unless otherwise specified, as soon as practicable after award of a contract or order, the manufacturer shall submit three complete connector assemblies. The samples

Source: https://assist.dla.mil -- Downloaded: 2026l0-18T12:38Z
Check the source to verify that this is the current version before use.

shall be representative of the construction, workmanship components, and materials to be used during production. When a contractor is in continuous production of these connector assemblies from contract to contract, submission of further first article samples on the new contract may be waived at the discretion of the procuring activity. Approval of the first article samples or the waiving of the first article inspection does not preclude the requirements of submitting to the quality conformance inspection. The first article samples shall be forwarded to the Supply Officer, Naval Air Engineering Center, Philadelphia, Pennsylvania 19112, Attention: Director, Aerospace Crew Equipment Laboratory. The first article samples shall be plainly identified by securely attached durable tags marked with the following information:

Sample submitted by (name) (date) for first article inspection in accordance with the requirements of MIL-C-19246C under contract no.

4.4.1.1.1 Upon completion of the first article inspection, all the applicable inspection reports and when applicable, recommendations and cements pertinent for use in monitoring production shall be forwarded to the Government quality control representative. One approved connector shall be returned to the manufacturer for use in monitoring production. The other connectors shall be consumed or destroyed in the first article inspection and shall not be considered as part of the quantity to be delivered under contract.

4.4.1.2 Applicable to the Army - For purchases by the Army, the quanitity and kind of samples, the name and address of the first article inspection activity and other pertinent information (see 4.4.1.1 and 4.4.1.1.1 for information) shall be as specified (see 6.2).

4.5 Quality conformance inspection - Quality conformance inspection shall consist of the following examinations and tests:

Visual examination
Dimensions
Disconnection force
Leakage (room temperature)
Relief valve setting
Pressure drop through MC-3A connector
Pressure drop through MC-3A connector with MS22058 connector
Bailout flow
Preparation for delivery

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

4.5.1 Sampling -

4.5.1.1 Inspection lot -

4.5.1.1.1 Connector - An inspection lot size shall be expressed in units of one connector made under essentially the same conditions and from the same materials and components. The sample unit shall be one connector.

4.5.1.1.2 Preparation for delivery - An inspection lot size shall be expressed in units of one fully prepared shipping container, containing connector assemblies fully prepared for delivery made from essentially the same materials and components. The sample unit shall be one shipping container, containing connectors fully prepared for delivery with the exception that it need not be sealed.

4.5.1.2 Sampling for tests and examinations of connector assemblies - The sample size, acceptance criteria, tests, and examinations required for the connector assemblies shall be as specified in Table I.

4.6 Test conditions -

4.6.1 Gas - Unless otherwise specified, the gas used in testing the connector assemblies shall be oxygen conforming to MIL-O-27210, type I.

4.6.2 Temperature and pressure - Unless otherwise specified, inspections shall be conducted at local ambient temperatures and barometric pressure. Corrections shall be made to provide agreement with the temperature and pressure calibration of the instruments. Inspection data provided by any instrument not calibrated to normal temperature and pressure (NTP) conditions shall be corrected to determine NTP requirements, NTP conditions are 29.92 inches of mercury and 70°F.

4.7 Inspection methods -

4.7.1 Visual examination -

4.7.1.1 Connector - Every connector shall be examined visually (for critical defects) to determine conformance to this specification and to MS22016. The classification of defects, Table II, shall be used to classify the defects found.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

TABLE I

SAMPLE SIZE, ACCEPTANCE CRITERIA, TESTS, AND EXAMINATIONS OF THE CONNECTOR ASSEMBLIES

INSPECTION	TYPE OF INSPECTION	METHOD	SAMPLE SIZE	ACCEPTANCE CRITERIA
Visual examination (See classification of defects)	Critical Minor	4.7.1.1	Every connector assembly for critical defects. Inspection Level II 1/ for minor defects.	Reject all units with any critical defects. An acceptable quality level of 2.5 defects per hundred units for minor defects.
Dimensions	Major	4.7.1.1.1	Inspection level S-2 1/	Acceptance number zero, rejection number 1.
Disconnection force	Critical	4.7.2	Every connector assembly.	Reject all defective units.
Leakage (room temperature)	Critical	4.7.3	Every connector assembly.	Reject all defective units.
Relief valve setting	Critical	4.7.4	Every connector assembly.	Reject all defective units.
Pressure drop through connector	Critical	4.7.5	Every connector assembly.	Reject all defective units.
Pressure drop MC-3A connector with MS22058 connector	Major	4.7.6	Inspection Level S-2 1/	Acceptance number zero, rejection number 1.
Bailout flow	Major	4.7.8	Inspection Level S-2 1/	Acceptance number zero, rejection number 1.
Preparation for delivery	Minor	4.7.12	Inspection Level S-2 1/	Total acceptable quality level of 4.0 percent defective

1/ The sample size shall be based only on the applicable sample size code letter corresponding to the specified inspection level of MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

4.7.1.1.1 Dimensions - The connector shall be checked dimensionally to determine conformance to MS22016.

4.7.1.2 Preparation for delivery - Each of the fully prepared shipping containers, containing connectors selected as a sample unit from the lot shall be examined to determine that the packaging, packing, and marking conform to this specification. The classification of defects, Table III, shall be used to classify the defects found.

4.7.2 Disconnection force - The connector shall be installed in a connector conforming to MS22058. The connector assembly shall then be subjected to a straight pull parallel to the longitudinal axis. An average of five pulls conducted at five second intervals shall be used as the disconnection force. The disconnection force shall be as specified in 3.5.1.

4.7.3 Leakage (room temperature) - The connector shall be installed in a connector conforming to MS22058. An internal pressure of 1 psi at room temperature shall be applied to the connects? assembly while it is subjected to a pull of not more than 2.0 pounds applied in any direction, The leakage shall be as specified in 3.5.2.

4.7.4 Relief valve setting - A piezometer ring, to measure the developed pressure, shall be located at 2 $\pm$ 0.5 inches from the oxygen-mask end of the connector. Pressure shall be applied to the oxygen mask end of the connector to obtain flows of 2.5, 12, and 50 liters per minute. The developed pressure shall then be measured at the specified flows and shall be as specified in 3.5.3.

4.7.5 Pressure drop through MC-3A connector - A piezometer ring shall be located at 2 $\pm$ 0.5 inches from the end of the connector. The ring shall be located on the mask end if suction is used or on the regulator end if pressure is used. The flow duct shall be constructed to assure a straigh flow for at least 6 inches immediately before entering the connector if pressure is used or immediately after leaving the connector if suction is used. The connector shall be mounted in a vertical position, mask end up, and so located as to permit a free flow of room air out of the connector if pressure is used or into the connector if suction is used. A pressure or suction shall be applied to the connector to obtain a flow of 15 liters per minute from the regulator end to the mask end. The pressure drop through the connector shall then be measure by means of the piezometer ring and shall be as specified in 3.5.4

Source: https://assist.dla.mil -- Downloaded: 202610-18T12:38Z
Check the source to verify that this is the current version before use.

TABLE II

CLASSIFICATION OF DEFECTS FOR VISUAL EXAMINATION

CRITICAL	MINOR
1. Material imperfections - foreign matter embedded. 2. Surface - unclean, rough, misaligned, or containing cracks, nicks, or other flaws. 3. Any component missing, malformed, fractured, or otherwise damaged. 4. Any component loose or otherwise not securely retained. 5. Incorrect assembling or improper positioning of components. 6. Any functioning part that works with difficulty. 7. Faulty workmanship or other irregulaities.	201. Marking - missing, insufficient, incorrect, illegible, or not permanent.

TABLE III

LIST OF DEFECTS FOR PREPARATION FOR DELIVERY

ITEM	DEFECTS
Exterior and interior markings	Missing, incorrect, incomplete, illegible; of improper size, location, sequence; or method of application; markings not the same on the interior and exterior containers.
Packaging and Packing materials	Any non-conforming component; any component missing, damaged, or otherwise defective.
Workmanship	Inadequate application of the components such as incomplete closure of the unit package, intermediate package, container flaps, loose strappings, etc.; bulging or distortion of the containers.
Exterior and interior weight or content	Number per container is more or less than required; gross or net weight exceeds the requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-18T12:38Z
Check the source to verify that this is the current version before use.

4.7.6 Pressure drop through MC-3A connector with MS22058 connector - A piezometer ring shall be located 2 ± 0.5 inches from the applicable end of the connector. For instructions concerning the use of the piezometer ring, see 4.7.5. The MC-3A connector shall then be installed in a connector conforming to MS22058. Pressure or suction shall be applied to the connector to obtain flows of 30, 50, 70, 90, and 135 liters per minute from the regulator end to the mask end. The pressure drop through the assembly shall then be measured by means of the piezometer ring and shall be as specified in 3.5.5.

4.7.7 Proof pressure - The connector shall be installed in a connector conforming to MS22058. The connector assembly shall be subjected to an internal pressure of 5 psi for a period of 3 minutes. The connector shall pass the performance requirements specified in 3.5.6. The connector shall then be. subjected to and pass the tests specified in 4.7.4 and 4.7.6.

4.7.8 Bailout flow - Attach the hose of an 1800 psi H-2 emergency bailout bottle to the emergency oxygen port of a MC-3A connector and connect a flowmeter to the mask end of the connector. The H-2 emergency bailout bottle shall be activated and the flow of oxygen through the connector measured. The bailout flow though the connector shall be as specified in 3.5.7.

4.7.9 Leakage (extreme temperature) -

4.7.9.1 Leakage (high temperature) - The connector shall be conditioned at a temperature of 160 ± 2°F for a period of 2 hours. After conditioning and while still at this temperature, the connector shall be subjected to and pass the leakage test specified in 4.7.3.

4.7.9.2 Leakage (low temperature) - The connector shall be conditioned at a temperature of -65 ± 2°F for a period of 2 hours. After conditioning, and while still at this temperature, the connector shall be subjected to and pass the leakage test specified in 4.7.3.

4.7.10 Endurance - The connector shall be inserted into and disconnected from a connector conforming to MS22058 for a total of 1000 times each. The insertions and disconnections shall be conducted at the rate of one complete cyle per five seconds. After 1000 cycles have been completed, the connector shall remain at rest for a period of 24 hours. The connector shall then be subjected to the test specified in 4.7.2 to determine the disconnection force after endurance. The disconnection force shall be as specified in 3.5.8. The connector shall then be subjected to and pass the tests specified in 4.7.3 and 4.7.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

4.7.11 Drop test - The connector shall be subjected to a free fall of 2 feet so that the end that connects with the MS22058 connector hits onto concrete or an equally firm surface. This procedure shall be repeated for a total of 100 times. The connector shall then be subjected to and pass the tests specified in 4.7.3, 4,7.4, and 4.7.6.

4.7.12 Salt spray - Both ends of the connector shall be unplugged exposing the interior as well as the exterior of the connector. The connector shall be subjected to the salt spray test specified in MIL-E-5272, Procedure II. The connector shall pass the performance requirements specified , in 3.5.10. Upon completion of the salt spray test, the connector shall then be subjected to and pass the tests specified in 4.7.2, 4.7.3, 4.7.4, and 4.7.6.

4.7.13 Disassembly and examination - At the conclusion of all the above inspections the connector shall be disassembled and examined. The connector shall pass the performance requirements specified in 3.5.11. The connector shall be disassembled and examined after any test in which connector failure indicates faulty construction.

4.7.14 Ozone-resistance test (elastomer components) - The three elastomer test slabs submitted for qualification inspections shall be tested for ozone-resistance. The test apparatus shall be in accordance with ASTM Method No. D1149. The test slabs shall be elongated 20 percent, placed in an ozone-free atmosphere for 24 hours, then placed in the ozone chamber. The chamber shall be adjusted to 100 $\pm$ 2°F and to give an exposure of ozone concentration of 120 $\pm$ 10 parts by volume of ozone per million parts by volume of air. The air-ozone velocity in the chamber shall be at least 2 feet per seconds. The material shall be exposed to these conditions for 60 minutes. The test slabs shall be examined under a 10X magnification and shall pass the performance requirements specified in 3.5.12.

5. PREPARATION FOR DELIVERY

5.1 Preservation and packaging - Preservation and packaging shall be Level A or C as specified (see 6.2).

5.1.1 Level A - Each connector shall be cleaned and packaged in accordance with MIL-P-116, Method IC-1(X). The connector shall be cleaned by any appropriate procedures of MIL-P-116 that do not employ petroleum or hydrocarbon products. Each connector shall then be packaged within a bag fabricated of transparent flexible barrier conforming to MIL-F-22191, Type II, and heat sealed. The bag shall be sufficiently oversized to provide cushioning protection for the item.

Source: https://assist.dla.mil -- Downloaded: 202610-18T12:38Z
Check the source to verify that this is the current version before use.

5.1.2 Intermediate packaging - Unless otherwise specified, ten connectors: packaged as specified in paragraph 5.1.1, shall be packaged within a container conforming to PPP-B-636, Type I, Class 2, W6c. The container shall be equipped with criss-crossed corrugated fiberboard separators which shall segregate and hold each connector snugly. The closure shall be accomplished by tape sealing all seams and joints with pressure sensitive tape conforming to PPP-T-76, two inch minimum width,

5.1.3 Level C - Each connector shall be cleaned and packaged in a reamer that will afford the item the degree of protection necessary to prevent chemical deterioration and physical damage during transit from contractor's plant to the receiving activity. No petroleum or hydrocarbon compound shall be used in the cleaning or packaging procedure.

5.2 Packing - Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A- Unless otherwise specified, no more than ten intermediate containers of ten connectors each, packaged as specified in paragraphs 5.1.1 and 5.1.2 shall be packed in an exterior shipping container conforming to PPP-B-636, Type I, Class 2. The intermediate containers shall be arranged within the shipping container in two stacks of five each in a manner that will conserve volume without impairing container rigidity. The gross weight of each packed shipping container shall not exceed the limitations of the container specification. The containeer closure shall be accomplished in accordance with procedures in the appendix of the specification.

5.2.2 Level B - Packing procedures shall be the same as specified in paragraph 5.2.1, except that containers to be used shall be domestic service grade conforming to PPP-B-636, Type I, Class 1.

5.2.3 Level C - The packaged connecters, which the carrier requires to be packed, shall be packed for domestic shipment in shipping containers that will assure safe arrival at the lowest possible transportation rate to the point of delivery. Containers and procedures employued shall conform to the latest issue of rules and regulations applicable to the mode of transportation utilized.

5.3 Marking - Nomenclature tags shall be inserted in each bag showing title, part number or other data as required by procuring activity. In addition, interior and exterior containers shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 202640-18T12:38Z
Check the source to verify that this is the current version before use.

5.3 Precautionary marking - The following precautionary marking shall appear on each package:

"CAUTION: DO NOT ALLOW CONTAMINANTS OF ANY KIND TO BE USED ON OR ABOUT THE OXYGEN CONNECTORS."

6. NOTES

6.1 Intended use - The connector covered by this specification is intended for use in connecting the oxygen mask hose to the oxygen regulator hose in the insllation of high-pressure and low-pressure breathing oxygen equipment in aircraft.

6.3 Ordering Data - procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Items of data required (see 6.3).
(c) Whether first article inspection is required for Bureau of Naval Weapons or Army purchases (see 4.4.1.1 and 4.4.1.2).
(d) Selection of applicable levels of preservation, packaging, and packing.
(e) Applicable methods of cleaning and preservation.

6.3 Data - For the information of Contractors and Contracting Officers, any of the data specified in applicable documents listed in Section 2 of this specification or referenced lower-tier documments need not be prepareed for the Government and shall not be furnished to the Government unless specified in the contract or order. The data to be furnished shall be listed on DD Form 1423(Contractor Data Requirements List), which shall be attached to and made a part of the contract or order. NavWeps Form 4200/15 (Drawings, Lists, and Specificatins Required) shall be attached where applicable.

6.4 With respect to products requiring qualification, awards will be made only for such products as have, prior to bid opening date, been tested and approved for inclusion in the applicable Qualified Products List whether or not such products have actually been so listed by that date. The attention of the suppliers is called to this requirement, and manufacturers are urged to arrange to have the products that they propose to offer to the Federal Government tested for qualification in order that they may be eligible to be awarded contracts or orders for the products covered by this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

The activity responsible for the qualified products lists is the Chief, Bureau of Naval Weapons, Department of the Navy, Washington, D.C. 20360; however, authorization for qualification of products shall be obtained from the Commanding Officer, Naval Air Engineering Center, Philadelphia, Pennsylvania 19112, Attention: Director, Aerospace Crew Equipment Laboratory. Prior to submission-of the samples for qualification inspection, the manufacturer shall submit a request to the Aerospace Crew Equipment Laboratory indicating a date on which the samples can be forwarded and also request an authorization number to accompany the samples.

Custodians:

Army-MO
Navy - WP

Preparing Activity:

Navy - WP

Project No. (1660-0032)

Review Activities:

Army - MO
Navy - WP

NOTICE : Review/user information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T12:38Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

MIL-C-19246C Connector, Oxygen Mask Hose, Type MC-3A

ORGANIZATION *(Of submitter)*	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*	DATE

DD FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE C-8279

Source: https://assist.dla.mil -- Downloaded: 202640-18T12:38Z
Check the source to verify that this is the current version before use.