MIL-M-10530B(EL)
2 April 1968
Superseding
MIL-M-10530A(EL)
2 May 1961

MILITARY SPECIFICATION
METEOROLOGICAL STATION, MANUAL AN/PMQ-1( )
(COMPONENTS OF)

1. SCOPE

1.1 This specification covers the following components of Meteorological Station, Manual, AN/PMQ-1( ).

Thermometer Support
Syringe
Gage, Precipitation ML-435( )/PMQ-1
Psychrometer ML-436( )/PMQ-1
Tripod, Anemometer MT-869( )/PMQ-1
Case, CY-952( )/PMQ-1
Carrying Case, Meteorological Equipment CY-1285( )/PMQ-1
Wet-Bulb Psychrometer Wick

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

L-P-380	Plastic Molding Material, Methacrylate
T-T-871	Twine, Cotton, Wrapping
PPP-B-636	Box Fiberboard
PPP-C-843	Cushioning, Material, Cellulose
PPP-F-320	Fiberboard, Corrugated and Solid, Sheet Stock and Cut Shapes
PPP-P-291	Paperboard, Wrapping, Cushioning

FSC 6660

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS (con't)

FEDERAL

PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water Resistant, for Carton Sealing.
PPP-T-97	Tape, Pressure-Sensitive Adhesive, Filament Reinforced

MILITARY

MIL-P-116	Preservation, Methods of
MIL-I-631	Insulation, Electrical, Synthetic-Resin, Composition, Non-Rigid
MIL-P-11268	Parts, Materials and Processes used in Electronic Communication Equipment
MIL-M-13231	Marking of Electronic Items
MIL-F-14072	Finishes for Ground Signal Equipment
MIL-B-43291	Box, Fiberboard, Corrugated, Double-Wall, Weather Resistant

STANDARDS

MILITARY

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MIL-STD-169	Extreme-Temperature Cycle

Source: https://assist.dla.mil -- Downloaded: 202610-17T14:12Z
Check the source to verify that this is the current version before use.

STANDARDS (con't)

MIL-STD-170	Moisture Resistance Test Cycle for Ground Signal Equipment
MIL-STD-454	Standard General Requirements for Electronic Equipment
MIL-STD-810	Environmental Test Methods

DRAWINGS

ELECTRONICS COMMAND

SC-DL-96613	Thermometer Support
SC-DL-96634	Syringe
SC-DL-96684	Gage, Precipitation ML-435( )/PMQ-1
SC-DL-96692	Psychrometer ML-436( )/PMQ-1
SC-DL-96763	Tripod, Anemometer MT-869( )/PMQ-1
SC-DL-96790	Case, CY-952( )/PMQ-1
SC-DL-96825	Carrying Case, Meteorological Equipment CY-1285( )/PMQ-1

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Description. - This equipment is composed of lightweight portable meteorological items suitably packaged in a carrying case for transportation by manpack.

3.2 Construction. - The equipment shall be constructed in accordance with the following drawings:

Source: https://assist.dla.mil -- Downloaded: 202610-17T14:12Z
Check the source to verify that this is the current version before use.

3.2 Construction (con't).

SC-DL-96613	Thermometer Support
SC-DL-96634	Syringe
SC-DL-96684	Gage, Precipitation ML-435( )/PMQ-1
SC-DL-96692	Psychrometer ML-436( )/PMQ-1
SC-DL-96763	Tripod, Anemometer MT-869( )/PMQ-1
SC-DL-96790	Case, CY-952( )/PMQ-1
SC-DL-96825	Carrying Case, Meteorological Equipment CY-1285( )/PMQ-1

3.3 Preproduction samples. - The contractor shall furnish two preproduction samples of each component listed in 1.1 for approval, if required by the invitation for bids and contract.

3.4 Parts, materials, and processes; general. - In addition to the requirements of this specification, the requirement of MIL-P-11268, including the selection requirements therein, shall apply.

3.5 Finish, protective. - Equipment shall be given protective finish in accordance with MIL-F-14072, and the equipment drawings (See 4.4).

3.6 Marking of equipment. - Marking shall conform to MIL-M-13231 and the equipment drawings.

3.7 Interchangeability. - Like units, assemblies, subassemblies, and replaceable parts shall be physically and functionally interchangeable, without modification of such items or of the equipment. (See 4.12). Individual items shall not be hand-picked for fit or performance; however, matched pairs or sets, when permitted, may be interchangeable as such. Reliance shall not be placed on any unspecified dimension, rating, characteristic, etc.

3.8 Detailed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

3.8.1 Weather recording form. - The weather recording form shall be Signal Corps Form 445.

3.8.2 Syringe. - The syringe shall be fabricated in accordance with the applicable drawings, and shall be of a Type A plastic in accordance with MIL-I-631.

3.8.3 Case CY-952( )/PMQ-1. - The carrying case shall be so constructed that if any of the components are removed, the remaining items will not be free to rattle or break while in transit.

3.8.4 Wick. - The wet-bulb thermometer wick shall be Wallace and Tiernan Part No. LP-316-1, or equal. Tropicalization treatment shall not be applied to the wick.

3.8.5 Pinchcock. - The pinchcock shall be Eimer and Amend No. 5-855, size A, or equal.

3.8.6 Psychrometer motor. - When fully wound, the spring motor in the psychrometer shall operate the fan at least 15 minutes at any temperature in the specified range (See 3.9.2(b).

3.8.7 Fungus. - Equipment and assemblies shall show no evidence of viable fungus or corrosion* when subjected to 28 days of fungus test in accordance with paragraph 4.10.

*Corrosion is defined as any visible degradation that can be attributed to flaky, pitted, blistered, or otherwise loosened finish or metal surface.

3.9 Service conditions. - The equipment as specified shall meet the following service conditions:

3.9.1 Equipment non-operating. - Except as otherwise specified, the psychrometer shall function satisfactorily and shall meet the requirements of the drawings and documents referenced herein after subjection to any one, or combination of the following conditions:

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

3.9.1 Equipment non-operating (con't). -

(a) Temperature. - Exposure in the range of +160°F to -80°F; exposure at the high and low temperature extreme shall be 24 hours or to practical thermal equilibrium.

(b) Relative humidity. - Up to 97 percent relative humidity for 20 hours; and exposure at 100 percent relative humidity; with condensation, for 4 hours.

(c) Immersion. - The equipment when packed in its carrying case shall be capable of withstanding immersion in water to a depth of 6 inches above the top for a period of 30 minutes without evidence of leakage.

3.9.2 Equipment operating. - Except as otherwise specified herein, the psychrometer motor shall meet the requirements of this specification under any probable combination of the following:

(a) Operation. - Operation of equipment, including the psychrometers spring motor, shall withstand at least one year of field service.

(b) Temperature. - The psychrometer motor shall operate in the range of -75°F to +150°F.

(c) Elevation. - Any elevation up to 10,000 feet above sea level.

(d) Relative humidity. - Any relative humidity up to 95 percent.

3.10 Technical literature, tools, and running spare parts. - Technical literature, tools, and running spare parts shall be furnished if specified in the contract. Running spare parts shall be identical to corresponding parts in the equipment furnished on the order. (See 6.2).

3.11 Workmanship. - The equipment shall be manufactured and assembled in accordance with requirement 9 of MIL-STD-454 and the applicable portions of the following paragraphs:

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

3.11 Workmanship (con't). -

(a) In MIL-P-11268:

Cleaning (Equipment)
Parts
Units

General Requirements for Securing Parts

Plastic Material and Parts

Tropicalization of Materials

Adhesives

Rubber

Webbing

Screws, other Threaded Devices, and Related Parts

Nuts, Anchor and Clinch

Form and Fit (Screws, other Threaded Devices and Related Parts)

Riveting

Welding:
Cleaning Prior to Welding Process
Cleaning After Welding

Felt

Duck (Fabric)

(b) In this specification:

3.8.3	Carrying Case
3.8.6	Psychrometer Motor

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. - Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-17T14:12Z
Check the source to verify that this is the currentversion before use.

4.2 Classification of inspection. - Inspection shall be classified as follows:

(a) Preproduction inspection (does not include preparation for delivery). (See 4.3)

(b) Inspection covered by subsidiary documents. (See 4.4)

(c) Quality conformance inspection:

(1) Quality conformance inspection of equipment before preparation for delivery. (See 4.5)

(2) Quality conformance inspection of preparation for delivery. (See 4.13)

4.3 Preproduction inspection. - This inspection will be performed by the Government unless otherwise specified in the contract. It shall consist of the inspection specified in subsidiary documents covering the items listed in 4.4, and the inspection specified for Group A, and Group C (See Table I and II respectively).

4.4 Inspection covered by subsidiary documents. - The following shall be inspected under the applicable subsidiary documents as part of the inspection of equipment before preparation for delivery:

Item	Where required
Finish	3.5
Marking	3.6

4.5 Quality conformance inspection of equipment before preparation for delivery. - The contractor shall perform the inspection specified in 4.4 and 4.5.1 through 4.5.4. This does not relieve the contractor of his responsibility for performing any additional inspection which is necessary to control the quality of the product and to assure compliance with all specification requirements. The Government will review and evaluate the contractor's inspection procedures and examine the contractor's inspection record. In addition the Government--at its discretion--may perform all or any part of the specified inspection, to verify the contractor's compliance with specified requirements. (See 6.4) Test equipment for Government verification inspection shall be made available by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

4.5.1 Group A inspection. - This inspection, including sampling, shall conform to Table I and the ordinary inspection procedures of MIL-STD-105. Group A inspection shall be performed in any order which is satisfactory to the Government.

TABLE I. - Group A Inspection

Inspection	Req Para	Insp Para	AQL Major	AQL Minor
Visual and mechanical	3.11	4.11	4.0%	6.5%
Interchangeability	3.7	4.12	4.0%	6.5%

4.5.2 Group B inspection. - Group B inspection not required.

4.5.3 Group C inspection. - This inspection shall consist of the test specified in Table II and shall be performed on sample units that have been subjected to and met Group A inspection. Sample unit shall be selected in accordance with para 4.5.3.1.

TABLE II. - Group C Inspection

Inspection	Req Para	Insp Para
Temperature	3.9.1(a) & 3.9.2(b)	4.6
Moisture Resistance	3.9.1(b) & 3.9.2(d)	4.7
Immersion	3.9.1 (c)	4.8
Altitude	3.9.2(c)	4.9
Fungus	3.8.7	4.10

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

4.5.3.1 Sampling for inspection. - Two sample units (other than preproduction samples) shall be selected from the first 50 items produced and one from each 50 items produced thereafter. The items shall be selected without regard to their quality and shall be selected at random from each group of units produced. The two preproduction samples shall be subjected to the inspection of 4.3.

4.5.3.2 Noncompliance. - The contractor shall immediately report in writing each Group C failure occurrence, including details of the failure and characteristics affected. The contractor shall immediately investigate the cause of failure and further report the results of investigation and details of the proposed corrective action on (1) the process and materials, as applicable, and, (2) all units of product which were manufactured under the same conditions and which the Government considered subject to the same failure. Reports shall be forwarded to the responsible technical activity designated in the contract through the Quality Assurance Representative. After corrective action has been taken, additional sample units shall be subjected to Group C inspection (all inspections, or the inspections which the sample failed, at the option of the Government) and Group A may be reinstituted; however, final acceptance and shipment will be withheld until the Group C reinspection results have shown that the corrective action was effective.

4.5.4 Disposition of Group C sample units. - Unless otherwise specified, sample units which have been subjected to and passed Group C inspection may be delivered on the contract or order.

4.6 Temperature test. - The psychrometer shall be subjected to the temperature cycle shown in MIL-STD-169, except the performance test shall be in accordance with the following:

(a) Take performance measurements of the psychrometer motor to establish a standards of performance before placing it in the temperature chamber (when fully wound the psychrometer motor shall operate smoothly for 15 minutes).

(b) During the temperature cycle, make performance checks of the psychrometer motor at +150 ±5°F and at -40 ±5°F.

(c) Reduce the temperature to -75 ±5°F; maintain for 6 hours. Check the psychrometer motor for performance as compared to step (a).

(d) After the cycle has been completed, repeat the performance measurements taken in step (a), repeatability of performance should be attained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

4.7 Moisture-resistance test. - The psychrometer shall be subjected to the humidity test described in the following subparagraphs:

4.7.1 Test procedure. - The psychrometer will be tested as follows:

(a) Measure performance as specified in para 4.6(a) to establish standards of performance before placing the psychrometer in humidity chamber.

(b) Dry at 130 ±5°F for 24 hours.

(c) Condition at 77 ±5°F and 40 to 50 percent relative humidity for 24 hours.

(d) Subject the components to continuous cycling for two 48-hour cycles. Detail requirements for relative humidity, temperature, and time for each cycle shall be in accordance with MIL-STD-170.

4.7.2 Performance. - Upon completion of the cycling, the psychrometer motor shall meet the performance requirements of this specification.

4.8 Immersion test. - The equipment as prepared for field transportation, except for the duck bags, shall be immersed in fresh water to a minimum depth of 6 inches above the top for one half (1/2) hour. Immediately prior to immersion, the temperature of the equipment shall be 40°F, or more, above the temperature of the water. The tank in which the equipment is immersed shall be of sufficient capacity to maintain the water within ±2°F of its initial temperature, or the temperature of the water shall be maintained within those limits by other means. After completion of the period of immersion, the equipment shall be removed from the water and wiped dry on exterior surfaces. When the equipment is opened, there shall be no evidence of leakage.

4.9 Elevation test. - The psychrometer placed in the altitude chamber shall be tested as follows:

(a) Operate the psychrometer motor to establish its proper performance by winding the spring and allowing it to run down.

Source: https://assist.dla.mil -- Downloaded: 202610-17T14:12Z
Check the source to verify that this is the current version before use.

4.9 Elevation test (con't). -

(b) Reduce the barometric pressure to 20.6 ±0.1 inches of mercury (to simulate 10,000 feet above sea level) and after temperature and pressure stabilization has been achieved, the performance of the psychrometer shall meet the requirements of this or any referenced subsidiary specifications.

(c) Return the chamber to ambient pressure and insure that satisfactory performance of the psychrometer motor is achieved.

4.10 Fungus test. - The equipment and assemblies shall be subjected to 28 days of fungus test specified in method 508 of MIL-STD-810. For gasket or other non-hermetically sealed equipments, the parts and materials shall be sprayed with the spore suspension. The interior of the case shall also be sprayed. The equipment shall be subjected to the test in this open condition.

4.11 Visual and mechanical inspection. - Parts and equipment shall be inspected for workmanship, mechanical fit, loose nuts and screws, and miscellaneous defects. Fastening devices shall be inspected for mechanical operation. Finishes shall be inspected for workmanship. Clearances and dimensions shall be measured. Markings shall be inspected for conformance with the applicable drawings. Examination also may be made for other visual or mechanical defects, similar to those described above, that are contrary to specified requirements for the equipment. Visual and mechanical defects shall be classified by the Government as major or minor in accordance with the definitions of MIL-STD-105.

4.12 Inspection for interchangeability. - The external and internal dimensions of cases, covers, and insertable assemblies, when such dimensions affect mating of parts and location of hinges and fasteners on separable parts or assemblies which must mate shall be gaged or measured to determine conformance to physical interchangeability requirements of 3.7. When a listed dimension is not within specified or design limits, it shall be considered a major defect.

4.13 Quality conformance inspection of preparation for delivery. - Preparation for delivery shall be inspected in accordance with MIL-P-116 to determine conformance to the requirements of section 5.

5. PREPARATION FOR DELIVERY

Source: https://assist.dla.mil -- Downloaded: 202640-17T14:12Z
Check the source to verify that this is the currentversion before use.

5.1 Preservation and packaging. - Preservation and packaging shall be Level A or C, as specified (See 6.2(c)).

5.1.1 Level A. -

5.1.1.1 Cleaning. - Components of Meteorological Station, Manual AN/PMQ-1( ) shall be cleaned in accordance with process-C-1 of MIL-P-116.

5.1.1.2 Drying. - Components of Meteorological Station, Manual AN/PMQ-1( ) shall be dried in accordance with the applicable procedure of MIL-P-116.

5.1.1.3 Preservation application. - None required.

5.1.1.4 Unit packaging. - Unit packaging shall be in accordance with methods prescribed in MIL-P-116 as specified herein.

5.1.1.4.1 Case CY-952/PMQ-1. - Each case shall be packaged method III as follows: close case and secure all fastenings. Cushion each case on all surfaces with cells or pads, or both fabricated of fiberboard conforming to PPP-F-320, grade W6c, designed to protect all projections and absorb the shock of impact in handling and transit. Place the cushioned case within a close-fitting fiberboard box conforming to PPP-B-636, W6c. Close the box in accordance with the appendix of the box specification. In addition, seal all seams and joints with tape, not less than two inches wide, conforming to PPP-T-76.

5.1.1.4.2 Case CY-1285( )/PMQ-1. - Each case shall be packaged method III as follows: fold each case into a bundle of proportionate dimension and tie in two places with cotton twine conforming to T-T-871. Place the bundled case within a close fitting fiberboard box conforming to PPP-B-636, W6c. Close box in accordance with the appendix of the box specification.

5.1.1.4.3 Intermediate package. - A quantity of 10 each Case CY-1285( )/PMQ-1 packaged as specified in paragraph 5.1.1.4.2 shall be arranged in a most compact manner within a close fitting box conforming to PPP-B-636, W5c. Close box in accordance with the appendix of the box specification. In addition, seal all seams and joints with tape not less than two inches wide, conforming to PPP-T-76.

Source: https://assist.dla.mil -- Downloaded: 202610-17T14:12Z
Check the source to verify that this is the current version before use.

5.1.1.4.4 Gauge ML-435( )/PMQ-1 and Psychrometer ML-436( )/PMQ-1 shall be individually packaged method III as follows: cushion each item by wrapping in cellulosic material conforming to PPP-C-843. Place each cushioned item within a close fitting fiberboard box conforming to PPP-B-636, W6c. Close box in accordance with the appendix of the box specification.

5.1.1.4.5 Intermediate package. - Gauge ML-435( )/PMQ-1 and Psychrometer ML-436( )/PMQ-1 packaged as specified in paragraph 5.1.1.4.4, like items shall be intermediate packaged in accordance with paragraph 5.1.1.4.3.

5.1.1.4.6 Syringe Plastic, Thermometer Support and Tripod Anemometer MT-869( )/PMQ-1 shall be individually packaged method III as follows: cushion each item by wrapping in paperboard conforming to PPP-P-291, place each cushioned item within a close fitting fiberboard box conforming to PPP-B-636, W6c. Closure shall be in accordance with the appendix of the box specification.

5.1.1.4.7 Intermediate package. - Syringe Plastic, Thermometer support and Tripod Anemometer MT-869( )/PMQ-1 packaged as specified in paragraph 5.1.1.4.6, like items shall be intermediate packaged in accordance with paragraph 5.1.1.4.3.

5.1.2 Level C. - Components of Meteorological Station, Manual AN/PMQ-1( ) shall be individually packaged in a manner which will afford adequate protection against corrosion, deterioration and damage during shipment from the supply source to the first receiving activity.

5.2 Packing. - Packing shall be Level A, B or C as specified (See 6.2(c)).

5.2.1 Level A. - A quantity of items packaged as specified in paragraph 5.1 bearing the same stock number shall be packed in close fitting box conforming to PPP-B-636, type CF, class weather-resistant, special requirements or MIL-B-43291. Closure shall be in accordance with the appendix of the applicable box specification. The shipping containers shall be reinforced by pressure-sensitive filament tape banding conforming to PPP-T-97, type IV in accordance with the appendix of the applicable box specification.

5.2.2 Level B. - A quantity of items packaged as specified in 5.1 bearing the same stock number, shall be packed in accordance with 5.2.1.

Source: https://assist.dla.mil -- Downloaded: 202640-17T14:12Z
Check the source to verify that this is the currentversion before use.

5.2.3 Level C. - A quantity of components of Meteorological Station, Manual AN/PMQ-1( ) packaged as specified in paragraph 5.1.2 bearing the same stock number shall be packed in shipping containers in a manner that will afford adequate protection against damage to the package and its contents during shipment from the supply source to the first receiving activity. Shipping containers shall comply with the rules and regulations of the common carrier as applicable to the mode of transportation.

5.3 Marking. - In addition to any special marking required by the contract or order, interior packages and exterior shipping containers shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. - The equipment consists of a group of portable meteorological items used at locations where it is impractical to establish weather stations.

6.2 Ordering data. - Procurement documents should specify the following:

(a) Title, number, and date of this specification and any amendments thereto.

(b) Type required.

(c) Levels of preservation, packaging and packing (See section 5).

(d) Whether preproduction samples are required.

(e) If rough handling and functional test is required.

(f) Place of final inspection.

(g) Technical literature required.

(h) Quantity of tools and running spare parts.

6.3 Nomenclature. - The parentheses in the nomenclature will be deleted or replaced by a letter identifying the particular design; for example: AN/PMQ-1W. The contractor should apply for nomenclature in accordance with the applicable clause in the contract. (See 1.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

6.4 Verification inspection. - Verification by the Government will be limited to the amount deemed necessary to determine compliance with the provisions established in this specification and contract.

CUSTODIAN:
ARMY EL

REVIEW ACTIVITIES:
AF 71, 13

PREPARING ACTIVITY:
ARMY EL
PROJECT NO. 6660-A126

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION
MIL-M-10530B(EL) METEOROLOGICAL STATION, MANUAL AN/PMQ-1( ) (COMPONENTS OF)

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*	DATE

DD FORM 1 OCT 64 1426

ESC-FM 2011-66

To detach this form, cut along this line

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T14:12Z
Check the source to verify that this is the current version before use.