MIL-P-28785(EC)
Amendment 2
29 August 1973
SUPERSEDING 1/
AMENDMENT 1
10 August 1973

MILITARY SPECIFICATION

POWER SUPPLY, VARIABLE OUTPUT (0 to 50 VOLTS, 0 to 5 AMPERES)

This amendment forms a part of Military Specification MIL-P-28785(EC), 31 July 1972.

Page 1

2.1 Add:

MIL-STD-461 - Electromagnetic Interference Characteristics, Requirements for Equipment

MIL-STD-462 - Electromagnetic Interference Characteristics, Measurement of

Page 2

Add new paragraphs 3.5.4 and 3.5.4.1:

"3.5.4 Electromagnetic compatibility. The units shall meet the requirements of MIL-STD-461 for Class IC, non-antenna C/E equipment.

"3.5.4.1 Radiated susceptibility. RS03 susceptibility test limits shall be in accordance with MIL-STD-461, except that over the frequency ranges 30-42 MHz and 132-150.8 MHz, the test field intensity shall be 100 volts/meter."

3.6.4: Delete and substitute:

"3.6.4 Power drain. When tested in accordance with 4.5.2.1 the primary current to the power supply shall not exceed 6 amperes ac when operating under full load of 50 volts and 5 amperes."

3.6.5: Delete and substitute:

"3.6.5 Line regulation. Line regulation shall be 0.01 percent +2 millivolts (mV) with an ac line variation of 105 to 125 Vac when tested in accordance with 4.5.2.5"

1/ CHANGES FROM PREVIOUS ISSUE. THE OUTSIDE MARGINS OF THIS DOCUMENT HAVE BEEN MARKED "#" TO INDICATE WHERE CHANGES (DELETIONS, ADDITIONS, AND SO FORTH) FROM THE PREVIOUS ISSUE HAVE BEEN MADE. THIS HAS BEEN DONE AS A CONVENIENCE ONLY AND THE GOVERNMENT ASSUMES NO LIABILITY WHATSOEVER FOR ANY INACCURACIES IN THESE NOTATIONS. BIDDERS AND CONTRACTORS ARE CAUTIONED TO EVALUATE THE REQUIREMENTS OF THIS DOCUMENT BASED ON THE ENTIRE CONTENTS AS WRITTEN IRRESPECTIVE OF THE MARGINAL NOTATIONS AND RELATIONSHIP TO THE LAST PREVIOUS ISSUE.

FSC 5820

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T15:18Z
Check the source to verify that this is the current version before use.

Page 4

3.7.1.2, item (g): Delete and substitute:

"(g) Over-voltage indicator light - The over-voltage indicator light shall light when the output voltage exceeds the voltage setting by one volt.

Page 8

Add new paragraph 4.5.1.4:

"4.5.1.4 Electromagnetic compatibility test. Compliance with the requirements of 3.5.4 shall be verified by test performed in accordance with MIL-STD-462."

4.5.2.1, third sentence: Delete and substitute:

"Attach a voltmeter and ammeter of suitable accuracy to establish conformance to 3.6.4 to output terminals."

4.5.2.2, first sentence: Delete "Using a voltmeter having an accuracy of 1/2 percent" and substitute "Using a voltmeter and ammeter of suitable accuracy to establish conformance to 3.6.8,"

4.5.2.5, line 2: Delete "having an accuracy of 1/2 percent" and substitute "and ammeter of suitable accuracy to establish conformance to 3.6.5"

4.5.2.6, lines 3 and 4: Delete "1/2 percent accuracy meter." and substitute "a voltmeter of suitable accuracy to establish conformance to 3.6.6."

Preparing activity:
Navy - EC

(Project 6820-N623)

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T15:18Z
Check the source to verify that this is the current version before use.