RR-T-636D
January 28, 1976
SUPERSEDING
RR-T-00636C
January 10, 1966

# FEDERAL SPECIFICATION

## TRAYS, INSTRUMENT, CORROSION-RESISTING STEEL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for use of all Federal agencies.

## 1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers corrosion-resisting steel instrument trays, suitable for hospital and medical use.

1.2 Classification. Corrosion-resisting steel instrument trays shall be of the following types and sizes, as specified (see 6.2):

Type I - With cover.
Size 1 - 8-7/8 by 5 by 2 inches.
Type II - Without cover.
Size 2 - 10-1/2 by 8 by 2 inches. As an alternate, 10 by 8-1/4 by 2-1/2 inches.
Size 3 - 15-1/2 by 9-1/2 by 2 inches.
Size 4 - 19-1/4 by 12-3/4 by 3/4 inches.
Size 5 - 20-5/8 by 12-5/8 by 4 inches.
Size 7 - 15-1/8 by 10-1/2 by 5/8 inches.
Type III - Perforated, Without cover.
Size 6 - 16-3/4 by 9-1/4 by 2-1/2 inches.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

## 2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issues in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

Federal Specifications:

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.

Federal Standards:

Fed. Std. No. 66 - Steel: Chemical Composition and Hardenability.
Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).
Fed. Test Method Std. No. 151 - Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)



Source: https://assist.dla.mil -- Downloaded: 202640-26T12:11Z
Check the source to verify that this is the currentversion before use.

Military Specifications:

MIL-L-10547 - Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

Military Drawings:

Defense Personnel Support Center Drawings

20019 - Tray, Instrument, Corrosion-Resisting Steel, 19-1/2 by 12-3/4 by 3/4 inch.
20460 - Trays, Instrument, Corrosion-Resisting Steel.
20686 - Tray, Instrument, Corrosion-Resisting Steel, 8-7/8 by 5 by 2 inches.
24019 - Trays, Instruments, Corrosion-Resisting Steel, Perforated.

(Copies of Military Specifications, Standards, and Drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Trays shall be constructed of corrosion-resisting steel, conforming to class 302 or 304 of Fed. Std. No. 66, and shall meet the requirements for chemical composition specified therein when tested as specified in 4.4.1.



Source: https://assist.dla.mil -- Downloaded: 202640-26T12:11Z
Check the source to verify that this is the current version before use.

3.2 Design and construction.

3.2.1 Body. The body of the tray shall be seamlessly die formed from a single piece of corrosion-resisting steel, and shall rest evenly on a flat surface without rock or wobble. In overall dimensions, the tray shall be square within ± 1/64 inch. Thickness of metal used, and the dimensions for each size, shall be as indicated in table I.

TABLE I. Body

| Size | Original metal thickness (inch) | Final thickness, not less than (inch) | Dimensions |
|---|---|---|---|
| 1 | .031 No. 22 United States Standard (U.S.S.) gage | .024 | Drawing 20686 |
| 2 | .031 (No. 22, U.S.S. gage) | .027 | Drawing 20460 |
| 3 | .044 (No. 19, U.S.S. gage) | .038 | Drawing 20460 |
| 4 | .037 (No. 20, U.S.S. gage) | .033 | Drawing 20019 |
| 5 | .050 (No. 18, U.S.S. gage) | .040 | Drawing 20460 |
| 6 | .037 (No. 20, U.S.S. gage) | .033 | Drawing 240[illegible] |
| 7 | .037 (No. 20, U.S.S. gage) | .033 | Drawing 20460 |

3.2.2 Cover. Cover for Size 1 shall be seamlessly die formed of corrosion-resisting steel, not less than .031 inch thick (No. 22, U.S.S. gage), which shall not be reduced below .024 inch at any point in the finished cover. Design and dimensions shall be as indicated on Drawing 20686. The rim of the cover shall be formed to fit snugly over the bead on the body of the tray, and shall have a lift handle of the design shown on Drawing 20686 made of corrosion-resisting steel not less than .044 inch thick (No. 19, U.S.S. gage), which shall not be reduced below .038 inch at any point in the finished handle. The lift handle shall be properly centered and spot-welded to the top of the cover. Cover may be of either design as shown on Drawing 20686. In overall dimensions, the cover shall be square within ± 1/64 inch. The cover shall not rock on the body of the tray.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

3.2.3 Perforations. The perforations in the Type III, Size 6 instrument tray shall be as shown on Drawing 24019.

3.3 Finish.

3.3.1 Grinding. The entire inside and outside surfaces of the tray and cover shall be finished with a No. 4 or better finish after all surface imperfections have been removed by previous grinding. This last grinding shall produce a uniform finished surface with all imperfections and marks of previous grinds removed. The thickness of the sheet forming the tray or cover shall not be reduced below the thickness specified in 3.2.1 and 3.2.2 by grinding to remove surface imperfections.

3.3.2 Alternate surface finish. As an alternate, the tray and cover shall be finished by electropolishing. The electropolishing shall result in a uniform, bright, and shiny surface which can be readily cleaned.

3.3.3 Cleaning. As a final operation, the tray and cover shall be thoroughly cleaned to remove foreign particles such as dirt, grease, graphite, and metal.

3.3.4 Resistance to corrosion. The finished trays and covers shall pass the salt-spray test specified in 4.4.2. Slight surface corrosion of the specimen which can be easily washed off shall not be considered as cause for rejection; but if the specimen should rust or show continuous bleeding from pits, it shall be considered as having failed the test.

3.4 Identification marking. Each tray shall be marked with the name or registered trademark of the manufacturer. Impressions shall be made on the bottom of the trays at even pressure so as not to penetrate the inside of the tray. Letters shall be not less than 1/4 inch high.

3.5 Workmanship. Workmanship shall be first class throughout. Trays shall be free from defects which detract from their appearance or may impair their serviceability, and shall be free of rough grind marks, trimming marks, pits, scales, and other defects or imperfections. All surfaces shall have a smooth finish and shall be free of burs or sharp edges.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

## 4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to the prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table II.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for examination

| | Inspection level | AQL (defects per hundred units) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 1.0 |
| Total (Major and minor defects combined) | II | 4.0 |
| For dimensional examination | S-1 | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table III. Not less than three samples shall be taken, regardless of lot size. Acceptance number shall be zero for all sample sizes.

TABLE III. Sampling for tests

| Component (unit of product) | Characteristic | Requirement | Test procedure | Inspection level |
|---|---|---|---|---|
| Tray | Chemical composition | 3.1 | 4.4.1 | 1/ |
| Tray | Corrosion-resistance | 3.3.4 | 4.4.2 | S-1 |

1/ Sample as specified in 4.2.2.1.

4.2.2.1 Raw material. Four ounces of each lot of metal used in the construction of the trays shall be selected for chemical analysis as specified in 4.4.1.

4.3 Examination. The trays shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed on table IV. Examination shall not be restricted to the classified possible defects listed below.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

TABLE IV. Classification of defects

| Categories | Defects |
|---|---|
| Major | |
| 101 | Tray not free of cracks, tears, holes, ripples. |
| 102 | Edges not free of burs or sharpness. |
| 103 | Tray rocks, wobbles, or does not rest evenly on a flat surface. |
| 104 | Cover rocks on tray (size 1 only). |
| Minor | |
| 201 | Finish not free of scratches or mars. |
| 202 | Tray not free of grease, oil, dirt, chips, or other foreign matter. |
| 203 | Finish not as specified. |

4.3.2 Dimensional examination. The trays shall be examined for defects in dimensions. Any dimension not within tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Tests for chemical analysis shall be conducted in accordance with the Chemical Analysis or the Spectrochemical Analysis test requirements of Fed. Test Method Std. No. 151.

4.4.2 Salt-spray. Salt-spray test shall be conducted in accordance with the Salt Spray Test requirements of Fed. Test Method Std. No. 151.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing and marking comply with the requirements of section 5. Defects shall be scored as specified in table V. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 expressed in units of percent defective.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

TABLE V. Examination for preparation for delivery

| Examination | Defect |
|---|---|
| Containers | Not as specified. |
| Quantity | Quantity in interior packages and shipping containers not as specified. |
| Packaging and Packing Materials | Any component missing or damaged. |
| Workmanship | Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted. |
| Markings | Omitted, incorrect, illegible, improper size, location, sequence, or method of application. |

4.5.1 Examination of closure, waterproofing and banding of container. When the shipping container is required to comply with PPP-B-636, the examination for defects in closure, waterproofing or banding shall be in accordance with the appendix of the box specification.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each tray, of the type and size specified, shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, class domestic. Box selected shall be within the size and weight limitation of the box specification. Tray and cover shall be wrapped individually with non-corrosive paper prior to packaging in the box, when trays are furnished with covers. Closure of boxes shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which shall afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. The number of trays as specified in table VI for the type and size shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

5.2.2 Level B. The number of trays as specified in table VI for the type and size shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

TABLE VI. Level B (types and sizes)

| Type | Size | Quantity per exterior container |
|---|---|---|
| I | 1 | 48 |
| II | 2 | 48 |
| II | 3 | 36 |
| II | 4 | 24 |
| II | 5 | 12 |
| II | 7 | 6 |
| III | 6 | 6 |

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with Fed. Std. No. 123.

5.3.2 Military activities.

5.3.2.1 Unit package and exterior container. Each unit package and exterior container shall be marked in accordance with MIL-STD-129.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. Trays are intended for use with the equipment listed below:

| | |
|---|---|
| Size 1 | To provide a suitable container for dressing sets which are placed on dressing carriages. |
| Sizes 2 and 3 | To hold instruments in hospital operating rooms and medical offices. |
| Size 4 | As a supplementary or replacement item for the tray provided as a component of STAND, SURGICAL INSTRUMENT, NSN 6530-00-027-5252, 6530-00-551-8681, or 6530-00-708-1590 or TABLE, ANESTHETIST'S NSN 6530-00-708-8425. |
| Size 5 | To hold sterile obstetrical forceps, orthopedic mallets, and similar bulky surgical items used in operating and delivery rooms. |
| Size 6 | As a container for instruments during sterilization in standard sterilizers. |
| Size 7 | For use with CABINET, DENTAL INSTRUMENT AND SUPPLIES NSN 6520-00-782-6499, 6520-00-890-1825, 6520-00-890-1831, or 6520-00-890-1832, DENTAL OPERATING UNIT NSN 6520-00-782-2674, and INSERT DRAWER, DENTAL INSTRUMENT AND SUPPLIES CABINET NSN 6520-00-890-1850. |

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

(a) Title, number, and date of this specification.
(b) Type and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following corrosion-resisting steel trays:

| National Stock Number | Tray Identification |
|---|---|
| | Type I - With cover |
| 6530-00-793-9600 | Size 1 - 8-7/8 by 5 by 2 inches |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

| National Stock Number | Tray Identification |
|---|---|
| | Type II - Without cover |
| 6530-00-793-9945 | Size 2 - 10-1/2 by 8 by 2 inches |
| 6530-00-794-0000 | Size 3 - 15-1/2 by 9-1/2 by 2 inches |
| 6530-00-793-9570 | Size 4 - 19-1/4 by 12-3/4 by 3/4 inches |
| 6530-00-794-0300 | Size 5 - 20-5/8 by 12-5/8 by 4 inches |
| 6530-00-926-9170 | Size 7 - 15-1/8 by 10-1/2 by 5/8 inches |
| | Type III - Perforated, without cover |
| 6530-00-914-0238 | Size 6 - 16-3/4 by 9-1/4 by 2-1/2 inches |

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available but is intended to cover the types which are normally procured to meet requirements.

MILITARY CUSTODIANS:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Civil Agency Coordinating Activities:
VA - DMS
GSA - FSS
HEW-HSM

Project No. 6530-1358

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See section 2 of this specification to obtain extra copies and other documents reference herein. Price 35 cents each.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
RR-T-636D TRAYS, INSTRUMENT, CORROSION-RESISTING STEEL

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

**DD FORM 1 JAN 72 1426** REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974-- 603-680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:11Z
Check the source to verify that this is the current version before use.