MILITARY SPECIFICATION

OVEN, THERMAL DRYING, ELECTRIC

1. SCOPE

1.1 Scope. This specification covers general purpose, electrically heated ovens for use in various heating, drying, and baking operations.

1.2 Classification.

1.2.1 Types, classes and styles. The ovens shall be of the following types, classes and styles as specified (see 6.2).

1.2.1.1 Types.

Type I Gravity/convection air circulation
Type II Forced air circulation

1.2.1.2 Classes. The oven shall be classified by maximum operating temperature as follows:

Class 1 As specified in Ordering Data (see 6.2)
Class 2 500 Degrees F
Class 3 650 Degrees F

1.2.1.3 Styles.

Style 1 Single door
Style 2 Double door

1.2.1.4 Size. Oven size shall be considered to be the free and clear dimensions within the oven, wall to wall, floor to ceiling, door to back, as specified in 6.2.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein.

STANDARDS

Military

MIL-STD-130	Identification Marking of U.S. Military Property
MIL-STD-882	System Safety Program, for Systems and Associated Subsystems and Equipment: Requirements for

Source: https://assist.dla.mil -- Downloaded: 20261O-23T09:56Z
Check the source to verify that this is the current version before use.

MIL-STD-889 Dissimilar Metals

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

NATIONAL BUREAU OF STANDARDS

H-28 Screw Thread Standards for Federal Services

(Application for copies should be addressed to the Superintendent of Documents, U. S. Government Printing Office, Washington, DC 20402.)

NATIONAL FIRE PROTECTION ASSOCIATION STANDARDS

NFPA NO 86.A Ovens and Furnaces
National Electrical Code

(Application for copies should be addressed to the National Fire Protection Association, 60 Batterymarch Street, Boston, MA 02110.)

NATIONAL ELECTRICAL MANUFACTURERS' ASSOCIATION STANDARDS

NEMA IC-1 Industrial Controls
NEMA MG-1 Motors and Generators

(Application for copies should be addressed to the National Electrical Manufacturers' Association, 155 East 44th Street, New York, NY 10017.)

3. REQUIREMENTS

3.1 First article. This specification includes provisions for first article testing.

3.2 Design. The oven shall be an upright, double-walled, fully insulated, electrically heated, automatically controlled, completely self-contained design of the type, class, style and size specified in 6.2.

3.2.1 Safety. The oven shall be provided with safety features and devices for the protection of personnel, equipment and the work being processed. Parts which are hazardous to personnel shall be guarded. Equipment and components shall be selected and applied which are "fail-safe," when possible. Where not possible, personnel and equipment shall be separately protected against hazardous failures and conditions. Equipment shall be analyzed for category III and category IV hazards as defined in MIL-STD-882. Any hazard so identified shall be brought to the attention of the procuring agency with recommended action for satisfying safety requirements. Action order of precedence shall be satisfied in accordance with MIL-STD-882, para 5.6.

NOTICE 1

Source: https://assist.dla.mil -- Downloaded: 20261O-23T09:56Z
Check the source to verify that this is the current version before use.

3.2.2 Material. All material used in the fabrication of the oven and related equipment shall be of a quality necessary to produce an oven to meet the requirements herein. Materials that are exposed to the internal environment of the oven work space shall be compatible with the atmosphere and temperature conditions encountered.

3.2.2.1 Protective treatment. When materials are used in the construction of the oven which are subject to deterioration when exposed to climatic and environmental conditions likely to occur during service usage, they shall be protected against such deterioration in a manner that will in no way prevent compliance with the performance requirements set forth in the specification. The use of dissimilar metals shall conform to the requirements of MIL-STD 889.

3.2.3 Lubrication. Provisions shall be made for the lubrication of all moving parts except non-additive, pack-seal bearings, to reduce wear and maintain efficiency. Lubrication points shall be readily accessible.

3.2.4 Accessibility. All parts subject to wear or breakage and all parts which require periodic maintenance shall be readily accessible for adjustment and replacement.

3.2.5 Interchangeability. All parts shall be manufactured to standards which permit replacement without modification of parts or equipment.

3.3 Construction. The oven shall be constructed of parts which are free of harmful defects and repairs. Welding shall be employed only where specified in the original design; fastening devices shall be installed in a manner to prevent change of tightness. All surfaces shall be cleaned and free of harmful or extraneous materials. All edges shall be either rounded or beveled unless sharpness is required to perform a necessary function. Gaps between mating parts conducive to heat losses shall be avoided. The structure shall be capable of withstanding without damage normal forces encountered during transportation/delivery and operation of the oven at its maximum capacity.

3.3.1 Painting. The oven exterior shall be cleaned, primed and painted in accordance with the manufacturer's standard commercial practice.

3.3.2 Threads. All threaded parts shall conform to NBS Handbook H-28.

3.4 Components and systems.

3.4.1 Air circulation.

3.4.1.1 Type I. Air shall circulate by gravity/convection in patterns which shall promote even heat distribution throughout the working chamber.

3.4.1.2 Type II. A fan or blower shall provide air movement within the oven to promote even heat distribution throughout the working chamber and provide means for accurate temperature control. The temperature variation between any two points within the oven shall be not greater than 10 degrees F.

ON MS27089
Source: https://assist.dla.mil -- Downloaded: 202640-23T09:56Z
Check the source to verify that this is the current version before use.

3.4.2 Cabinet. The cabinet shall be double-walled construction. The minimum outside dimensions of the cabinet shall be consistent with the required thickness of the thermal insulation between the walls to ensure minimal heat loss. Construction may be either welded or securely fabricated with sheet metal screws, bolts and lock nuts or a combination. The cabinet shall provide space for all wiring, heating elements, controls and for type II ovens, a forced air circulating system.

3.4.3 Inner surfaces. The inner walls of the oven, including the sides, back, top, bottom and air ducts, defining the entire heated surface shall be fabricated from aluminized steel, type I, or other compatible steel of, at least, equivalent corrosion resistance quality, not less than 20 gage.

3.4.4 Outer walls. The outer walls of the cabinet shall be fabricated from steel of not less than 18 gage. Construction of any seams shall cause no buckling or bulging after fabrication.

3.4.5 Structural bracing. The unexposed faces of the inner and outer walls may be either preformed or braced to provide stiffness. Reinforcing shall be provided at points of attachment of hinges and other load supporting fittings. The entire oven shall be sufficiently braced to provide a rigid structure even with the door(s) open. The bottom of walk-in ovens shall be reinforced to withstand, at least, 150 pounds per square foot. Oven bottoms, other than walk-in ovens, shall be of sufficient strength to withstand a concentrated 50 pound load anywhere on the oven bottom.

3.4.6 Ventilation. For type II ovens, the ventilators or openings shall be provided in the walls or top of the cabinet for the entrance and exit of fresh air. The location, size and design shall provide for a minimum of approximately one complete air change per minute in the oven, except when otherwise specified in 6.2. The ventilators shall have adjustable dampers, manually controlled from the outside of the oven. Provisions shall be made for connection of the exhaust ventilator opening to an exhaust duct. The oven ventilation air requirements shall conform with NFPA No 86.A, Chapter IV, Articles 400 through 420.

3.4.7 Instrument penetration. There shall be a penetration through the walls of the cabinet, near the top or through the walls of the cabinet, for the insertion inside the oven of temperature sensing element(s), other than those which are part of the normal oven instrumentation. The pipe or tubing penetration shall be sealed to and securely attached to the interior and exterior walls of the oven and shall be fitted with a plug or cap at the outside exposed end, to effectively seal the penetration when it is not in use.

3.4.8 Shelf supports. Shelf supports shall be attached to the right and left walls of the interior of the oven. The supports shall permit shelves to be spaced at vertical intervals which are adjustable from top to bottom in accordance with the requirements specified in paragraph 6.2. Shelf supports shall be made from at least the same quality steel as used for the oven interior. Unloaded shelves shall remain horizontal when withdrawn approximately one-half the depth of the cabinet chamber.

Source: https://assist.dla.mil -- Downloaded: 2026-0-23T09:56Z
Check the source to verify that this is the current version before use.

3.4.9 Cabinet support. Unless otherwise specified in para 6.2, the cabinet shall be constructed in such a manner as to permit forklifting during installation or subsequent movement. The supports shall be of sufficient strength to support the oven assembly and the work load. If oven bottom insulation is not provided, the underside of the oven shall be raised several inches off the floor/cabinet to permit free air circulation beneath the oven.

3.4.10 Doors. A side hinged single door (Style 1) or double doors (style 2) as specified in para 6.2, of double wall construction similar to the cabinet shall be provided and shall permit full width and height access to the working area when open. The door(s) when shut shall be close fitting within the door opening to minimize heat loss and shall seat tightly against a suitably resilient seal around the edge to prevent air leakage. Door(s) shall be reinforced at attachment points of hinges, handles, and latches. Suitable stops shall be provided, if required, to prevent damage or interference with other oven parts or components when the doors are swung open. Walk-in ovens shall have door(s) which open safely and easily from both the outside and the inside.

3.4.11 Hardware. Door hinges shall be of the heavy duty type, painted to match the cabinet, if not otherwise protected against corrosion. Latches shall be a "snap-action" type which provide positive and fast release of the closed door(s).

3.4.12 Thermal insulation. Thermal insulation shall be heat-resisting material which is non-combustible, non-hydroscopic, and non-powdering. Powdered, shredded or granular insulation shall not be used. The insulation shall be precompressed or bonded together to prevent settling. As installed, the insulation shall be not less than three inches thick and shall be sufficient to limit the heat loss of the oven to no more than 0.050 KW (171 BTU) per hour per square foot of oven interior surface as determined from the size dimensions of 1.2.1.4. The insulation shall not support any load except its own weight and shall be walled off by steel ducts where openings from the inner to outer walls are necessary.

3.4.13 Shelves. Shelves for other than walk-in ovens, shall be approximately the same size as the width and depth of the oven working chamber. The number required for each oven shall be as specified in 6.2. When shelves or partial shelves are required for walk-in ovens, the type, size, number and location shall be as specified in 6.2. All shelves and partial shelves shall be fabricated from expanded mesh, corrosion-resistant steel. Shelves shall be easily removable from the oven. The shelves shall not deflect more than one fourth inch or receive any permanent deformation when tested in conformance with 4.6.1.

3.4.14 Temperature indicator. A remote dial type temperature indicator is required. The instrument shall be graduated in degrees F, approximately five feet six inches above ground level and grouped with other instruments and controls. The thermometer may be an individual unit or combined with the thermostat in 3.4.16.

Source: https://assist.dla.mil -- Downloaded: 202640-23T09:56Z
Check the source to verify that this is the currentversion before use.

3.4.14.1 Operating temperature range. The operating temperature range of the oven shall be as specified in 6.2.

3.4.15 Heating units. Oven heat shall be provided with electric resistance heating elements located outside of, or otherwise shielded from the active oven work space so that the oven load is not subject to direct radiant energy from the heating elements. The number and size of the elements shall be sufficient to raise and stabilize the oven temperature from ambient level to maximum operating temperature in fifty minutes. The elements shall be sufficiently ventilated to maintain them in the "black" heat or low-glow heat range. The heaters shall conform to the requirements of the National Fire Code for ovens and furnaces, NFPA No 86.A concerning electric heating systems.

3.4.16 Thermostat. The oven temperature adjustment shall be by means of a thermostat control. The thermostat shall be located on the outside of the oven and shall be fully enclosed to protect the instrument against dust, dirt, moisture or corrosive media. Adjustment shall be by means of a knob or lever and graduated reference dial permitting selective temperature control within the operating temperature range of the oven. The thermostat shall be capable of automatically maintaining the oven temperature as set, to within $\pm$10 degrees F. The thermostat sensing element shall be suitably located inside the oven to ensure sensitivity to overall oven temperature fluctuations.

3.4.17 Excess temperature cut-out. An excess temperature cut-out shall be provided and shall be a type which is adjustable over a minimum range of 150 degrees F below to 50 degrees F above the maximum operating temperature of the oven. The cut-out shall function to remove power from the oven when the set over-temperature point is reached. The power shall remain off until reset manually and shall comply with the requirements defined in NFPA Article 530.

3.4.18 Over-current protection. Over-current protection for each major element in the heater circuit(s) shall be in compliance with National Electrical Code, Article 240, and the National Fire Code, Article 570, Safety Control Application for Heating Systems Lacking Fuel Hazards.

3.4.19 Cut-out switch. A cut-out switch shall be mounted on the door/door frame and shall function in the control circuit so that current to the heating units shall be disconnected when the oven doors are opened and reconnected when the doors are securely closed.

3.4.20 Indicator light. The oven shall be equipped with a red indicator light which will function when the oven heaters are energized. The light shall signal automatic temperature control by the thermostat by indicating the make and break of the heater circuits. It shall be located near the door(s) and be visible when facing the doors.

3.5 Electrical requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-23T09:56Z
Check the source to verify that this is the current version before use.

3.5.1 Electrical system. The oven electrical system shall comply with the requirements of the National Electrical Code and NFPA No. 86.A, concerning oven electrical safety and malfunction shutdown.

3.5.2 Power supply. The oven shall operate on a 230 volt, 60 Hertz, 3 phase AC power supply, unless otherwise specified in 6.2.

3.5.3 Wiring. The oven wiring shall be completely enclosed in easily accessible control and wiring compartment(s) integral with the cabinet. Any wiring exposed to high temperatures shall be covered with suitable heat and moisture resisting insulation. Electrical connections shall not be adversely affected by the heat produced as a result of normal use of the oven.

3.5.4 Fan drive. An electric motor driven fan or blower shall be provided to circulate air within the type II oven. The motor shall conform to NEMA MG -1 and shall have adequate thermal and over-current protection. The fan drive may be direct or indirect provided that components such as the motor which are not heat-resisting are suitably insulated from elevated temperatures. The assembly shall be statically and dynamically balanced.

3.5.5 Electrical connections. Electrical connections of conductors and terminal parts shall be of the mechanical pressure or solder type. When soldered connections are used, the conductors and terminal parts shall be mechanically secured. Flux for soldering shall be of a resin base.

3.5.6 Dielectric strength. The electrical system, excluding control and recording instruments shall be capable of withstanding without damage for a period of one minute an impressed voltage of 1,000 volts plus twice the rated voltage without breaking down and shall register a resistance value of not less than one megohm at 500 volts DC. The dielectric strength shall apply between all insulated circuits and insulated circuits and external metal parts.

3.6 Nameplate. Unless otherwise specified (see 6.2), a corrosion-resistant metal nameplate, marked in accordance with MIL-STD-130 shall be securely attached to the oven. The nameplate shall contain the following information:

Nomenclature
Manufacturer's Name
Manufacturer's Model
Manufacturer's Serial Number
Power Input Characteristics and Ratings
Inside Dimensions (Width, Depth, Height - In Inches)
Maximum Operating Temperature, Degrees F
Date of Manufacture
Contract or Order Number
Federal Stock Number
U. S.

4. QUALITY ASSURANCE PROVISIONS

Source: https://assist.dla.mil -- Downloaded: 202640-23T09:56Z
Check the source to verify that this is the currentversion before use.

4.1 Responsibility for inspection. Unless otherwise specified in the contract or order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.2 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection (see 4.3).
b. Quality conformance inspection (see 4.4).

4.3 First article inspection. Unless otherwise specified (see 6.2), first article inspection shall consist of the examination in 4.5 and all tests in 4.6. Failure of the oven to pass the examination or any test shall require re-evaluation and correction/resolution, followed by re-examination or retest as necessary to clear the problem.

4.3.1 First article test report. The following information shall be included as part of first article test reports:

a. Test article descriptive statistics (actual measure or rating, as opposed to specification limits) such as key dimensions, weights, wattage, horsepower, pressure, temperature, loads, ranges, and so forth.

b. Test article photographs describing major external and internal detail.

4.4 Quality conformance inspection. Quality conformance inspection shall be performed on each item prior to being offered for acceptance under the contract. Unless otherwise specified (see 6.2), quality conformance inspection shall consist of the examination in 4.5, the test in 4.7, and the inspection in 4.8. Failure of the oven to pass the inspection shall be cause for rejection.

4.5 Examination. The oven and equipment shall be examined for design, construction, material, components, electrical equipment and workmanship, to determine compliance with the requirements of this specification. Particular attention shall be given to safety and to identification of potentially hazardous design and/or operating conditions.

4.6 Tests. All tests shall be conducted at an ambient temperature of 77 degrees F ±18 degrees F.

4.6.1 Shelf deflection test. A shelf shall be supported in the same manner as it is supported in the oven and shall have a load of 50 pounds placed at its center for a period of five minutes and then removed to determine conformance to 3.4.13. The base area of the load shall not exceed 30 square inches.

Source: https://assist.dla.mil -- Downloaded: 202640-23T09:56Z
Check the source to verify that this is the current version before use.

4.6.2 Bottom deflection and shelf support rigidity test. A shelf shall be supported in the same manner as it is supported in the oven and shall have a 50 pound load placed at its center for a period of five minutes and then removed to determine compliance with 3.4.13. The base area of the load shall not exceed 30 square inches. The shelf supports shall be tested by placing the shelf in its normal totally inserted position in the oven, with the 50 pound load placed in the center of the shelf. After two and one-half minutes, the shelf shall be withdrawn approximately one-half the depth of the oven and allowed to remain for an additional two and one-half minutes. After removal of the load and shelf, the supports shall show no sign of permanent deformation.

4.6.2.1 Bottom deflection test. For walk-in ovens, a person weighing at least 150 pounds shall walk randomly around and about the oven floor for five minutes. The oven bottom shall not buckle permanently, sag, or sustain any other permanent deformation or damage. The floor/bottom of ovens not large enough for or classified as walk-in, shall be tested by placing a 50 pound load, of no more than 30 square inches bearing area, at a minimum of five locations about the floor. The load shall remain in place at each location for at least three minutes. No permanent deformation or damage to the floor shall result.

4.6.3 Thermometer accuracy test. Each thermometer or temperature indicator shall be compared with a National Bureau of Standards (NBS) certified thermometer to determine conformance to the accuracy requirements in 3.4.14.

4.6.4 Cut-out switch test. The doors shall be closed and the heating elements energized. The doors shall than be alternately opened and closed several times to determine that the cut-out switch performs in accordance with 3.4.19.

4.6.5 Indicator light test. With the oven energized, the indicator light shall be observed to assure that it cycles on and off with the heating elements in conformance with 3.4.20.

4.6.6 Heating time test. With no load in the oven, the doors closed, and commencing at ambient temperature, the oven shall be placed in operation and tested for compliance with the heating time requirements set forth in 3.4.15.

4.6.7 Heat loss test. The oven shall be operated for a period of time sufficient to allow complete thermal stability at maximum operating temperature. Then, without altering any controls or settings, the oven heater power consumption shall be accurately measured for a period of at least one hour. The power consumption so measured and reflected in KW/HR shall not exceed 0.05 KW per square foot of oven actual interior surface area as measured in 1.2.1.4.

4.6.8 Temperature variation. (Type II Ovens Only) At least four identically reading thermocouples with suitable indicating instruments shall be located within the oven to determine overall temperature variation and to determine that the limits of 3.4.1.2 have not been exceeded. The effective work space shall be defined as being 1/6 less than the actual inside dimensions of the work chamber. The test shall be conducted between 55-60 degrees F above existing ambient conditions and maximum operating temperature.

Source: https://assist.dla.mil -- Downloaded: 202640-23T09:56Z
Check the source to verify that this is the current version before use.

4.6.9 Dielectric test. The circuits of the electrical system excluding control or recording instruments shall withstand without damage for a period of one minute, an impressed voltage of 1,000 volts plus twice the rated voltage. The impressed voltage shall be applied between all insulated circuits and insulated circuits and the metal frame.

4.6.10 Insulation resistance test. The insulation shall be subjected to an electrical resistance test after the dielectric test in 4.6.9 has been completed. The resistance between all insulated circuits and the metal frame shall be not less than one megohm at 500 DC.

4.6.11 Operational test. The oven shall be operated for a sufficient period of time to insure that all systems operate and function correctly. The thermostat shall successfully complete at least 50 cycles of operation and the oven shall then be tested to determine compliance with 3.4.16. All safety items and safety interlocks shall be checked for proper operation by simulating failure of key components such as fans, sticking thermostat, sticking main contactor and/or by purposely mis-operating the oven controls to whatever extent that might be inadvertently possible during normal use.

4.7 Performance. The oven temperature shall be set at a convenient point within the upper temperature range and operated for a sufficient period of time to demonstrate that the oven will perform continuously and satisfactorily and that all oven controls and components function properly.

4.8 Inspection of preparation for delivery. Preservation, packaging, packing and marking for shipment and storage shall be inspected to determine conformance to the requirements of section 5.

5. PREPARATION FOR DELIVERY

5.1 Preservation, packaging, packing and marking. Preservation, packaging, packing and marking shall be in accordance with the applicable requirements set forth in the solicitation schedule.

6. NOTES

6.1 Intended use. The oven(s) covered within the limits of this specification are intended for use in thermal drying and curing various products, components and end items in shops and laboratories, without the need for continuous manual attention.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number, and date of this specification.

b. Type of oven required (1.2.1.1).

c. Class of oven required (1.2.1.2).

d. Style of oven required (1.2.1.3 and 3.4.10).

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T09:56Z
Check the source to verify that this is the current version before use.

e. Size of minimum, clear, interior dimensions required (1.2.1.4).

f. When first article is required (3.1).

g. Ventilation, if different (3.4.6).

h. Shelf supports, if different (3.4.8).

i. Cabinet supports, if required (3.4.9).

j. Number of shelves required for other than walk-in ovens (3.4.13).

k. Number of full size shelves for walk-in ovens, if required and the type, size, number and location of partial shelves, if required (3.4.13).

l. Operating temperature range required (3.4.14.1).

m. Operating power supply, if different (3.5.2).

n. Nameplate, if different (3.6).

o. First article inspection, if required (4.3).

p. Quality conformance inspection, if different (4.4).

6.3 Supersession data. This specification includes the Air Force requirements of Federal Specification W-O-830.

Custodian:
Air Force - 84

Preparing activity:
Air Force - 84

Project number:
4430-F012

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T09:56Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION

MIL-O-83990 (USAF)

ORGANIZATION	CITY AND STATE
Warner Robins ALC/MMEME Robins AFB Ga.	Warner Robins Ga 31098.

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (*Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity*)

SUBMITTED BY (*Printed or typed name and activity*)	DATE

DD FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE. AF-WP-O-MAY 64 505

Source: https://assist.dla.mil -- Downloaded: 202610-23T09:56Z -
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEPARTMENT OF THE AIR FORCE

Warner Robins ALC/MMEME
Robins AFB Ga 31098

UNITED STATES AIR FORCE
OFFICIAL BUSINESS

Warner Robins ALC/MMEME
Robins AFB Ga 31098

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T09:56Z
Check the source to verify that this is the current version before use.