MIL-B-8598A
5 May 1970
SUPERSEDING
MIL-B-8598(ASG)
17 August 1953

MILITARY SPECIFICATION

BULB, ELECTRICAL RESISTANCE, THERMOMETER, FREE AIR

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers flush-mounted electrical resistance bulbs, with a range from -70° to +70° Centigrade (C).

1.2 Classification. The bulbs shall be of the following types, as specified herein:

Accuracy 1.0°C - Part No. MS28038-1
Accuracy 0.5°C - Part No. MS28038-2

1.2.1 The Air Force designates the Part No. MS28038-1 bulb as type B-1 and the Part No. MS28038-2 bulb as type B-3. These designations are not applicable to Navy activities.

2. APPLICABLE DOCUMENTS

* 2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-636	Box, Fiberboard

Military

MIL-P-116	Preservation, Methods Of
MIL-D-1000	Drawings, Engineering And Associated Lists
MIL-R-6855	Synthetic Rubber Sheets, Strips, Molded Or Extruded Shapes
MIL-S-7742	Screw Threads, Standard, Optimum Selected Series: General Specification For

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures And Tables For Inspection By Attributes
MIL-STD-129	Marking For Shipment And Storage
MIL-STD-130	Identification Marking Of US Military Property
MIL-STD-143	Specifications And Standards, Order Of Precedence For The Selection Of
MIL-STD-810	Environmental Test Methods
MIL-STD-1186	Cushioning, Anchoring, Bracing, Blocking, And Waterproofing; With Appropriate Test Methods
MS24693	Screw, Machine, Flat Countersunk Head, 100°, Cross Recessed, JNC-2A And UNF-2A
MS28038	Bulb-Electrical Resistance, Thermometer, Free Air

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

* 2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

UNIFORM CLASSIFICATION COMMITTEE

Uniform Freight Classification Rules

(Application for copies should be addressed to the Uniform Classification Committee, Room 202, Union Station, 516 West Jackson Boulevard, Chicago IL 60606.)

3. REQUIREMENTS

* 3.1 Qualification. The bulbs furnished under this specification shall be products which are qualified for listing on the applicable qualified products list at the time set for opening of bids (see 4.4 and 6.3).

* 3.2 Materials.

* 3.2.1 Metals. Metals shall be of the corrosion-resistant type, unless suitably protected to resist corrosion during normal service life.

* 3.2.2 Synthetic materials. Synthetic materials shall be of the thermal-setting type composed of a suitable filler and phenol condensation product binder.

* 3.2.3 Fungus-proof materials. Materials which are not nutrients for fungi shall be used to the greatest extent practicable. In cases where materials that are nutrients for fungi must be used, such materials shall be treated with a fungicidal agent as approved by the procuring activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 3.2.4 Protective treatment. When materials are used in the construction of the bulb that are subject to deterioration when exposed to climatic and environmental conditions likely to occur during service usage, they shall be protected against such deterioration in a manner that will in no way prevent compliance with the performance requirements of this specification. The use of any protective coating that will crack, chip, or scale with age or extremes of climatic and environmental conditions shall be avoided.

* 3.2.5 Selection of materials. Specifications and standards for all materials, parts, and Government certification and approval of processes and equipment, which are not specifically designated herein and which are necessary for the execution of this specification, shall be selected in accordance with MIL-STD-143, except as provided in the following paragraph.

* 3.2.5.1 Standard parts. Standard parts (MS or AN) shall be used wherever they are suitable for the purpose, and shall be identified on the drawing by their part numbers. Commercial utility parts such as screws, bolts, nuts, cotter pins, et cetera, may be used, provided they possess suitable properties and are replaceable by the stnadard parts (MS or AN) without alteration, and provided the corresponding standard part numbers are referenced in the parts list and, if practicable, on the contractor's drawings. In the event there is no suitable corresponding standard part in effect on date of invitation for bids, commercial parts may be used provided they conform to all the requirements of this specification.

* 3.3 Design and construction. The bulb shall conform to MS28038.

* 3.3.1 The bulb shall be so constructed that no parts will work loose in service. It shall be built to withstand the strains, jars, vibrations, and other conditions incident to shipping, storage, installation, and service, without failure.

* 3.3.2 Temperature-resistance characteristics. The temperature-resistance characteristics of the bulb shall conform to Table I.

TABLE I. Temperature Resistance Characteristics Aircraft Thermometer Bulb

°C	Ohms Resistance	Tolerance-Ohms	
		MS28038-1	MS28038-2
-70	68.27	±0.50	±0.20
-60	71.19	±0.50	±0.20
-50	74.24	±0.40	±0.20
-40	77.39	±0.40	±0.20
-30	80.56	±0.40	±0.20
-20	83.77	±0.40	±0.20
-10	87.04	±0.40	±0.20
0	90.38	±0.40	±0.20
+10	93.80	±0.40	±0.20
+20	97.31	±0.40	±0.20
+30	100.91	±0.40	±0.20
+40	104.60	±0.40	±0.20
+50	108.39	±0.40	±0.20
+60	112.28	±0.50	±0.20
+70	116.27	±0.50	±0.20

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 3.3.3 Resistance element. The resistance element shall be sealed to withstand immersion of the bulb in liquid baths as specified in the applicable tests of Section 4.

* 3.4 Interchangeability. All parts having the same manufacturer's part number shall be directly and completely interchangeable with each other with respect to installation and performance. Changes in manufacturer's part numbers shall be governed by the drawing number requirements of MIL-D-1000.

* 3.5 Finish. All metals shall be suitably protected to meet the requirements of this specification.

* 3.6 Weight. The weight of the complete bulb shall not exceed 0.25 pound.

* 3.7 Screw threads. The threads of all machine screws greater than 0.060 inch in diameter shall conform to MIL-S-7742

* 3.8 Gaskets. Three gaskets constructed from black molded gasket rubber in accordance with MIL-R-6855, class III, grade 60, of each type as shown on MS28038 shall be furnished with each bulb.

* 3.9 Mounting lugs and screws. Each mounting lug shall be fitted with a self-locking nut, with NO. 6-32 NF-2B threads, securely attached to prevent rotation when the screw is installed. The contractor shall furnish sufficient MS24693 screws for installing the bulb.

* 3.10 Performance. The bulb shall be capable of meeting the requirements specified in Section 4, when subjected to the applicable tests.

* 3.11 Identification of product. The bulb shall be marked for identification in accordance with MIL-STD-130. The following information shall be legibly and permanently marked on the flat of the boss upon which the electrical receptacle is mounted, or on a nameplate in such a manner as to be clearly seen when the bulb is installed:

MS Part No.
1/ _____
Manufacturer's Part No.
Manufacturer's Serial No.
Contract or Order No.
Manufacturer's name or trade-mark

1/ For Air Force procurement only, insert type B-1, or type B-3, as applicable.

* 3.11.1 Use of AN or MIL designations. AN or MIL designations shall not be applied to a product, except for qualification test samples, nor referred to in correspondence, until notice of approval has been received from the activity responsible for qualification.

* 3.12 Envelope. An envelope, containing the mounting screws and gaskets, shall be furnished with each bulb. The following information shall be printed on the face of the envelope:

IMPORTANT
THIS ENVELOPE CONTAINS MOUNTING SCREWS AND GASKETS

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 3.13 Workmanship. The bulb, including all parts and accessories, shall be constructed and finished in a thoroughly workmanlike manner. Particular attention shall be given to neatness, marking of parts and assemblies, and freedom of parts from burrs and sharp edges.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.2 Classification of inspections. The examination and testing of the bulb shall be classified as follows:

a. Qualification inspection: Those tests accomplished on samples submitted for qualification as a satisfactory product. (4.4)

b. Quality conformance inspection: Those tests accomplished on bulbs manufactured and submitted for acceptance under contract. (4.5)

* 4.3 Test conditions.

* 4.3.1 Standard atmospheric conditions. Unless otherwise specified, all tests shall be made at an atmospheric pressure of approximately 29.92 inches mercury and at a temperature of approximately +25°C. When tests are made with atmospheric pressure and temperature substantially different from the above values, proper allowance shall be made for the difference from the specified conditions.

* 4.3.2 Resistance values. Whenever practicable, the resistance values given for test results shall be the averages of direct and reversed readings to eliminate thermal-electric potential errors.

* 4.3.3 For test purposes, the sensitive portion of the bulb shall be considered as from the front of the bulb to the front of the mounting flange as shown on MS28038. The head of the bulb shall be considered as from the front of the mounting flange to the top of the electrical receptacle as shown on MS28038.

* 4.3.4 Unless otherwise specified, all temperature-resistance measurements shall be made in an agitated liquid bath with the sensitive portion of the bulb immersed.

* 4.3.5 Test voltage. Unless otherwise specified, all resistance measurements shall be made with 1.5 ±0.2 volts direct current (DC) applied to the bulb.

* 4.3.6 Master. The master shall be defined as a temperature-measuring device with known calibration.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 4.4 Qualification inspection. The qualification inspection shall consist of all tests specified herein.

* 4.4.1 Prior qualification. Unless otherwise specified by the procuring activity, bulbs which have not previously passed qualification inspection or which have passed the qualification inspection and have been modified in any manner, shall satisfactorily pass qualification inspection prior to the acceptance of any bulb.

* 4.4.2 Sampling instructions. Qualification inspection samples shall consist of three bulbs. Samples shall be identified as required and forwarded to the activity responsible for qualification, designated in the letter of authorization from the qualifying activity. (See 6.3)

* 4.5 Quality conformance inspection. The inspection shall consist of individual tests (4.5.1), sampling tests (4.5.2), and special tests (4.5.3).

* 4.5.1 Individual tests. Each bulb submitted for acceptance under contract shall be subjected to the following examination and test.

* 4.5.1.1 Examination of product. The bulb shall be inspected to determine compliance with the requirements specified herein with respect to material, workmanship, and marking.

* 4.5.1.2 Bulb adjustment test. The resistance of the bulb shall be measured at room temperature. The measured resistance shall be within 0.40 ohm for the Part No.MS28038-1, and within 0.20 ohm for the Part No. MS28038-2 of the resistance specified in Table I for the exact temperature of the master. To determine the true resistance equivalent to the temperature of the master, interpolation of the values given in Table I may be necessary.

* 4.5.2 Sampling plan A tests. Sampling shall be performed in accordance with MIL-STD-105, Inspection Level II AQL 2.5 percent.

* 4.5.2.1 Bulb resistance. The resistance of each bulb shall be measured at temperatures of -50° ±5°C, 0° ±5°C, and +50° ±5°C in an agitated liquid bath. If desired by the inspector, resistance measurements at any other points in Table I may also be made. The measured resistance shall be within the tolerances specified in Table I for the exact temperature of the master. To determine the true resistance equivalent to the temperature of the master, interpolation of the values given in Table I may be necessary.

* 4.5.2.2 Vibration. The bulb shall be subjected to vibrations as specified in Procedure 1, Parts 1, 2 and 3, Method 514, Vibration of MIL-STD-810. The period of vibration shall be as specified by Time Schedule 1 of Table 514-11 and the double amplitude and frequency range shall be in accordance with Curve Z of Figure 514-1. An indicator or a wheatstone bridge shall be connected to the bulb during this test. Induced pointer oscillation or variation greater than 2.0 degrees (0.80 ohm) for Part No. MS28038-1, or 1.0 degree (0.40 ohm) for Part No. MS28038-2 shall be cause for failure. Following vibration, the bulb shall be subjected to the bulb adjustment test (4.5.1.2). The measured resistance shall be within 0.40 ohm for Part No. MS28038-1 and within 0.20 ohm for Part No. MS28038-2 of the resistance specified in Table I for the exact temperature of the master. In addition, no looseness nor damage to any part of the bulb shall result from this test. (In accordance with recommended test levels according to MIL-STD-810 for aircraft equipment location as stated in 6.1.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 4.5.3 Sampling plan B tests. Sampling shall be in accordance with MIL-STD-105, Special Inspection Level S-1, AQL 6.5 percent.

* 4.5.3.1 Detail examination. The bulb shall be critically examined to determine full compliance with regard to weight, physical dimensions, markings, and similar detail requirements.

* 4.5.3.2 Response time. The bulb, with sensitive portion immersed, shall be quenched from an agitated liquid bath at 70°C to an ice bath at 0°C. The bulb shall be agitated in the ice bath. The time required for the bulb temperature to fall from 70° to 10°C shall be noted. The time required for the bulb temperature to change from 70° to 10°C shall not exceed 3 seconds.

* 4.5.3.3 Thermal conductivity. With heated air, the head of the bulb shall be maintained at a temperature of 70° ±2°C. The sensitive portion of the bulb shall be subjected to a temperature of -50° ±2°C in an agitated liquid bath. The measured resistance of the bulb shall not differ by more than 1 ohm from the resistance determined in the bulb resistance test (4.5.1.2) for the exact temperature of the master.

* 4.5.3.4 Self-heating. The bulb shall be placed in an agitated liquid bath at a temperature of 25° ±2°C. A current of 150 milliampere (mA) shall be applied to the bulb for a minimum period of 2 minutes and the maximum rise in resistance measured during this period. The rise shall not exceed 3 ohms.

* 4.5.3.5 Thermoelectric potential. The potential produced in the bulb winding measured with a potentiometer when the sensitive portion of the bulb is subjected to a temperature of 70° ±2°C in an agitated liquid bath shall not exceed 1 millivolt (mV).

* 4.5.3.6 Cold resistance. The bulb shall be subjected to a temperature of -65°C for a period of 15 hours. With the bulb at -65°C, it shall be subjected to vibration, as specified herein, for a period of 10 minutes. The vibration frequency shall be 3,000 cycles per minute for this test. There shall be no damage to the bulb as a result of this test.

* 4.5.3.7 High temperature. The bulb shall be subjected to a temperature of 71°C for a period of 15 hours. With the bulb at 71°C, it shall be subjected to vibration, as specified herein, for a period of 10 minutes. The vibration frequency shall be 3,000 cycles per minutes for this test. There shall be no change to the bulb as a result of this test.

* 4.5.3.8 Humidity. The bulb shall be subjected to humidity tests, Procedure I, of MIL-STD-810. There shall be no rusting or corrosion of the bulb or any other damage as a result of this test.

* 4.5.3.8.1 Insulation leakage after humidity. A potential of 100 volts DC shall be applied between any electrical receptacle pin contact and any metal part of the bulb. Air drying of the bulb receptacle connecting leads and outer housing prior to conducting this is permissible. The insulation resistance measured between any pin and any metal part of the bulb shall be equal to or greater than 20 megahertz (MHz).

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 4.5.3.9 Fungus. A fungus test in accordance with fungus tests, Procedure I, of MIL-STD-810 shall be conducted to determine the bulb's resistance to fungus. Upon completion of the test period, there shall be no signs of fungus growth on the bulb

* 4.5.3.10 Bulb adjustment. Following the completion of the foregoing tests, the bulb adjustment test (see 4.5.1.2) shall be performed. The measured resistance shall be within 0.40 ohm for Part No. MS28038-1, and within 0.20 ohm for Part No. MS28038-2 of the resistance specified in Table I for the exact temperature of the master.

* 4.5.3.11 Tensile strength. The mounting flange of the bulb with the sensitive element removed shall be mounted on the movable head of an Amsler testing machine, or equivalent apparatus, with the diameter of the bulb in a horizontal plane so that the mounting lugs receive no added support. A suitable pin shall be inserted through the hole in the mounting nut and attached to a pull strap in the stationary head of the machine. Each mounting lug of the completely assembled bulb shall withstand for 1 minute, without fracture, a load of 175 pounds applied along the insert nut axis and toward the sensitive portion of the bulb.

* 4.5.3.12 Mounting nuts. Each mounting nut shall have a load of 25 pounds applied to the nut in a manner tending to pull the nut out of the lug toward the rear of the bulb. The mounting lug shall not become loose in the lug and shall not rotate more than a fraction of a turn under a torque of 10 pound-inches.

* 4.5.3.13 Shock. The bulb, nonoperating, shall be subjected to Procedure 1, Method 516, Shock Test of MIL-STD-810. The wave form of the shocks and time duration shall be in accordance with Figure 516-1, Procedure 1 for Flight Vehicle Equipment. Following this test the bulb shall be subjected to the bulb adjustment test (4.5.1.2). The tolerance requirements of 4.5.1.2 shall apply.

* 4.5.3.14 Salt fog. The bulb shall be subjected to a Salt Fog test in accordance with Method 509 of MIL-STD-810. External connections shall be made to the bulb in such a manner as to simulate installed conditions. Following the salt fog exposure, the bulb shall be washed with distilled water. Following the 48-hour storage the bulb shall be subjected to and meet the requirements of the bulb adjustment test (4.5.1.2) and the insulation leakage after humidity test (4.5.3.8.1). In addition evidence of corrosion shall be cause for failure.

* 4.5.4 Rejection and retest. When sampling tests (4.5.2) are specified on a number of items that are selected as representative of a certain lot, and one or more of this number fail to meet the specified tests(s), acceptance of all items shall be withheld until the extent and cause of failure is determined. For operational reasons, individual tests (4.5.1) may be continued pending investigation of a sampling test failure, but the final acceptance of the product is contingent upon the inspector's decision regarding the over-all conformance of the product to the specification requirements. If investigation indicates that the defects may exist on items previously accepted, full particulars concerning the defect(s) found, including recommendations for correction, will be furnished to the procuring activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 4.6 All parts, specimens, or assemblies destroyed in making tests required by this specification or drawings, to determine compliance with the specification or drawings, shall be in addition to the quantity specified in the contract or purchase order and shall be furnished without increasing the cost of the contract or order.

* 4.7 Preservation, packaging, packing, and marking. Preparation for delivery shall be inspected for conformance to Section 5.

5. PREPARATION FOR DELIVERY

* 5.1 Preservation and packaging. Preservation and packaging shall be level A or C as specified (see 6.2).

* 5.1.1 Level A. Each bulb shall be preserved and packaged in accordance with MIL-P-116, Method IC-1(no preservative) and placed in a snug-fitting container conforming to grade W6c, PPP-B-636.

* 5.1.2 Level C. Each bulb shall be preserved and packaged in a manner which will afford adequate protection against corrosion, deterioration and physical damage during shipment from supply source to the first receiving activity. The supplier may use his commercial practice providing it meets this requirement.

* 5.2 Packing. Packing shall be level A, B, or C as specified (see 6.2).

* 5.2.1 Level A. Bulbs preserved and packaged as specified in 5.1.1 shall be packed in weather-resistant shipping containers conforming to PPP-B-636. The grade of each shipping container, determined by the gross weight and size, shall comply to the applicable special requirements table of the specification. Interior cushioning, blocking, bracing and waterproofing shall be in accordance with MIL-STD-1186. Closure and strapping shall be in accordance with the container specification. Containers shall be of uniform shape and size and of minimum cube and tare.

* 5.2.2 Level B. Level B shall be same as level A, except that waterproofing is not required and domestic containers may be used.

* 5.2.3 Level C. Packages that require overpacking for acceptance by the carrier shall be placed in exterior type shipping containers in a manner that will insure safe transportation at lowest rate to the point of delivery. Containers shall comply with Uniform Freight Classification Rules, or regulations of other carriers as applicable to the mode of transportaiton.

* 5.3 Marking. Interior and exterior containers shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The temperature bulb covered by this specification is intended for mounting flush on aircraft skins for precision free air temperature measurement.

6.2 Ordering data . Procurement documents should specify the following:

a. Title, number, and date of this specification.

b. Levels of packaging and packing required (see Section 5).

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

* 6.3 With respect to products requiring qualification, awards will be made only for products which are at the time set for opening of bids, qualified for inclusion in the applicable Qualified Products List whether or not such products have actually been so listed by that date. The attention of the suppliers is called to this requirement, and manufacturers are urged to arrange to have the products that they propose to offer to the Federal Government tested for qualification in order that they may be eligible to be awarded contracts or orders for the products covered by this specification. The activity responsible for the Qualified Products List is the San Antonio Air Materiel Area, Service Engineering Division, ATTN: SANE, Kelly AFB, Texas 78241, and information pertaining to qualification of products may be obtained from that activity.

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Air Force - 82
Navy - AS

Review Activities:
Air Force - 82
Navy - AS

User Activity:
Air Force - 11

Preparing Activity:
Air Force - 82

Project No. 6685-0322

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD, FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64 WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T10:39Z
Check the source to verify that this is the current version before use.