FED. SUP CLASS
6640



MS35980-1 thru MS35980-4



MS35980-5 thru MS35980-7

PART NUMBER	CAPACITY mL	GRADUATED INTERVAL mL	NUMBERED INTERVALS mL
MS35980-1	0.1 ± .005	0.01	0.01
MS35980-2	0.2 ± .008	0.01	0.02
MS35980-3	1.0 ± .02	0.01	0.10
MS35980-4	2.0 ± .02	0.10	0.20
MS35980-5	5.0 ± .04	0.10	1.0
MS35980-6	10.0 ± .06	0.10	1.0
MS35980-7	25.0 ± .10	0.10	1.0

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A or B.

DESIGN: Pipets shall be of one piece construction and designed to permit complete emptying and thorough cleaning.

Tip of the pipet shall be tempered to reduce chipping and breakage.

MS35980-5, MS35980-6, and MS35980-7 shall have mouthpieces that are reduced in diameter.

All pipets shall be calibrated "to deliver" at 20°C with etched or fused on graduation marks.

NOTES:

1. Color coding bands on pipets are optional.

Ⓑ 2. In the event of a conflict between the text of this standard and the references cited herein, the text of this standard shall take precedence.

3. Referenced Government documents of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DODISS) specified in the solicitation form a part of this standard to the extent specified herein.

APPROVED 23 DEC 1959 REVISED Ⓐ 14 JUL 1969 Ⓑ 16 MAR 1987

P.A.	EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust	MB MS 99		PIPET, MEASURING general purpose	MS35980
Procurement Specifications: NONE		SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672-1 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1407-3

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

REVIEWER ACTIVITIES: ARMY - ME, NAVY - OS, YD AIR FORCE - 82

USER ACTIVITIES:

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.