NOTICE.—When ... ernment drawings, specifications, ... other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

FED. SUP CLASS
6640

Cancelled with a superseding standard
Use MS36196

PART NUMBER	RATED DISTILLATION gal/hour	MAX OVER-ALL HEIGHT in.
MS36192-1	2	30
MS36192-2	5	40

MATERIAL: The apparatus shall be made of materials that will meet the design, finish, and performance requirements of this standard.

DESIGN: Each unit shall be of such design as to perform the intended function and be capable of operation by heating with steam. Apparatus is equipped with a constant bleeder device.

FINISH: Metal surfaces which come into contact with vapors or distillate shall have a coating of pure block tin, 0.0001 inch minimum thickness or stainless steel. All other surfaces shall resist corrosion.

PERFORMANCE: The apparatus shall produce distilled water for "Water for Injection," conforming to the requirements of the United States Pharmacopeia. Water distillation rate shall be not less than the rated gallons per hour.

NOTES:

1. Steam pressure, 40 to 60 ... satisfactory operati...
2. Items that ... to rec...

CANCELLED
CANCELLED AFTER 7 FEBRUARY 1980 WITH A SUPERSEDING STANDARD MS36196

... date

APPROVED 23 JUN 1961 REVISED 7 February 1980

P.A. CmlC Other Cust BuMed DAY	TITLE DISTILLING APPARATUS, LABORATORY water, steam heated	MILITARY STANDARD MS36192	
Procurement Specification GG-D-465	SUPERSEDES:	SHEET 1 OF 1	

DD FORM 1 SEP 57 672 (Coordinated) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

Project Number 6640-0780

☆U.S. GOVERNMENT PRINTING OFFICE: 1980—603-121/1056

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.