FED. SUP CLASS
6630

| PART NUMBER | ENVELOPES pH VALUE |
|---|---|
| MS36452-1 | 4.01 ± .01 |
| MS36452-2 | 6.86 ± .01* |
| MS36452-3 | 7.00 ± .03 |
| MS36452-4 | 7.413 ± .01* |
| MS36452-5 | 9.18 ± .01 |
| MS36452-6 | 10.01 ± .01* |
| MS36452-7 | 12.45 ± .05 |
| MS36452-8 | ASSORTMENT PACK |

*pH Value when mixed by accurate analytical methods.

MATERIAL: Buffers shall consist of powdered ACS grade salts.

Bottle shall be of polyethylene.

DESIGN: Buffer salts shall be packaged in plastic envelopes. Each envelope of buffer salts shall be sufficient to make 500 ml of solution.

PURPOSE: To provide precise means for calibrating meters to a specific pH reference.

COMPONENTS: Parts (Kits) -1 through -7 shall consist of one 500 ml polyethylene reagent bottle and 12 envelopes of (one pH value) buffer salts, and instructions.

Part -8 shall consist of 14 envelopes (2 each of envelopes -1 through -7)

NOTES:
1. The pH values listed are for solution in distilled water at 25°C.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.

USER ACTIVITIES: ARMY - CE
NAVY - NC

REVIEWER ACTIVITIES: ARMY - MD

APPROVED 7 OCTOBER 68 REVISED (A) 28 JAN 1982

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(A) ENTIRE STANDARD REVISED

| P.A. EA | INTERNATIONAL INTEREST | TITLE | MILITARY STANDARD |
|---|---|---|---|
| Other Cust MS 99 | | STANDARD SOLUTION KIT, HYDROGEN ION METER | MS36452 |
| Procurement Specifications NONE | SUPERSEDES: | | SHEET 1 OF 1 |

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6630-0191

REVISION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:41Z
Check the source to verify that this is the current version before use.