GG-R-1549A
12 June 1986
(Superseding)
GG-R-001549 (DSA-DM)
21 November 1968

FEDERAL SPECIFICATION

RONGEUR, Dental

This specification is approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal Agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers rongeurs suitable for use in surgical procedures.

1.2 Classification. Rongeurs shall be of the following types, as specified (see 6.1).

Type I. Mead, No. 1A.
Type II. No. 4A.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

Federal Specifications

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Fiberboard.

Federal Standards

FED-STD-123 Marking for Domestic Shipment (Civilian Agencies).

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal specifications, standards, and commercial item descriptions as outlined under General Information in the Index of Federal Specifications, Standards and Commercial Item Descriptions. The Index, which includes cumulative bimonthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402.)

(Copies of listed federal and military standards, specifications, Commercial Item Descriptions (CIDs), handbooks and associated documents listed in the Department of Defense Index of Specifications and Standards (DoDISS), should be obtained from the DoD Single Stock Point, Commanding Officer, Naval Publications and Forms Center, 5801 Tabor Avenue, Philadelphia, PA 19120. Copies of industry association documents should be obtained from the sponsor. Copies of all other listed documents should be obtained from the contracting activity or as directed by the contracting activity.)

(Federal Government activities may obtain copies of Federal standardization documents and the Index of Federal Specifications, Standards and Commercial Item Descriptions from established distribution points in their agencies.)

Military Specifications

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Military Standards

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MIL-STD-147	Palletized Unit Loads.
MIL-STD-794	Parts and Equipment, Procedures for Packaging of.

DRAWINGS

Defense Personnel Support Center (DPSC)

20508	Rongeur, Mead, No. 3A
20509	Rongeur, No. 4A

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity).

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.

ASTM E 92 - Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The rongeur shall be in accordance with the applicable requirements of GG-I-526, with the options and additional requirements specified herein.

3.2 Materials. The rongeur shall be class 4, type 420 or 410X stainless steel.

3.3 Hardness. After proper stress relieving and passivation, the Rockwell hardness of the jaws and handles shall be C40 to C55 (392-595 Vickers). The Rockwell hardness of the springs shall be C42-C52 (412-544 Vickers). Hardness shall be determined as specified in 4.4.1.

3.4 Design and dimensions. The rongeur shall be suitable for performing an alveolectomy, breaking down the alveolar process and other surgical procedures, and shall conform to Defense Personnel Support Center (DPSC) Drawings 20508 and 20509 for types I and II, respectively.

3.4.1 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design, function, fit, assembly and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

3.5 Construction.

3.5.1 Cutting edges. Cutting edges shall be properly sharpened to cut cleanly throughout their entire length, when tested as specified in 4.4.2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

3.5.2 Locks. The locks of the rongeur shall be either lap joint (scissors) type or box lock type. The edges of the joints shall be properly beveled to prevent pinching of tissues. Pin in lock area shall be flush with the surface of the joint in both open and closed positions. The rongeur shall be free of excessive stiffness and shall not have excessive play when the rongeur is in the open or closed position.

3.6 Finish. The rongeur shall have a mirror finish, except that the handles shall have a satin finish, and the concave surfaces of the beaks shall have a finely honed surface.

3.7 Corrosion resistance. Corrosion tests (copper sulfate and boil test) as specified in GG-I-526 shall apply to the entire rongeur.

3.8 Identification marking. Each rongeur shall be legibly and permanently marked with the applicable instrument number. Rongeurs of foreign manufacturer shall be marked with the name of the country of origin, together with the name or registered trademark of the contractor and manufacturer.

3.9 Workmanship. Rongeur shall be free from defects which detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in table I and table II (see 4.2.1 and 4.2.2) and table III (see 4.3).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I. Unit of product for examination purposes shall be one rongeur.

Table I. Examination Criteria

Type defect	Inspection level	AQL(percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one rongeur. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Table II. Tests

Characteristic	Requirement	Test Procedure	Inspection Level
Hardness	3.3	4.4.1	S-2
Cutting performance	3.5.1	4.4.2	S-2
Copper Sulfate Test	3.7	4.4	S-2
Boil Test	3.7	4.4	S-2
Composition	3.2	4.4	*

*Sufficient amount of each lot of each metal used in the construction of the rongeurs shall be selected for chemical analysis.

4.3 Examination. The rongeurs shall be examined for defects including, but not limited to, those listed in table III.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the currentversion before use.

Table III. Classification of Defects

Category	Defect
Major A:	
101	Style and design not as specified for style of instrument.
102	Instrument not constructed as specified.
103	Cutting edges of beaks not properly aligned.
104	Beaks not free from burrs or cracks.
105	Edges of the joint of the rongeur not properly beveled.
106	Joint of the rongeur not free from excessive stiffness.
107	Joint of the rongeur has excessive play.
Major B:	
151	Beak dimensions not as specified.
Minor A:	
201	Interfaces of lock joint have gross crevices.
202	Instrument not permanently and legibly marked as specified (see 3.8).
203	Instrument finish not as specified.
204	Burrs, scratches, or sharp edges (other than 104 and cutting edges) present.
Minor B:	
251	Dimensions (other than 151) not as specified.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements. Test procedures shall be as specified in GG-I-526 and as specified herein.

4.4.1 Hardness test. Hardness shall be determined in accordance with ASTM Standard E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid Hardness).

4.4.2 Cutting performance test. The rongeur shall cut 0.015 inch thick cardboard as specified in 3.5.1.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with group A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified, (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.1.1 Level A.

5.1.1.1 Cleaning. Rongeurs shall be cleaned in accordance with MIL-P-116.

5.1.1.2 Drying. Rongeurs shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservation applications. Preservatives shall not be used.

5.1.1.4 Unit package. Each rongeur shall be clean, dry and packaged in accordance with MIL-P-116, method IC-2.

5.1.1.5 Intermediate package. Six unit packages shall be packaged in a snug fitting box conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.1.2 Level C. The level C packaging for the rongeurs shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified, (see 6.1).

5.2.1 Level A. Sixteen intermediate packages (96 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Sixteen intermediate packages (96 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for rongeurs shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length, and shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Rongeurs packed as specified in 5.2.1 and 5.2.2, shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Rongeurs, packed as specified in 5.2.3, shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Civil agencies. Marking shall be as specified in FED-STD-123.

5.3.2 Military activities.

5.3.2.1 Levels A, B, and C. Each unit package, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

a. Title, number, and date of this specification.

b. Type required (see 1.2).

c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

6.2 National Stock Number (NSN) coverage. This specification covers the following items listed in the Federal Supply Catalog:

NSN	Item Identification
6520-00-531-3150	RONGEUR, Mead, No. 1A
6520-00-531-3250	RONGEUR, No. 4A

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Review Activities:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1855

Civil Agency Coordinating Activity: VA-OSS

Location: NCR

05422

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/43662

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.