GG-S-001634(DSA-DM)
January 30, 1970

INTERIM FEDERAL SPECIFICATION

SCALERS, DENTAL, Single-Ended

This Interim Federal Specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this Interim Federal Specification.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers single-ended dental scalers suitable for use in periodontal procedures.

1.2 Classification. The scalers shall be of the following types and styles (see 6.1):

Type I. Jaquette
- Style A. No. 1
- Style B. No. 2
- Style C. No. 3

Type II. Younger-Good, No. 15
Type III. Gerfing
Type IV. No. B

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards, and drawings. The following documents, of the issue in effect on date of invitations for bid or requests for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-123 Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151 Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D. C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston; New York; Atlanta; Chicago; Kansas City, Mo.; Fort Worth; Denver; San Francisco; Auburn Washington; and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

SPECIFICATIONS

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.
MIL-P-116 Methods of Preservation.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the currentversion before use.

DRAWINGS

Defense Personnel Support Center (DPSC)

20239	Scalers, Dental, Jaquette.
20377	Scaler, Dental, Younger-Good, No. 15.
20542	Scaler, Dental, Zerfing.
23021	Scaler, Dental, No. B.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitations for bid or requests for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM	Designation	E18	Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
ASTM	Designation	E92	Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103)

3. REQUIREMENTS

3.1 General requirements. The scalers shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Material. Scalers shall be fabricated from class 4, type 440A martensitic stainless steel.

3.3 Hardness. The operative ends and shanks of the scalers shall have a Rockwell hardness of C52 to C56 after proper stress relieving, and the handles shall have a minimum Rockwell hardness of C20, when tested as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

3.4 Style, design, and dimensions. The style, design, and dimensions of the scalers shall be in accordance with the Defense Personnel Support Center (DPSC) Drawings listed in Table I for the applicable scaler type, style, design, and number.

Table I. Applicable DPSC Scaler Drawings

Type	Style	Design and Number	DPSC Drawing
I	A	Jaquette No. 1	20239
I	B	Jaquette No. 2	20239
I	C	Jaquette No. 3	20239
II	--	Younger-Good, No. 15	20377
III	--	Gerfing	20542
IV	--	No. B	23021

The scaler handles shall be straight and flat.
The scalers shall be suitable for use in removing supragingival [illegible] deposits from the teeth.

3.5 Finish.

3.5.1 Operative end and shank. Operative ends (blades), including sharpened edges, shall have a fine honed finish such as is produced by using grade FF emery on a leather wheel. The shanks shall have a mirror or highly polished finish.

3.5.2 Handle. The handles of the scalers shall have a mirror or highly polished finish. Handles shall have 40 ± 10 fillet or notch-type serrations per inch on seven faces, and one face shall be smooth for identification marking. The depth of the serrations shall be 0.002 to 0.004 inch.

3.6 Corrosion resistance. The scalers shall be properly passivated and shall meet the copper sulfate and boil test requirements of GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

3.7 Identification marking. The scalers shall be permanently and legibly marked in accordance with GG-I-526. Types I, II, and III scalers shall also be marked with the name of the applicable design, or the abbreviation thereof, namely JAC, YNG, AND ZER, respectively. In addition, types I, II, and IV scalers shall be marked with the applicable instrument number five to twelve millimeters from the proximal (nonoperative) end of the handle.

3.8 Workmanship. Workmanship shall be first class throughout. The scalers shall be free from defects which detract from their appearance or may impair their serviceability, proper functioning, or intended use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw material, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the material, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in tables II and IV (see 4.2.1 and 4.3) and table III (see 4.2.2).

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the currentversion before use.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table II. The unit of product for examination purposes shall be one scaler.

Table II. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table III, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one scaler. Where feasible, the same unit of product shall be used for the determination of two or more test characteristics.

Table III. Sampling Criteria for Tests

Characteristic	Requirement	Test procedure	Inspection level
Composition analysis	3.2	4.4.2	1/
Hardness	3.3	4.4.1	S-2
Copper sulfate test	3.6 and GG-I-526	GG-I-526	S-2
Boil test	3.6 and GG-I-526	GG-I-526	S-2

1/ A sufficient quantity from each lot of raw metal used in the construction of the scalers shall be selected for composition analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.3 Examination. The scalers shall be examined to determine compliance with the requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include but shall not be limited to, the defects listed in table IV.

Table IV. Classification of Defects

Category	Defect
Major A:	
101	Type not as specified.
102	Style and design not as specified.
103	Scaling edge not properly sharpened.
104	Scaler not free from feathers or burrs.
105	Operative end finish not as specified.
Major B:	
151	Operative end dimensions not as specified.
Minor A:	
201	Handle or shank finish not as specified.
202	Handle or shank finish not free from coarse grind marks or scratched surface areas.
203	Handle not straight and flat.
204	Identification marking not present or improperly located, illegible, or incomplete.
205	Scaler not free from dirt, polishing compound, or foreign matter.
206	Packaging or packing not as specified.
207	Marking on packaging and pack not as specified.
Minor B:	
251	Dimensions, other than 151, not as specified.
252	Handle serrations not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.4 Test procedures.

4.4.1 Hardness. The hardness shall be determined in accordance with ASTM Designation E18 or E92.

4.4.2 Composition analysis. In the absence of certificates of quality, the scalers shall be subjected to chemical analysis in accordance with Method 111 or 112 of FED-STD-151. In the event of conflict, Method 111 shall govern.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A, or C, as specified (see 6.1).

5.1.1 Unit of issue. One scaler, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. No preservative is required.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class domestic. Boxes shall be closed to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers, of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B, or C, as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and shall be constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table II, using type III, class 1 plywood; or shall be designed for type 2 load and shall be constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers, of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with FED-STD-123.

5.3.1.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.1.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type, or type and style (type I only) of scaler required.
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, styles, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types and styles which are normally procured to meet Federal Government requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:54Z
Check the source to verify that this is the currentversion before use.