FED. SUP CLASS
6640

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - AS, YD
AIR FORCE - 82



PART NUMBER	CAPACITY mL*	LENGTH OF STEM mm	STOPPER NUMBER ℸ	STOPCOCK NUMBER ℸ
MS35950-1	60	220 ± 10	16	2

* +5% to 35% in excess of listed capacity.

MATERIAL: Funnel, stem, stopper, and stopcock shall be of borosilicate glass in accordance with ASTM E438, Type I, Class A. Stopcock plug shall be of borosilicate glass as per specification above or polytetrafluoroethylene in accordance with Federal Specification L-P-403.

DESIGN: Funnel shall be in accordance with Federal Specification NNN-F-800, Type III. Funnel shall be symmetrical in shape with stem and stopper, and fitted with a stopcock. Funnel shall have a roughened area suitable for marking. Stopcock shall have a leakproof fit with the funnel with which it is originally supplied. Each stopcock plug shall be provided with a retaining device to prevent accidental dislodging. Funnel shall have a dropper with 1.5 mm opening inside bulb at top of stem for dispensing single drops.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 18575 and 22527.

APPROVED 25 Sep 1959 REVISED (A) 20 June 1969 (B) 11 Dec 80 (C) 25 JAN 1984

(C) ENTIRE STANDARD REVISED

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA Other Cust MS 99		FUNNEL, SEPARATORY, FOR SINGLE DROPS	MS35950
Procurement Specifications NNN-F-800	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1205

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:31Z
Check the source to verify that this is the current version before use.