DATA ITEM DESCRIPTION

Form Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 110 hours per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.

1. TITLE	2. IDENTIFICATION NUMBER
SUMMARY OF PRECISION MEASUREMENT EQUIPMENT LABORATORY (PMEL) TOTAL QUALITY PROGRAM REVIEWS	DI-QCIC-81111A

3. DESCRIPTION / PURPOSE

3.1 This report allows the staff to monitor and manage the PMEL program. The data provides figures on monthly production, daily workable backlog, percent ot items awaiting parts, and Quality Process Reviews (QPRs).

4. APPROVAL DATE (YYMMDD)	5. OFFICE OF PRIMARY RESPONSIBILITY (OPR)	6a. DTIC APPLICABLE	6b. GIDEP APPLICABLE
950731	AFSPC/LGMOP		

7. APPLICATION / INTERRELATIONSHIP

7.1 This data item description (DID) contains the format and content preparation instructions for the data product generated by the specific and discrete task requirement as delineated in the contract.

7.2 This DID supersedes DI-MISC-81111.

8. APPROVAL LIMITATION	9a. APPLICABLE FORMS	9b. AMSC NUMBER
	NONE	F7146

10. PREPARATION INSTRUCTIONS

10.1 Format. Contractor format is acceptable.

10.2 Content. Data per calendar month shall include:

a. Total number of items scheduled into PMEL.

b. Total number of items processed through the PMEL.

c. Total number of items produced and certified (subject to Quality Process Review (QPR)).

d. Total number of QPRs completed.

e. Total number of QPRs with nonconformities and a list of items with nonconformity codes (See Table 1).

f. Total number of In Process Reviews (IPRs) completed.

(Continued on page 2)

11. DISTRIBUTION STATEMENT

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T10:43Z
Check the source to verify that this is the current version before use.

Block 10, Preparation Instructions (continued)

g. Total number of IPRs with nonconformity's and a list of items with nonconformity codes (See Table 1).

h. Total number of Working Standard Reviews (WSRs) completed.

i. Total number of WSRs with nonconformity's and a list of items with nonconformity codes (See Table 1).

j. Total number of items Awaiting Parts (AWP) and number of items AWP over 45 days.

k. Total number of items deferred (other than AWP) and, number of items deferred (other than AWP) over 30 days, and their status.

TABLE I Nonconformity codes

A2	The item is completely inoperative.
A3	The item has a function that is inoperative.
A4	The item does not meet calibration accuracy's for all parameters certified.
A5	The item has a physical deficiency which affects or may affect its operation (a knob that binds or metallic FOD)
B5	The item has a physical deficiency which does not affect its operation (cracked knob, missing screw).
A6	The item is not clean and its condition could affect operation (filter clogged which could cause overheating, precision machined surfaces dirty).
B6	The item is not clean, its condition does not affect operation.
A7	The item has a documentation error which affects the accuracy or reliability (T.O. directed limitation omitted, date due calibration is extended past the authorized interval).
B7	The item has a documentation error which does not affect the accuracy or reliability.
A8	The item has an intermittent function which affects the accuracy or reliability.
B8	The item has an intermittent function which does not affect the accuracy or reliability.
L1	Current technical order not used, or technical order used contained an error requiring an AFTO Form 22 (affects the accuracy or reliability of item).
L2	Current technical order not used, or technical order used contained an error requiring an AFTO Form 22 (does not affect the accuracy or reliability of item).
L3	Incoming or outgoing inspections were not performed or were performed incorrectly (affects safety or operation of item).
L4	Incoming or outgoing inspections were not performed or were performed incorrectly (does not affect safety or operation of item).
L5	A full calibration could not be performed on the item (other than approved limitations)
L6	An overdue standard was used or a standard was incorrectly substituted during calibration of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.

L7	Item was calibrated in an out-of-tolerance environmental condition (condition could affect accuracy of calibration).
L8	Calibration fixtures, accessories, or tools were not used or were improperly used to facilitate calibration or repair of the item.
L9	Technician was not signed off on task in On-the-Job-Training (OJT) records.
M1	Item was placed in awaiting parts status an excessive amount of times.
M2	Replaced parts were not documented in Maintenance Data Collection (MDC).
M3	Incorrect Urgency of Justification Code (UJC) or Serviceability Recoverability Designator (SRD) codes were used when ordering parts.
M4	Materiel Deficiency Report (MDR)/Quality Deficiency Report (QDR) not submitted for defective parts or defective new item.
M5	Incorrect part number ordered or received causing delay
M6	Parts on order excessive amount of time with no follow-up action.
M7	Parts on bench stock, shop stock, or work order residue not utilized.
P1	Item received dirty or had physical defect that was not documented.
P2	Item received without required documentation, technical data, or accessories.
P3	Item received overdue calibration an excessive amount of time.
P4	Item in awaiting maintenance, hold, or deferred status an excessive amount of time.
P5	Item data not correctly reflected in PMEL Automated Management System (PAMS) database or item part number was not in 33K-1-100 (Series) and an AFTO Form 45 was not submitted.
P6	Item remained in awaiting customer pickup status an excessive amount of time.
P7	Item not appropriately documented or packaged for off-base shipment.
N1	Indicates the Process Review is in-progress or not yet complete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.