DATA ITEM DESCRIPTION

Title: Summary of Precision Measurement Equipment Laboratory (PMEL) Total Quality Program Reviews

Number: DI-QCIC-81111B
Approval Date: 19990520
AMSC Number: F7332
Limitation:
DTIC Applicable:
GIDEP Applicable:
Office of Primary Responsibility: AFMETCAL Det1/MLSP
Applicable Forms:
Use, Relationship:
This report allows the MAJCOM and AFMETCAL Det 1 staffs to monitor and manage the individual PMEL quality program operations. Monthly summaries of quality program related information are required from each PMEL. The summaries contain the quantity and the reason for the defects found during the Reviews performed by the PMEL under the Total Quality Program. The reason for the defects are assigned specific codes. Table 1 contains those codes assigned, as well as the type of defect. It also drives some standardization in the input.

a. This Data Item Description contains the format, and content preparation instructions for the data product generated by the specific and discrete task requirement as delineated in the contract**.**

b. This DID supersedes DI-QCIC-81111A.

Requirements:
1. Contractor format is acceptable.

2. Reference Documents: The applicable issue of the documents cited herein, including the approval dates of any applicable amendments, notices and revisions, shall be as cited in the current issue of the Department of Defense Index of Specifications and Standards (DODISS) at the time of the solicitation; or for non-DODISS listed documents, as stated herein.

3. Content of the summaries shall include the quantity and reason for the defects found. Use Table 1 for those specific codes for the defects found in accordance with T.O.00-20-14(current edition). This T.O. can be obtained at AFMETCAL Detachment 1/MLSR, 813 Irving-Wick Dr. W. Ste 4M, Heath OH 43056-6116:

A. In Process Reviews (IPRs)
   Total Input (total Job Control Numbers (JCN) created)
   Total number of random IPRs selected
   List of IPR items with critical non-conformities (NC) (i.e. list PMEL Automated Management System(PAMS) label number and NCs)
   Percent of total input selected for random IPR
   Total number of non-randomly selected (target) IPRs
   List of non-randomly selected (target) IPRs completed
   List of target IPRs with critical non-conformities (PAMS label number and NCs)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.

B. Quality Process Reviews (QPRs)
- Total production subject to QPR
- Total number of items selected for QPR
- Total number of QPRs completed
- Percent of total production selected for QPR
- QPR items with critical non-conformities (PAMS label number and NCs)

C. Working Standards Review (WSR)
- Total number of standards available for WSR
- Total number of items selected for WSR
- Total number of WSRs completed
- Total number of critical/high use WSRs completed
- Percent of standards selected for WSR
- List of WSRs with critical non-conformities (PAMS label number and NC)

D. List of all bypassed reviews with justification for each review bypass

Table 1

SUMMARY OF Total Quality Plan Non-conformity CODES
(Critical non-conformity codes are identified with an asterisk (*))
General Codes:

A02*	The item is unsafe or hazardous to use
A03*	The item is completely inoperative
A04*	The item has a function that is inoperative
A05*	The item does not meet calibration uncertainties for all parameters certified
B05	The item has a physical deficiency which affects or may affect its operation
A06*	The item has a physical deficiency which does not affect its operation
B06	The item is not clean and its condition could affect the item’s operation
A07*	The item is not clean and its condition does not affect the item’s operation
B07	The item has a documentation error which affects the item’s accuracy or reliability
A08*	The item has a documentation error which does not affect the item’s accuracy or reliability
B08	The item has an intermittent function which affects the item’s accuracy or reliability
	The item has an intermittent function which does not affect the item’s accuracy or reliability

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.

Specific Process Codes:

L01*	Current technical order not used, or technical order used contained an error requiring an AFTO Form 22 (affects the item's accuracy or reliability)
L02	Current technical order not used, or technical order used contained an error requiring an AFTO Form 22 (does not affect the item's accuracy or reliability)
L03*	Incoming or outgoing inspections were not performed or were performed incorrectly (affects the item's safety or operation)
L04	Incoming or outgoing inspections were not performed or were performed incorrectly (does not affect the item's safety or operation)
L05*	A full calibration could not be performed
L06*	An overdue standard was used or a standard was incorrectly substituted during calibration of the item
L07	The item was calibrated in an out-of-tolerance environmental condition (condition could affect accuracy of calibration)
L08	Calibration fixtures, accessories, or tools were not used or were improperly used
L09	Technician was not signed off on task in OJT records

Material Control/Repair Process Codes:

M01	Item was in awaiting parts status an excessive amount of time
M02	Replaced parts were not documented in MDC
M03	Incorrect UJC or SRD codes were used when ordering parts
M04	Deficiency Report (DR) not submitted for defective parts or defective new item
M05	Incorrect part(s) ordered or received causing delay
M06	Part(s) on order an excessive amount of time with no follow-up action
M07	Part(s) on bench stock, shop stock, or work order residue not used

Production Control Process Codes

P01	Item received dirty or had physical defects that were not documented
P02	Item received without required documentation, technical data, or accessories
P03	`Item received overdue calibration an excessive amount of time
P04	Item in awaiting maintenance, hold, or deferred status an excessive amount of time
P05	Item's data not correctly reflected in PAMS database or item part number was not in 33K-1-100 and an AFTO Form 45 was not submitted
P06	Item remained in awaiting customer pickup status an excessive amount of time
P07	Item not appropriately documented or packaged for off-base shipment

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.

Status of Review:

N00 No Defect
N01 Review in process-not completed

Root Cause Codes:

Non-conformities attributed to calibration tech data

C01 Wrong or invalid calibration technical data used
C02 Calibration technical data in error (wrong specifications, required equipment not accurate enough to perform calibration
C03 Calibration technical data ambiguous (allows for more than one interpretation)

Non-conformities attributed to environmental conditions exceeding limits

E01 Temperature exceeded limits at time of calibration
E02 Humidity exceeded limits at time of calibration

Non-conformities attributed to internal failure of the TMDE reviewed

F01 Component failure caused degradation or hard failure
F02 Failure related to previous faulty maintenance or repair
F03 Failure related to thermal heating

Non-conformities related to human error

H01 Human error: Used when a lack of proficiency is not a contributing factor and a simple oversight or omission on the part of the individual is determined to be the root cause

Nonconformity attributed to an inadequate or faulty process

I01 Process inadequate or faulty: Used when root cause analysis indicates all other root cause codes are not applicable. Pertains mainly to processes not associated with the calibration process

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.

Non-conformities attributed to the standards used to perform calibration

S01 Standard was out of tolerance
S02 Standard was overdue calibration
S03 Standard was limited for the function or accuracy used
S04 Standard was intermittently malfunctioning
S05 Accessories or calibration fixtures were faulty

Non-conformities attributed to training deficiencies

T01 OJT insufficient (trainer demonstrates task proficiency)
T02 OJT insufficient (trainer does not demonstrate task proficiency)
T03 Incorrect substitution of standards
T04 Basic technical school deficiencies
T05 Advanced technical school deficiencies

Non-conformities attributed to Owner/User of equipment being reviewed

U01 A defined Owner/User responsibility was not accomplished (Owner/User was not properly trained on responsibilities)
U02 A defined Owner/User responsibility was not accomplished (Owner/User was properly trained on responsibilities)
U03 Owner/User did not receive required schedules, listings, or documentation required to fulfill responsibilities
U04 Unique mission requirements precluded Owner/User from accomplishing required responsibilities

End Of DI-QCIC-81111B

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:43Z
Check the source to verify that this is the current version before use.