MIL-C-37008A
31 MARCH 1987
SUPERSEDING
MIL-C-37008
19 September 1973

MILITARY SPECIFICATION

CLAMP, ARTERY (DeBakey)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an aortic aneurysm clamp of the DeBakey design.

1.2 Classification. The clamp shall be of three sizes, as specified in the invitation for bids:

Sizes	Overall Length (in inches)
1	10
2	12-1/4
3	11-3/4

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Set-up.

MILITARY

MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

2.1.2 Drawings. The following drawings form a part of this specification to the extent specified herein. Unless otherwise specified, the issue shall be that in effect on the date of the solicitation.

DRAWINGS

Defense Personnel Support Center (DPSC)

20900	Teeth Standard For Cardiovascular Instruments
20905	Clamps, DeBakey, Aortic Aneurysm

(Copies of specifications, standards and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM) STANDARDS

E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

E 92 - Vickers Hardness of Metallic Materials, Tests for.

E 140 - Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness and Knoop Hardness).

F 899 - Stainless Steel Billet, Bar and Wire for Surgical Instruments.

F 921 - Standard Definitions of Terms Relating to Hemostatic Forceps.

F 1026 - General Workmanship and Performance Measurements of Hemostatic Forceps.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The clamp shall be in accordance with the applicable requirements of GG-I-526, and as specified herein. The terms used to describe the clamp in this specification shall be in accordance with ASTM F921, and as specified herein.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:46Z
Check the source to verify that this is the current version before use.

3.2 Material. The clamp shall be fabricated from Class 4 martensitic stainless steel, in accordance with ASTM F899.

3.3 Hardness. After proper stress relieving, the clamp shall have a Rockwell hardness of C40 to C48, with the exception of the screw or pin.

3.4 Style, design and dimensions. Shall be an aortic aneurysm clamp of the DeBakey design. The clamp shall be in accordance with the Defense Personnel Support Center Drawing Nos. 20900 and 20905.

3.5 Locks. The clamp shall have a box type lock. The box lock construction shall be be in accordance with the applicable requirements of ASTM F1026.

3.6 Finish. Finish shall be satin. Locks and ratchets shall be smooth and shall be polished, but need not be buffed to a mirror finish.

3.6.1 Passivation. The clamp shall be properly passivated.

3.8 Identification markings. In addition to the requirements specified in ASTM F1026, clamps of foreign manufacture shall be marked "stainless."

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of assuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling in quality conformance does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to acceptance of defective material.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection shall be in accordance with GG-I-526.

4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.4 Certificates of quality. Certificate of quality supplied by the manufacturer of the clamp may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1. For examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicted in Table I. Unit of product for examination purposes shall be one clamp.

TABLE I. Sampling criteria for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as indicated in Table II. The unit of product for test purposes shall be one clamp.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

TABLE IV. Sampling Criteria for tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
Hardness	3.3 and GG-I-526	GG-I-526	S-2	*
Dimensions	3.4	---	S-2	2.5
Copper sulfate test	GG-I-526	GG-I-526	S-2	*
Boil test	GG-I-526	GG-I-526	S-2	*

* Acceptance number shall be zero for all sample sizes.

** Four ounces of each lot of metal used in the construction of the clamp shall be selected for chemical analysis. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The clamp shall be examined to determine compliance with all requirements contained in this specification. Examination shall as specified in Tables III and V. Examination, however, shall not be restricted to defects listed in Table V. Any syringes in the sample containing one or more defects shall be rejected and, if the number of defective syringes in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE IV. Classification of defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified.
102	Clamp not free from cracks, pits, dents, nicks, sharp edges and burrs.
103	Serrations not as specified.
104	Ratchet teeth not uniform, or function properly.
105	Clamp not of box lock design.
106	Clamp not free from excessive play in box lock area.
107	Box lock does not provide a smooth opening and closing action.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

TABLE IV. Classification of defects (Cont'd)

Categories	Defects
Minor 201 202	Finish not mirror or matte, and not uniform. Identification markings not present, not correct, not complete, or not legible.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness test. Hardness test shall be determined in accordance with ASTM Standards E 18 or E 92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Standard E 140.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Clamps shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each clamp shall be packaged in accordance with MIL-P-116, Method IC-1.

5.1.1.4 Intermediate package. Twelve unit packages containing clamps of one size only shall be packaged in a box conforming to PPP-B-566, PPP-B-636, Class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

5.1.2 Level C. The level C preservation for clamps shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Eight intermediate packages containing clamps of one size only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601 overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Eight intermediate packages containing clamps of one size only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636 class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. The level C packing for clamps shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in a exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length, and shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Clamps packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Clamps packed as specified in 5.2.3 shall be unitized as specified in 5.2.5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Intended use. The clamp is intended for use for temporary occlusion of the aorta in resection of fusiform aneurysms of the aorta and restoration of continuity by aortic grafts. It is also used for occlusion of the neck of sacciform aneurysms of the aorta in tangential excision of such aneurysms and subsequent aortorraphy.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(d) National Stock Number.

6.3 This specification covers the following National Stock Numbers:

National Stock No.	Item Identification
6515-00-690-3194	CLAMP, ARTERY, DeBAKEY, Curved Jaws, Aortic Aneurysm, 10 Inch, 8.5 cm Jaw
6515-00-690-3195	CLAMP, ARTERY, DeBAKEY, Curved Jaws, Aortic Aneurysm, 12-1/4 Inch, 12 cm Jaw
6515-00-690-3196	CLAMP, ARTERY, DeBAKEY, Curved Jaws, Aortic Aneurysm, 11-3/4 Inch, 10.5 cm Jaw

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

6.4 Subject term (key word) listing.

Artery clamp
Clamp
DeBakey clamp

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3472

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:46Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-C-37008A	CLAMP, ARTERY (DeBakey)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:46Z
Check the source to verify that this is the current version before use.