MIL-C-37028A
22 September 1978
SUPERSEDING
MIL-C-37028
25 January 1974

MILITARY SPECIFICATION

CATHETERIZATION SET, URETHRAL, DISPOSABLE WITH 30cc SYRINGE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of sterile disposable assembly, prepackaged, ready for use, suitable for use in urethral catheterization of human adults.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636 - Boxes, Shipping, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-Packaging, Methods of.
MIL-S-36586 - Sterilization Test Strip Set, Bacterial Spore.
MIL-P-37513 - Povidone-Iodine Solution, USP.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION (FDA)

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States (USP)

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. The material used shall be as specified herein and of the highest quality, suitable in all respects for the purpose intended. The materials of the components shall be as follows:

3.1.1 Tray. The tray shall be fabricated from inert, virgin plastic.

3.1.2 Absorbent balls or premoistened swab tips. Absorbent material shall be of natural cotton or viscose rayon.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:48Z
Check the source to verify that this is the current version before use.

3.1.3 Absorbent waterproof underpad. Absorbent waterproof underpad shall be three-ply, laminated, wet strength alpha cellulose stock, coated on one side, or backed with a suitable plastic film to provide an impermeable barrier.

3.1.4 Fenestrated drape. Fenestrated drape shall be three ply, laminated, wet strength alpha cellulose stock.

3.1.5 Lubricating jelly. Lubricating jelly, with or without a topical anesthetic, shall be sterile, bacteriostatic, and water soluble.

3.1.6 Cleaning solution. Cleaning solution shall be 10% Povidone Iodine Solution, USP, in accordance with requirements of MIL-P-37513, except that the packaging shall not apply.

3.1.7 Specimen container. Specimen container shall be made of suitable glass or plastic.

3.1.8 Examining gloves. Plastic examining gloves shall be made from an inert, flexible, elastomeric film or sheet.

3.1.9 Syringe. The disposable syringe shall be made of a suitable plastic material and shall have a Luer taper tip with cap.

3.1.10 Forceps. Forceps, when supplied, shall be fabricated of a suitable plastic.

3.2 Style, design, and construction. Sets shall be sealed, as specified in 5.1.1.1, ready for use and disposable. All components of the set, except the cleaning solution, shall be sterile. The set shall provide for preparation of the patient, protection of contiguous areas of the bed, obtaining and transporting specimens and clean-up of the patient in nursing-unit procedures. The set shall consist of the following components which shall be in accordance with the requirements specified herein:

Components of the set	Quantity required
Plastic tray (700 ml min.)	1
Absorbent balls or premoistened swabs	5 (min.)
Absorbent waterproof underpad	1
Fenestrated drape sheet	1
Bacteriostatic lubricating jelly	1 pkg. (5 gm min.)
Specimen container (60 ml min.)	1
Plastic examining gloves	1 pair
Syringe (30 ml) filled with sterile distilled water	1
Cleaning solution (Povidone-Iodine Solution)	1 container (30 ml)
Forceps, plastic (if applicable)	1

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:48Z
Check the source to verify that this is the currentversion before use.

3.2.1 Plastic tray. The tray shall be 2-1/2 ± 1/8 inches (6.35 ± .32 centimeters (cm) deep by 4-1/2 + 1/8 inches (11.43 ± .32 cm) wide by 8-1/2 ± 1/8 inches (21.59 ± .32 cm) long. Tray may be compartmented or noncompartmented as long as a separate container is provided for absorbent balls and/or swabs. The tray (main portion if compartmented) shall have a fluid capacity of at least 700 ml. The long sides of the tray shall be graduated from 100 to 700 ml in 100 ml increments. Graduations shall be legible from outside or inside the tray.

3.2.2 Absorbent balls or premoistened swab tips. The absorbent material, when dry, shall measure approximately 1-1/8 inches (2.86 cm) minimum, in diameter, and shall meet the requirements specified in 3.1.2. Material shall be suitable for use as absorbent wipes and sponges.

3.2.3 Absorbent waterproof underpad. The underpad shall have dimensions of 19 + 1/2 inches (48.26 ± 1.27 cm) by 16-1/2 ± 1/2 inches (41.91 ± 1.27 cm), and shall meet the requirements specified in 3.1.3.

3.2.4 Fenestrated drape. The fenestration shall provide for access to the urethral orifice while protecting the contiguous pubic areas and shall have dimensions of 3-3/4 ± 1/4 inches (9.52 ± .64 cm) square. The drape shall have overall dimensions of 18 ± 1/4 inches (45.72 ± .64 cm) by 13 ± 1/4 inches (33.02 ± .64 cm). Drape shall meet the requirements specified in 3.1.4.

3.2.5 Bacteriostatic lubricating jelly. At least 5.0 grams (gm) of sterile, bacteriostatic, water-soluble lubricating jelly, with or without a topical anesthetic, shall be supplied with each set. Jelly shall be packaged in a suitable foil envelope which shall be labeled to meet the requirements of the Food, Drug and Cosmetic Act. Lubricant shall be tested as specified in 4.4.1 and 4.4.2.

3.2.6 Cleaning solution. A minimum of 30 millimeters (ml) of the solution shall be furnished in a suitable plastic or glass container. As an alternate, when premoistened swabs are furnished, a sufficient quantity of the solution shall be contained therein to saturate the swabs. Labeling shall meet the requirements of the Food, Drug, and Cosmetic Act.

3.2.7 Specimen container. The container shall have a minimum capacity of 60 ml and shall be fitted with a water-tight closure.

3.2.8 Examining gloves. The gloves shall be of universal size, prepowdered and packaged, or folded, to facilitate putting them on the wearers hands without violating aseptic technique. Gloves shall be free of seams or welts in the palm or finger-tip areas. Gloves shall conform to the human hand after not more than one minute's wear.

3.2.9 Syringe. Syringe shall be graduated from 0 to 30 ml in 1 ml increments. All calibrations shall be clear and legible. Syringe shall be filled with a nominal 30 ml of Sterile Water for Injection, USP.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

3.2.10 Forceps. Forceps shall be spring type, 5 ± 1/4 inches long. The forceps need not be furnished when supplied with premoistened swabs.

3.3 Sterility. All components of the set, except the cleaning solution, shall be sterile. The lubricating jelly, gloves and syringe water shall be sterile when tested with a biological indicator as specified in 4.4.2.

3.3.1 Certificates of sterility. A certificate of sterility, stating that samples from each sterility lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item Identification.
b. National Stock Number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of tests.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, ATTN: DPSC-ATQ, 2800 South 20th Street, Philadelphia, PA, 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, the certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and sterilized simultaneously in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.4 Instructions. Each set shall include procedural instructions for use.

3.5 Marking. The lubricating jelly foil package and cleaning solution container or premoistened swabs container shall be labeled to meet the requirements of the Food, Drug, and Cosmetic Act.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

3.6 Expiration dating period. The expiration dating period shall be not less than 36 months..

3.7 Delivery. Not more than 4 months of the expiration dating period shall have elapsed at the time of delivery to the Government.

3.8 Workmanship. Workmanship shall be first class throughout. Catheterization sets, proper and individual components, shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.1.2 Inspection. Inspection, as used in this specification, is defined as bo examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the components, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:48Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one set.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for test purposes shall be one set. Where possible, the same sample unit shall be used for the determination of one or more test characteristics.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Lubricating jelly	Bacteriostasis	3.2.5	4.4.1	S-2 1/	2/
	Sterility	3.3	4.4.2	3/	2/
Gloves	Sterility	3.3	4.4.2	3/	2/
Water in syringe	Sterility	3.3	4.4.2	3/	2/

1/ Sampling shall be on component lots.
2/ Acceptance number shall be zero.
3/ Sampling shall be as indicated in the USP.

4.3 Examination. The sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in tables I and II. Examination, however, shall not be restricted to defects listed in table III. Any sets in the sample containing one or more defects shall be rejected and, if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Set incomplete (one or more components missing).
102	Plastic or rubber components not smoothly finished, not free from pits, blisters, scratches, tears, cracks, nicks, bubbles, roughness, tackiness, bloom or crazing, as applicable.
103	Components not free from foreign matter.
104	Sheet or drape not free from cuts or tears.
105	Dimensions of trays not as specified.
106	Labeling of cleaning solution, premoistened swabs, or lubricating jelly containers not as specified.
107	Tray does not have a minimum capacity of 700 ml.
108	Gloves not pre-powdered.
109	Specimen container not water-tight.
Minor	
201	Packaging, packing or marking not as specified.
202	Instructions missing, incomplete or illegible.
203	Poor workmanship (not resulting in a major defect).
204	Dimensions of sheets, drapes, balls or swabs not as specified.
205	Specimen container does not have a minimum capacity of 60 ml.
206	Quantity of lubricant or cleaning solution not as specified.
207	Tray leaks.
208	Premoistened swabs do not adhere firmly to handles.

4.3.1 Packaging inspection. The inspection of the packaging, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Bacteriostatic test. The lubricating jelly shall be bacteriostatic, as supplied, when tested by the 'Agar-Plate Method', to determine the zone of inhibition as described in the Methods of Testing Antiseptics and Disinfectants of the FDA, Circular No. 198.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:48Z
Check the source to verify that this is the current version before use.

4.4.2 Sterility test. The sets shall be tested for sterility in accordance with the method specified in the USP for the appropriate component using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in 4.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization and 30 for type II sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in table III.

4.4.2.1 Inoculated product or simulated product. For each type of sterilization, the organism shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. steraothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.2.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.2 and 4.4.2.1. The method of sterilization will control the one organism in the spore strip to be tested for, as shown in 4.4.2.1. When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.2.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein, shall submit proof that the alternate method provides equal assurance, concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

4.4.2.4 Product testing. Testing for sterility of the product, after receipt by the Government, shall be without a biological indicator and in accordance with the USP.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit packaging. Components comprising each set shall be placed within the plastic tray. The plastic tray shall be wrapped in such a manner as to assure an uncontaminated working surface when opened. The wrapped tray may be packaged in a paperboard box. The wrapped tray, with or without the box, shall be further packaged in a clear, polyethylene bag having a minimum wall thickness of 0.003 inch. Excess air shall be evacuated from the bag and the closure effected by heat sealing. The package shall be capable of maintaining sterility of contents unless damaged or opened. As an alternate, each wrapped tray shall be packaged in a styrene container having a peel-back, plastic-coated paper, heat sealed to the top of the container and shall not be capable of being resealed. The container shall be capable of maintaining sterility of contents unless damaged or opened. The complete package shall pass the leak test as specified in MIL-P-116.

5.1.2 Level C. Each disposable catheterization set shall be clean, dry and packaged sterile in such a manner as to afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity. The package shall be capable of maintaining sterility of contents unless damaged or opened.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Twenty unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Twenty unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged units shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Unit package. In addition to any commercial markings, each unit package shall be marked in accordance with MIL-STD-129 and the requirements of the Food, Drug, and Cosmetic Act. Lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat) shall be shown, following the level of packing and date of packing. Example: "A-1/78-G." Final sterilization shall be on or after indicated packaging date. In addition, marking shall include the legend: "CONTENTS OF INNER WRAP ARE STERILE UNLESS PACKAGE IS DAMAGED OR OPENED EXCEPT FOR PVP IODINE SOLUTION WHICH IS NOT TERMINALLY STERILIZED", or similar statement.

5.3.2 Shipping containers. Shipping container shall be marked in accordance with MIL-STD-129. Lot (control) number and the method of sterilization as specified in 5.3.1 shall also be shown.

6. NOTES

6.1 Intended use. The catheterization sets are intended for use in adult urethral catheterization procedures and shall provide for preparation of the patient, protection of contiguous areas of the bed, obtaining and transporting specimens and clean up of the patient in nursing unit procedures.

6.2 Ordering data. Procurement documents shall specify the following:

a. Title, number, and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-117-9756	CATHETERIZATION SET, URETHRAL, Disposable, with 30 ml syringe.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:48Z
Check the source to verify that this is the current version before use.

6.4 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD-MB

Preparing activity:
DLA-DM

Project No. 6515-2072

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-C-37028A CATHETERIZATION SET, URETHRAL, DISPOSABLE, with 30 ml syringe

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE-ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:48Z
Check the source to verify that this is the current version before use.