MILITARY SPECIFICATION

FORCEPS, KIDNEY PEDICLE, GUYON-PEAN, 9 INCH

This Amendment forms a part of Military Specification MIL-F-37003 dated 11 September 1973, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

Add the following paragraph:

"3.8 Jaw and ratchet operation. When tested as specified in 4.4.1, the jaw tips shall close first and progressively tighten with each ratchet position. Each ratchet position shall hold firmly and release instantly."

Page 5:

4.2.2 Delete entirely and substitute:

"4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. When feasible the same sample unit shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one forceps.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Forceps	Hardness	3.2	GG-I-526	S-2	1.5
	Distortion test	GG-I-526	GG-I-526	S-2	2.5
	Dimensions	3.4	---	S-2	2.5
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Box lock stiffness	3.5	GG-I-526	S-2	6.5
	Jaw and ratchet operation test	3.8	---	S-2	2.5

."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:45Z
Check the source to verify that this is the current version before use.

Page 5: cont'd

Add the following paragraph:

"4.2.3 Raw materials. Four ounces of each lot of metal used in the manufacture of the forceps shall be selected for chemical analysis."

Page 6:

Add the following paragraph:

"4.4.1 Jaw and ratchet operation test. The forceps shall be positioned to the first ratchet position, but not locked. Forceps with jaw tips not touching shall be rejected; further, should all the teeth touch, the forceps shall be rejected. The forceps shall then be engaged to each of the ratchet positions. Forceps that do not progressively tighten with each ratchet advance shall be rejected."

Page 9:

6.4 Delete entirely.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1877

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:45Z
Check the source to verify that this is the current version before use.