FED. SUP CLASS
6640



PART NUMBER	INSIDE DIAMETER mm	LENGTH mm	CAPACITY mL
MS36173-1	4.75 to 5	63 to 72	1

MATERIAL: Test tube shall be of borosilicate glass in accordance with ASTM E438, Type I, Class A.

DESIGN: Tube shall be in accordance with Federal Specification DD-T-200, Type I, as illustrated above.

The scale shall extend downward in 1 mm divisions to 40 mm.

Graduations shall be numbered at 0, 10, 20, 30, and 40 mm.

Graduation markings and numbers shall be clearly legible and shall remain so under ordinary usage.

Band 10 mm wide shall be sandblasted around top of tube.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 53629 and 84116.

APPROVED 30 Dec 1960 REVISED (A) 7 DEC 1983

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(A) ENTIRE STANDARD REVISED

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA Other Cust MS 99		TEST TUBE, BLOOD SEDIMENTATION, CUTLER TYPE	MS36173
Procurement Specifications: DD-T-200	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1208

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.