FED. SUP CLASS
6640

PART NUMBER	RATED DISTILLATION LITERS (GAL)/HOUR		MAXIMUM OVERALL HEIGHT cm (in.)	
MS36197-1	1	(.37)	87	(34)
MS36197-2	3	(.89)	115	(45)
MS36197-3	6	(1.58)	115	(45)
MS36197-4	8	(2)	33	(13)
MS36197-5	11	(3)	116	(45.5)

Ⓐ MATERIAL: The distillation components of the apparatus shall be made of borosilicate glass as specified in ASTM E 438, Type I, Class A, and teflon. The cabinet shall be aluminum.

DESIGN: Each unit shall be of such design as to perform the intended function and be capable of operation by heating with electricity of 110 to 240 volts AC or DC. All electrical components shall comply with the standards of the Underwriters Laboratories, Inc. Apparatus shall be equipped with automatic cutoff device to avoid overheating. The cabinet shall have a removable cover for servicing.

FINISH: All metal parts shall have a baked epoxy painted coating. All other surfaces shall resist corrosion.

PERFORMANCE: The apparatus shall produce distilled water conforming to the requirements of the United States Pharmacopeia when tested as specified in Federal Specifica- tion, GG-D-465A, paragraph 4.4.1. Water distillation rate shall be not less than the rate specified in the table.

NOTES:

1. Items that vary in design but which otherwise conform to requirements herein are considered one interchangeable Military Standard Item.

Ⓐ 2. In the event of a conflict between the text of this standard and the references cited herein, the text of this standard shall take precedence.

Ⓐ 3. Referenced Government documents of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DoDISS) specified in the solicitation form a part of this standard to the extent specified herein.

Ⓐ 4. FSCMs: 14674 and 05852

APPROVED 07 FEB 1980 REVISED A 27 MAR 1987

USER ACTIVITIES:
ARMY - MD
NAVY - AS, YD
AIR FORCE - 62

REVIEWER ACTIVITIES

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document

P. A. EA Other Cust MS, MB 99	INTERNATIONAL INTEREST	TITLE DISTILLING APPARATUS, LABORATORY water, unitized glass	MILITARY STANDARD MS 36197
PROCUREMENT SPECIFICATION NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672-1 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1415-3

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:36Z
Check the source to verify that this is the current version before use.