FED. SUP CLASS
6640

PART NUMBER	METAL TUBE	GLASS TUBE DIA., MM
MS36255-1		20.0
MS36255-2		22.0
MS36255-3		25.0
MS36255-4		28.0
MS36255-5		35.0
MS36255-6		41.0
MS36255-7		46.0

CANCELLED

CANCELLED AFTER 20 AUG 1991

NO SUPERSEDING STANDARD

MATERIAL: Adapters shall be in accordance with General Specification, DD-G-541, type 1 class A.

DESIGN: The tubes shall form a metal to glass seal. The kovar length, either 25 or 51 mm, overall length 20 or 23 cm respectively.

USE: Adapters are used for making glass to metal connections.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

(A) CANCELLATION

APPROVED 5 JAN 1983 REVISED (A) 20 AUG 1991

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA		ADAPTER, LABORATORY	MS36255
Other Custodians MS 99			
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

CANCELLATION

Project Number 6640-1725-

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

USER ACTIVITIES: ARMY - MD, EM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.