FED. SUP CLASS
6640

REVIEW ACTIVITIES: ARMY - MD, ME
NAVY - OS, AS, YD



PART NUMBER	APPROXIMATE BOWL CAPACITY milliliters	APPROXIMATE DIMENSIONS inches (mm)				RUBBER STOPPER SIZE NUMBER
		HEIGHT OVERALL	TOP OD OF BOWL	TOP OD OF BASE	MAXIMUM OPENING OF JAWS	
MS 36319-1	300	9 (229)	3-1/8 (80)	2 (51)	2-3/4 (70)	8

MATERIAL: Bowl and base shall be borosilicate glass, Federal Specification DD-G-541, Type I, Class A.

Clamp shall be of metal suitable for the purpose intended.

Stopper shall be rubber.

FINISH: Metal clamp shall be blue anodized.

DESIGN: Mating surfaces of the bowl and base shall be ground to a perfect fit.

The base shall have a fritted glass support to hold the filtering disk.

The spring clamp shall hold the mating surfaces of bowl and base together firmly, preventing leakage when the filter disk is in place.

The holder shall have an approximate effective filtering area of 9.6 square centimeters.

NOTE: 1. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 3 Sep 1963 REVISED (A) 31 MAR 1981

(A) ENTIRE STANDARD REVISED

This standard has been approved by the Department of the ... and is mandatory for use by that activity. All other military activities are required to employ this standard where suitable.

P.A.	TITLE	MILITARY STANDARD
EA	HOLDER, MICROPOROUS BACTERIAL FILTERING DISK	MS36319
Other Cust 99 MS		
Procurement Specification MIL-H-36376	SUPERSEDES:	SHEET 1 OF 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:28Z
Check the source to verify that this is the current version before use.

DD FORM 1 SEP 65 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED. Project Number 6640-0713