INCH-POUND
MIL-C-42066A
19 June 2000
SUPERSEDING
MIL-C-42066
15 June 1989

DETAILED SPECIFICATION

CARRIER, LITTER, WHEELED

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE.

1.1 Scope. This specification covers a wheeled, litter carrier for military field use. It is designed for individual issue and intended for use as described in section 6 of this specification.

2. APPLICABLE DOCUMENTS.

2.1 Government documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issue of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and Supplement thereto, cited in the solicitation (see paragraph 6.2).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Supply Center Philadelphia, ATTN: Directorate of Medical Materiel, DSCP-MBP, 700 Robbins Avenue, Building 6A, Philadelphia, PA 19111-5092, by using the self-addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

P-C-436	-	Cleaning Compound, Alkali, Boiling Vat (Soak) or Hydrosteam
FF-S-92	-	Screw, Machine, Slotted, Cross Recessed or Hexagon Head
FF-S-107	-	Screw, Tapping and Drive
QQ-A-200/8	-	Aluminum Alloy Tubes, Rods, Bars and Shapes, 6061
QQ-A-200/9	-	Aluminum Alloy 6063, Bar, Rod, Shapes, Tube, and Wire, Extruded
QQ-A-250/11	-	Aluminum alloy 6061, Plate and Sheet
QQ-P-35	-	Passivation Treatments for Corrosion-Resistant Steel
QQ-S-763	-	Steel Bars, Wire, Shapes and Forgings
QQ-W-766	-	Steel, Sheets, and Strip, corrosion Resisting
QQ-W-423	-	Steel, Wire, Corrosion Resisting
TT-E-527	-	Enamel, Alkyd, Lusterless, Low Voc Content
PPP-B-621	-	Box, Wood, Nailed and Lock-Corner

MILITARY

MIL-B-543	-	Buckles, Tongueless, and Webb Strap
MIL-T-5038	-	Tape, Textile and Webbing, Textile, Reinforcing, Nylon
MIL-S-7097	-	Steel, Carbon (1018, 1020 or 1025)
MIL-C-8514	-	Coating Compound, Metal Pretreatment, Resin-Acid
MIL-L-10547	-	Liner, Case and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible
MIL-P-15024	-	Plates, Tags, and Bands for Identification of Equipment
MIL-G-21164	-	Grease, Molybdenum Disulfide, for Low and High temperatures, NATO code No. G-353
MIL-P-23377	-	Primer Coatings: Epoxy High-Solids
MIL-W-45205	-	Welding, Gas Metal-Arc and Gas Tungsten-Arc Aluminum Alloys, Readily Weldable for Structures (Excluding Armor)
MIL-C-45662	-	Calibration Systems Requirements
MIL-C-46168	-	Coating, Aliphatic, Polyurethane, chemical Agent Resistant
MIL-C-53039	-	Coating, Aliphatic, Polyurethane, Single Component, Chemical Agent Resistant

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-595 - Colors Used In Government Procurement

MILITARY

MIL-STD-171 - Finishing of Metal and Wood Surfaces

(Unless otherwise indicated, copies of federal and military specifications, standards, and handbooks are available from Defense Printing Service Detachment Office, 700 Robbins Avenue, Building 4D (NPM-DODSSP), Philadelphia, PA 19111-5094).

2.1.2 Other Government documents, drawings and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

PUBLICATION

Medical Marking Standard No. 1

(Application for copies of the Marking Standard should be addressed to Defense Supply Center Philadelphia, ATTN: DSCP-MSCBP, 700 Robbins Avenue, Bldg 6A South, Philadelphia, PA 19111-5092.)

DRAWING

Defense Supply Center Philadelphia (DSCP)

60007FD - Carrier, Litter, Wheeled (Cage 89875)

(Application for copies of the drawing should be addressed to Defense Supply Center Philadelphia, ATTN: DSCP-MSCBA, 700 Robbins Avenue, Building 6A, Philadelphia, PA 19111-5092.)

2.2 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DOD adopted are those listed in the issue of the DODISS cited in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of the documents cited in the solicitation (see paragraph 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM documents.

A380 - Steel Parts, Equipment and systems, Cleaning and Descaling Stainless
D1974 - Standard Practice for Methods of Closing, Sealing and Reinforcing Fiberboard Shipping Containers
D5118 - Standard Practice for Fabrication of Fiberboard Shipping Boxes

IEEE/ASTM SI 10-1997 Standard for Use of International System of Unit (SI: The Modern Metric System)

(Applications for copies should be addressed to the American Society for Testing and Materials (ASTM), 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.)

AMERICAN NATIONAL STANDARDS INSTITUTE

ANSI documents.

ANSI/ASQC Z1.4 - Sampling Procedures and Tables for Inspection by Attributes

ANSI/AWS A2.4 - Welding AND Brazing and Nondestructive Examination

(Applications for copies should be addressed to the American National Standards Institute, 11 West 42nd Street, New York, NY 10036-8002)

2.3 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see paragraphs 4.5, 6.2 and 6.4). Any changes or deviations of carriers from the approved first article during production will be subject to the approval of the contracting officer. Approval of the first article shall not relieve the contractor of his obligation to furnish carriers conforming to this specification.

3.2 Materials. Materials shall be as specified herein and as shown on the applicable drawings. Materials not specified shall be selected by the contractor and shall be subject to all provisions of the specification.

3.2.1 Aluminum and aluminum alloys.

3.2.1.1 Aluminum plate and sheet, 6061 T6 alloy. Unless otherwise specified, the aluminum plate and sheet used in the construction of the items included in this specification shall meet the requirements of QQ-A-250/11 when inspected and tested as specified in paragraph 4.

3.2.1.2 Aluminum extrusions, 6061 T6 alloy. Unless otherwise specified, the extrusions (bar, rod, shapes, tube, and wire) used in the construction of the items included in this specification shall meet the requirements of QQ-A-200/8 when inspected and tested as specified in paragraph 4.

3.2.1.3 Aluminum extrusions, 6063 T52 alloy. Unless otherwise specified, the extrusions (bar, rod, shapes, tube, and wire) used in the construction of the items included in this specification shall meet the requirements of QQ-A-200/9 when inspected and tested as specified in paragraph 4.

3.2.2 Corrosion resisting steel (CRES), stainless steel.

3.2.2.1 Corrosion resisting steel plates, sheet, and strips. Unless otherwise specified, corrosion resisting steel plates, sheets, and strips used in the construction of the items included in this specification shall meet the requirements of QQ-S-766 when inspected and tested as specified in paragraph 4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.2.2.2 Corrosion resisting steel wire, stainless steel. Unless otherwise specified, corrosion resisting steel wire used in the construction of the end items included in this specification shall meet the requirements of QQ-W-423 when inspected and tested as specified in paragraph 4.

3.2.2.3 Corrosion resisting steel bars and rods. Unless otherwise specified, corrosion resisting steel bars and rods used in the construction of the items included in this specification shall meet the requirements of QQ-S-763 when inspected and tested as specified in paragraph 4.

3.2.3 Carbon steel. Unless otherwise specified, carbon steel 1018 bars and forging stock used in the construction of the items included in this specification shall meet the requirements of MIL-S-7097 when inspected and tested as specified in paragraph 4.

3.2.4 Nylon webbing. Unless otherwise specified, nylon webbing used in the construction of items included in this specification shall meet the requirements of MIL-T-5038 when inspected and tested as specified in paragraph 4.

3.2.5 Hardware and fasteners. Unless otherwise specified, all hardware and fasteners used in the construction of the items included in this specification shall be corrosion resisting materials or have corrosion resistant finishes. All hardware shall be inspected and tested as specified in paragraph 4.

3.2.5.1 Machine screws, corrosion resisting steel. Unless otherwise specified, all machine screws used in the construction of the items included in this specification shall meet the requirements of FF-S-92 when tested as specified in paragraph 4. All screws shall meet applicable material requirements for corrosion resistant steel.

3.2.5.2 Self-tapping screws, corrosion resisting steel. Unless otherwise specified, all self-tapping screws used in the construction of the items included in this specification shall meet the requirements of FF-S-107 for corrosion resisting steel screws except that all screws must be made of AISI 300 series stainless steels. All screws shall be inspected and tested as specified in paragraph 4.

3.2.6 Wheels. The carrier wheels shall be fabricated of toughened glass-reinforced nylon resin, that shall spring back from an impact load, with steel bearings, pneumatic tubes and tires.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.3 Design. The carrier shall be ruggedly designed to move patients over all types of terrain, through the doors of all hospital shelters, and to be maneuvered in wards, operating rooms, and x-ray for depositing or picking up patients. When the legs are down in locked position, on a level surface, the wheels shall be near but have no contact with the level surface. This configuration shall serve as a parking brake and also afford a stable examining table. The carrier shall also be able to be easily and quickly set up for use and collapsed for storage. The carrier shall be fabricated and constructed according to all drawings of this specification in paragraph 2. Each complete carrier shall not weigh more than fifty pounds (50 lbs.).

3.4 Dimensions. The dimensions and tolerances of the carrier included in this specification shall conform to DSCP drawing 60007FD.

3.5 Function and performance.

3.5.1 Operational and storage environment. The carrier covered by this specification shall operate reliably in basic cold and basic hot environments with daily cycles of temperature, solar radiation, and relative humidity. Storage, transit conditions, and additional environmental elements shall be as shown in Table I.

3.5.2 Function. The carrier covered by this specification shall function as designed when fabricated and assembled according to the drawings, standards, and specifications listed herein. When assembling the carrier, the portions of the frame shall snap firmly into place and be held there by the ball assembly. The wheels shall slide easily into the axle assemblies and be retained after the securing pin has been installed. The wheels shall rotate freely. There shall be little or no weight on the wheels when the legs are locked in the down position. When released, the legs shall move freely up against the frame and be held there by the action of the spring. A standard litter (NSN) National Stock Number 6530-00-783-7905, shall be held securely between the litter holders. The handwheel shall turn freely.

3.5.3 Operation and maintenance manual. An operation and maintenance manual shall be supplied with each carrier. The manual shall be secured to a non-contacting structural part of

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

the carrier and accessible in the operational or folded storage position. The attachment of the manual to the carrier shall not cause damage or eradication of manual text during shipping or storage as specified in paragraph 6.3.

TABLE I. Summary of temperature, solar radiation and relation humidity daily cycles

Climatic Design Type	Daily Cycle	Operational Conditions			Storage and Transit Conditions	
		Ambinet Air Temperature °F (°C)	Solar Radiation Bph (W/m^2)*	Ambient Relative Humidity Percent	Induced Air Temp. °F (°C)	Induced Relative Humidity Percent
Hot	Hot Dry	90 to 120 (32 to 49)	0 to 355 (0 to 1120)	3 to 8	91 to 160 (33 to 71)	1 to 7
	Hot-Humid	88 to 105 (31 to 41)	0 to 343 (0 to 1080)	59 to 88	91 to 160 (33 to 71)	14 to 80
Basic	Constant High Humid	Nearly constant 75 (24)	Negligible	95 to 100	Nearly constant 80 (27)	95 to 100
	Variable High Humid	78 to 95 (26 to 35)	0 to 307 (0 to 970)	74 to 100	86 to 145 (30 to 63)	19 to 75
	Basic Hot	86 to 110 (30 to 43)	0 to 355 (0 to 1120)	14 to 44	86 to 145 (30 to 63)	5 to 44
	Basic Cold	-5 to -25 (-21 to -32	Negligible	Tending toward saturation	-13 to -28 (-25 to -33)	Tending toward saturation
Cold	Cold	-35 to -50 (-37 to -46)	Negligible	Tending toward saturation	-35 to -50 (-37 to -46)	Tending toward saturation
Severe Cold	Severe Cold	-60 (Cold soak) (-51)	Negligible	Tending toward saturation	-60 (-51)	Tending toward saturation

*Bph (British thermal units per square foot per hour), W/m^2

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

(Watts per square meter)

3.6 Fabrication.

3.6.1 Welding. Unless otherwise specified, welding shall be durable, sound in every detail, and in accordance with the latest practices recommended by the American Welding Society in ANSI/AWS A2.4.

3.6.1.1 Arc welding aluminum alloys. All arc welding of aluminum alloys shall be according to MIL-W-45205 for class A weldments.

3.6.1.2 Plug welding. All plug welds shall be in accordance with ANSI/AWS A2.4. Opposing plug welds, 3/8-inch diameter, used on the litter holder assembly shall be capable of withstanding a shear load of 2500 pounds.

3.6.1.3 Heat treating. All 6061 welded aluminum assemblies shall be solution heat treated according to the following procedure:

a. Weldments shall be subjected to a controlled temperature of 985°F until all portions of the assembly have reached that temperature.

b. The weldments shall be quenched in a liquid medium until their temperature reaches 100°F or lower. Liquid medium temperature should not be allowed to exceed 100°F.

c. Weldments, if not held to their original shape by jigs, shall be readjusted immediately while they are in a soft pliable condition. If weldments cannot be immediately readjusted, they may be kept in a pliable condition by holding them in an environment of 25°F or lower.

d. After assuring the weldments are within the tolerances of their original shape, they shall be subjected to a temperature of 350°F for 8 hours and then allowed to cool to room temperature.

3.6.1.4 Cleaning aluminum alloys. All aluminum alloys shall be cleaned not more than 48 hours before welding. The aluminum shall be cleaned with an alkaline nonetching material conforming to P-C-436 or equal.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.7 Finish and color. After the last sentence add, "As an alternate to enamel finish, metal surfaces shall have an Epoxy Powder Coating (EPC) or an Urethane Polyester Coating (UPC) formulated for electrostatic spray equipment."

TABLE II. Finishing of surfaces.

Procedure	Paragraph number	*Finish number	Finish material specification
	Aluminum Alloy Parts		
Clean, nonetching alkaline	5.1.3.4	4.9*	P-C-436
	Corrosion Resisting Steel (CRES)		
Clean, passivate	5.1.3.2*	5.5.1*	QQ-P-35 or ASTM-A-380
	All Metal Parts		
Surface treatment, wash primer	5.3*	7.4*(alum) 5.5.1* (CRES)	MIL-C-8514
Prime, epoxy primer	5.3.1*(Alum) 5.3* (CRES)	20.24*	MIL-P-23377, type I (deep yellow)
Top coat, polyurethane	5.3.5*	20.24*	MIL-C-46168 or MIL-C-53039

* MIL-STD-171, Finishing of Metal and Wood Surfaces.

3.7.1 Color. Unless otherwise specified, all surfaces including wheels shall be painted with a lusterless paint system conforming to green color 34094 of FED-STD-595.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.7.1.1 Camouflage color pattern. When specified, the camouflage lusterless color pattern shall consist of contrasting color blotches randomly arranged over a solid ground color to form a jungle or a desert camouflage color pattern (see paragraph 6.2f of FED-STD-595). Four colors shall be used and closely match 30099, 34108, 34258and 37038 of FED-STD-595 for the jungle camouflage color pattern. Six colors shall be used and closely match 30099, 30117, 30277, 30324, 36306, and 37038 of FED-STD-595 for the desert camouflage color pattern.

3.7.2 Film thickness. Unless otherwise specified, the thickness of the dry paint film shall be in accordance with MIL-STD-171, paragraphs 5.2.3 and 5.2.3.1. The powder coating shall be 1.5 to 2.5 ml thick, fast curing and shall provide superior adhesion with a touch mar-resistant surface that shall withstand high impact without chipping or cracking.

3.7.3 Surfaces not to be painted. Surfaces not to be painted shall be designated on the contract drawings.

3.7.4 Stenciling and marking. Unless otherwise specified, the stenciling and marking shall be applied in lusterless white color 37875 per FED-STD-595 on green painted surfaces. Stenciling and marking shall be made with a lusterless enamel conforming to TT-E-527.

3.8 Identification.

3.8.1 Plate. Each carrier shall have a permanently fastened aluminum identification plate according to the applicable drawing and MIL-P-15024. All lettering on the plate shall be at least 1/8 inch (0.125 inch) high, legible, permanent and contain the following information:

Contract number
Name or Registered Trademark of manufacturer
Manufacturer's Code Number
Design Activity Code Number
U.S. (Government Ownership Designation)
Item Name
National Stock Number (NSN)
Serial Number

3.8.2 Patient orientation. On the vertical outward facing surface of the crossbar assembly, a 3/4 inch high, centrally located phrase "HEAD END" shall be stenciled on one assembly and "FOOT END" on the second assembly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.9 Lubrication. The screw rod threads, which are rotated by the handwheel of the clamp assembly, shall be coated with grease conforming to MIL-G-21164.

3.10 Workmanship. All rough and sharp edges, burrs, and slivers shall be removed. Metal forming shall be done by methods that will insure uniformity of size and shape. Shearing and clipping shall be done accurately and cleanly on all portions of the work and shall be neatly finished. All bolt and machine screw holes shall be accurately and cleanly punched or drilled. Holes for machine screws and bolts shall be sized to provide proper clearance for the size screw or bolt used. Bolts, nuts, and screws shall be drawn tight unless otherwise specified.

All welds shall be produced by qualified welding operators with satisfactory welding equipment. Surfaces to be welded shall be clean and free of grease, oil, oxides, or other foreign material. All weld joints shall blend into adjacent metal in gradual smooth curves. Welds shall be smooth and free of slag, excessive undercut, or excessive splatter.

All coats of finish materials shall be applied smoothly without excessive buildup, runs, wrinkles, orange peel, or surface irregularities.

When accepted by the Government, carriers shall be free of scratches, dents, gouges, and other appearance blemishes and defects.

4. QUALITY ASSURANCE PROVISIONS.

4.1 Product for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Records. The contractor shall maintain records for each lot of all required inspections, examinations, and tests performed on items in that lot to ensure quality of the delivered product. These records shall allow identification and traceability of the lot numbers of raw materials used and of manufacturing and assembly processes, including final packaging.

4.1.2.1 Retention of records. Records of inspections, examinations, and tests performed by or for the contractor specified by this section shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, dimensional, and functional investigation with the use of special laboratory appliances or procedures) and testing (determination by laboratory and technical investigation, destructive and nondestructive, the physical, electrical, and chemical properties) of materials, components, product, and packaging to determine compliance with this specification.

4.1.4 Test equipment and inspection facilities. Test and measuring equipment and inspection facilities of sufficient accuracy, quality, and quantity to permit performance of the required inspection shall be established and maintained by the contractor. The establishment and maintenance of a calibration system to control the accuracy of the measuring and test equipment shall be in accordance with MIL-C-45662 or alternate system acceptable to the Government.

4.1.4.1 Calibration system, alternate. Prior to beginning construction of the carriers, a contractor may submit and request approval of an alternate calibration system to be used in place of a system that meets MIL-C-45662 requirements. The alternate

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

system shall be adequate to control the accuracy of measuring and test equipment and be acceptable to the Government.

4.2 Classification of inspections. The inspections specified herein are classified as follows:

a. Materials inspection (see paragraph 4.3).

b. First article inspection (see paragraph 4.5).

c. Quality conformance inspections (see paragraph 4.6).

d. Inspection of packaging (see paragraph 4.7).

4.3 Materials inspection. Materials inspection shall consist of certification supported by verifying data that the materials listed in Table III and used in fabricating the carriers are in accordance with the applicable referenced specifications or requirements prior to such fabrication (see paragraph 4.8.3).

TABLE III. Materials inspection.

Material	Requirement Paragraph	Applicable Specifications
Aluminum	3.2.1.1	QQ-A-250/11
	3.2.1.2	QQ-A-200/8
	3.2.1.3	QQ-A-200/9
Steel, corrosion resisting	3.2.2.1	QQ-S-766
	3.2.2.2	QQ-W-423
	3.2.2.3	QQ-S-763
Steel, carbon	3.2.3	MIL-S-7097
Tape, nylon webbing	3.2.4	MIL-T-5038
Screws, machine, CRES	3.2.5.1	FF-S-92
Screws, self-tapping, CRES	3.2.5.2	FF-S-107
Compound, cleaning	3.6.1.4	P-C-436
	3.7	P-C-436
Treatment, passivation	3.7	QQ-P-35

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE III. Materials inspection. - Continued.

Material	Requirement Paragraph	Applicable Specifications
Primer, wash	3.7	MIL-C-8514
Primer, epoxy	3.7	MIL-P-23377
Coating, polyurethane, chemical agent resistant	3.7	MIL-C-46168

4.4 Inspection conditions. Unless otherwise specified herein, or in the individual specification, all measurements and tests shall be made at temperatures of 15°C to 35°C (59°F to 95°F) at air pressure of 650 to 800 millimeters of mercury, and relative humidity of 45 percent to 75 percent. Whenever these conditions must be closely controlled in order to obtain reproducible results; for referee purposes, temperature, relative humidity, and atmospheric pressure conditions of 25° +0° -2°C (77.0° -3.6°F), 50 percent ±2 percent, and 650 to 800 millimeters of mercury, shall be specified.

4.4.1 Permissible temperature variation in environmental chambers. When chambers are used, specimens under test shall be located only within the working area defined as follows:

a. Temperature variation within working area: The controls for the chambers shall be capable of maintaining the temperature of any single reference point within the working area within ±2°C (3.5°F).

b. Space variation within working area: Chambers shall be so constructed that, at any given time, the temperature of any point within the working area shall not deviate more than 3°C (5.4°F) from the reference point, except for the immediate vicinity of specimens generating heat.

4.4.2 Reference conditions. Reference conditions as a base for calculations shall be 25°C (77.0°F) for temperature, or an alternate temperature of 20°C (68.0°F), 760 millimeters of mercury for air pressure, and the relative humidity of 50 percent.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

4.5 First article inspection. First article inspection shall be performed by the contractor, after award of contract and prior to production, at a location acceptable to the Government. First article inspection shall be performed on pre-production sample units that have been produced with equipment and procedures normally used in production. First article approval is valid only on the contract under which it is granted unless extended by the Government to other contracts.

4.5.1 Sample size. Three carriers shall be subjected to first article inspection.

4.5.2 Inspection routine. The pre-production sample units shall be subjected to the inspections specified in Table IV.

TABLE IV. Group A inspection.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	Minor
		Before Subassembly		
Leg 30027FD				
27.37 length	3.4	4.8.1		4.0
Handle 20083FD				
8-1/8 width	3.4	4.8.1		4.0
5 height	3.4	4.8.1		4.0
3 R.	3.4	4.8.1		4.0
Brace 30022FD				
20.00 width	3.4	4.8.1		4.0
16.25 height	3.4	4.8.1		4.0
3 R. (2 places)	3.4	4.8.1		4.0
Support 20082FD				
14.000/13.970 length	3.4	4.8.1		4.0
Gusset 20067FD				
3R.	3.4	4.8.1		4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - Continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	Minor
Before Subassembly				
Knuckle, Male 20068FD				
0.310/0.305 Blade width	3.4	4.8.1		4.0
0.875 location of 0.320/0.315 hole	3.4	4.8.1		4.0
0.315/0.320 hole diameter	3.4	4.8.1		4.0
0.005/0.000 clearance to I.D. of 30022FD	3.4	4.8.1		4.0
Knuckle, Female 20069FD				
0.323/0.318 slot 0.875 location of 0.320/0.315 diameter hole	3.4	4.8.1		4.0
0.320/0.315 diameter hole	3.4	4.8.1		4.0
0.005/0.000 clearance to I.D. of 30022FD	3.4	4.8.1		4.0
Spacer, Small 20071FD				
0.861/0.846 length	3.4	4.8.1		4.0
Channel, Narrow 30024FD				
10.312 location of hole to slot	3.4	4.8.1		4.0
1.000/0.985 width	3.4	4.8.1		4.0
0.635 location of 0.320 diameter hole	3.4	4.8.1		4.0
0.44 R.	3.4	4.8.1		4.0
0.320 R	3.4	4.8.1		4.0
Stem 20059FD				
0.374/0.373 shank diameter	3.4	4.8.1		4.0
0.625 head diameter	3.4	4.8.1		4.0
0.925/0.915 location of 0.127/0.125 hole	3.4	4.8.1		4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - Continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	Minor
	Before Subassembly			
Stem 20059FD cont'd.				
0.127/0.125 diameter	3.4	4.8.1		4.0
0.375 +/- 0.005 head height	3.4	4.8.1		4.0
0.380/0.375 diameter spherical	3.4	4.8.1		4.0
0.192/0.187 depth of 0.380/0.375 diameter spherical	3.4	4.8.1		4.0
0.500 R of head	3.4	4.8.1		4.0
Housing, Pin, Weldment 20076FD				
1.762 length of housing	3.4	4.8.1		4.0
1.637 depth of 0.600/ 0.580 I.D. of housing	3.4	4.8.1		4.0
0.600/0.580 I.D. of housing	3.4	4.8.1		4.0
0.277/0.262 depth of 0.760/0.750 c'bore in housing	3.4	4.8.1		4.0
0.125 location of 0.127/0.125 hole in housing	3.4	4.8.1		4.0
0.127/0.125 diameter hole in housing	3.4	4.8.1		4.0
Spring, Torsion, Right 30033FD				
6-1/4 length	3.4	4.8.1		4.0
3-3/8 length of long leg	3.4	4.8.1		4.0
0.63 length of short leg	3.4	4.8.1		4.0
0.75 length of short leg tip	3.4	4.8.1		4.0
15° angle of short leg	3.4	4.8.1		4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
Before Subassembly				
Spring, Torsion, Left 30032FD				
6-1/4 length	3.4	4.8.1		4.0
3-3/8 length of long leg	3.4	4.8.1		4.0
0.63 length of short leg	3.4	4.8.1		4.0
0.75 length of short leg tip	3.4	4.8.1		4.0
15° angle of short leg	3.4	4.8.1		4.0
Holder, Spring 11044FD				
1.12 length	3.4	4.8.1		4.0
Pin, Quick Release 92384A016 (Mfr P/N)				
1-1/2 grip length	3.4	4.8.1		4.0
3/16 shank diameter	3.4	4.8.1		4.0
Operates smoothly	3.4	4.8.1		4.0
Surface finish free of defects	3.4	4.8.1		4.0
0.127/0.125 hole	3.4	4.8.1		4.0
Stop, Screw 20053FD				
0.525 overall length	3.4	4.8.1		4.0
0.420 length of body	3.4	4.8.1		4.0
0.605/0.595 flange OD	3.4	4.8.1		4.0
0.495 body diameter	3.4	4.8.1		4.0
0.314/0.312 inside diameter	3.4	4.8.1		4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
Before Subassembly				
Bushing, Axle 20062FD				
0.633/0.630 inside diameter	3.4	4.8.1		4.0
1.000 outside dimension to fit 20060FD	3.4	4.8.1		4.0
0.06 x 45° chamfer	3.4	4.8.1		4.0
Spacer, Large 20090FD				
1.000/0.985 length	3.4	4.8.1		4.0
Axle 20066FD				
7.06 overall length	3.4	4.8.1		4.0
3.540/3.530 location of groove	3.4	4.8.1		4.0
0.488/0.485 diameter of small groove	3.4	4.8.1		4.0
0.049/0.046 width of groove	3.4	4.8.1		4.0
2.797 location of large groove	3.4	4.8.1		4.0
0.250 diameter of large groove	3.4	4.8.1		4.0
0.312 width of large groove	3.4	4.8.1		4.0
0.6249/0.6244 outside diameter	3.4	4.8.1		4.0
0.03 x 45° chamfer, two places	3.4	4.8.1		4.0
Presence of 1/4-28 UNF-2B thread	3.4	4.8.1		4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
	Before	Subassembly		
Frame, Head, Assembly 60010FD	3.4	4.8.1		4.0
Brace perpendicular to crossbar 0.060	3.4	4.8.1		4.0
Braces parallel 0.060	3.4	4.8.1		4.0
Assembly flat 0.060	3.4	4.8.1		4.0
16-1/4 brace spacing	3.4	4.8.1		4.0
1-7/8 crossbar location	3.4	4.8.1		4.0
1-3/16 location of spring guide, two places	3.4	4.8.1		4.0
8-1/8 location of spring guide	3.4	4.8.1		4.0
Location of 6-32 UNC-2B thread	3.4	4.8.1		4.0
Presence of stem and grooved pin	3.4	4.8.1		4.0
Surface preparation as required for welding	3.6.1.4	4.8.1	2.5	
Welds as specified for visual examination	3.6.1.1	4.8.1		4.0
Welds as specified for mechanical properties	3.6.1	4.8.2	2.5	
Plug welds (3/8 inch diameter) withstand shear load of 2500 pounds	3.6.1.2	4.8.2	2.5	
	Before	Subassembly		
Check dimensions and alignment of all heat-treated assemblies prior to hardening phase	3.6.1.3	4.8.1	2.5	

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
	Before	Subassembly		
Frame, Foot, Weldment 60011FD				
Brace perpendicular to crossbar 0.060	3.4	4.8.1		4.0
Braces parallel 0.060	3.4	4.8.1		4.0
Crossbars parallel	3.4	4.8.1		4.0
Assembly flat 0.060	3.4	4.8.1		4.0
16-1/4 brace spacing	3.4	4.8.1		4.0
1-7/8 crossbar location	3.4	4.8.1		4.0
1-3/16 location of spring guides, two places	3.4	4.8.1		4.0
8-1/8 location of spring guide	3.4	4.8.1		4.0
Location of 6-32 UNC-2B thread	3.4	4.8.1		4.0
1.250 location of axle housing assembly	3.4	4.8.1		4.0
2.437 location of pin housing weldment	3.4	4.8.1		4.0
Free operation of pin in pin housing	3.5.2	4.8.1		4.0
Surface preparation as required for welding	3.6.1.4	4.8.1	2.5	
Welds as specified for visual examination	3.6.1.1	4.8.1		4.0
Welds as specified for mechanical properties	3.6.1.1	4.8.2	2.5	
Plug welds (3/8 inch diameter) withstand shear load of 2500 pounds	3.6.1.2	4.8.2	2.5	

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
	Before	Subassembly		
Frame, Foot, Weldment 60011FD (contd.)				
Check dimensions and alignment of all heat-treated assemblies prior to hardening phase	3.6.1.3	4.8.1	2.5	
Frame, Leg, Weldment 60012FD				
Support perpendicular to leg 0.060	3.4	4.8.1		4.0
Supports parallel 0.060	3.4	4.8.1		4.0
Legs parallel 0.060	3.4	4.8.1		4.0
Assembly flat 0.060	3.4	4.8.1		4.0
14.0 Leg spacing	3.4	4.8.1		4.0
5-5/8 Support location	3.4	4.8.1		4.0
24-5/8 Support location	3.4	4.8.1		4.0
2.13 Wheel holder spacing	3.4	4.8.1		4.0
4.25 Wheel holder spacing	3.4	4.8.1		4.0
0.62 Wheel holder location	3.4	4.8.1		4.0
25° Wheel holder position	3.4	4.8.1		4.0
3-13/16 Spring holder location	3.4	4.8.1		4.0
5/32 Spring holder location	3.4	4.8.1		4.0
4.0 Handle location	3.4	4.8.1		4.0
0.72 Collar location	3.4	4.8.1		4.0
1.07 Collar spacing	3.4	4.8.1		4.0
Surface preparation as required for				

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

welding	3.6.1.4	4.8.1	2.5	

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
Before Subassembly				
Frame, Foot, Weldment 60012FD (contd.)				
Welds as specified for visual examination	3.6.1.1	4.8.1		4.0
Welds as specified for mechanical properties	3.6.1.1	4.8.2	2.5	
Plug welds (3/8 inch diameter withstand shear load of 2500 lbs.	3.6.1.2	4.8.2	2.5	
Check dimensions and alignment of all heat-treated assemblies prior to hardening phase	3.6.1.3	4.8.1	2.5	
Carrier, Litter, Wheeled 60007FD				
Wheel assembly free fit in long axle housing	3.5.2	4.8.1	2.5	
Wheel assembly free fit in short axle housing	3.5.2	4.8.1	2.5	
Quick release pin free fit in axle housing	3.5.2	4.8.1	2.5	
Quick release pin retains wheel assembly	3.5.2	4.8.1	2.5	

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
	Before	Subassembly		
Carrier, Litter, Wheeled 60007FD (contd.)				
Head frame and foot frame latch in operating position	3.5.2	4.8.1	2.5	
Legs latch in operating position	3.5.2	4.8.1	2.5	
Litter holders accept standard litter (NSN-6530-00-783-7905) Litter holders securely clamp onto standard litter	3.5.2	4.8.1	2.5	
Maximum of six pound-inches of torque operates hand wheel through full range of movement	3.5.2	4.8.4	2.5	
Straps secure carrier in folded position	3.5.2	4.8.1		4.0
Carrier, Litter, Wheeled-Workmanship				
Sheet Metal: Rough or sharp edges, burrs, and slivers	3.10	4.8.1		4.0
Shearing and clipping inaccurate & ragged	3.10	4.8.1		4.0
Bolt, screw, and rivet holes not clean and round, inaccurately located, wrong				

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

size	3.10	4.8.1		4.0

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	Minor
	Before	Subassembly		
Carrier, Litter, Wheeled-Workmanship (contd.)				
Arc welding:				
Excessive undercut, surface cracks, cavities inclusions, welds ground, rough	3.6.1, 3.6.1.1, 3.6.1.2, 3.10	4.8.1	2.5	
Riveting:				
Pull stem sticking out of head, rivets not tight, poorly formed heads, flat rivets not flush or below surface	3.10	4.8.1		4.0
Assembly:				
Improper assembly, loose fasteners, scratches, dents, gouges, chips and appearance defeats	3.10	4.8.1		4.0
Finish:				
Materials not as specified	3.7	4.8.3	2.5	
Process & sequence not as specified	3.7	4.8.1	2.5	
Film thickness less than specified	3.7.2	4.8.2		4.0
Color not as specified	3.7.1	4.8.1		4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE IV. Group A inspection. - continued.

Inspection	Requirement paragraph	Test method paragraph	AQL (Percent Defective) Major	AQL (Percent Defective) Minor
	Before	Subassembly		
Carrier, Litter, Wheeled-Workmanship (contd.)				
Moving parts bind due to paint. Excessive paint buildup, runs, wrinkles, orange peel, over spray, skips, misses, & other surface defects.	3.7, 3.10	4.8.1		4.0
Camouflage color	3.7.1.1	4.8.1		4.0
Identification and marking:				
Identification plate omitted or not as specified	3.8	4.8.1		4.0
Instructions omitted or not as specified	3.8	4.8.1		4.0
scratches, dents,				
Technical manual	3.5.3	4.8.1		4.0

4.5.3 Failures. One or more failures shall be cause for refusal to grant first article approval.

4.6 Quality conformance inspection.

4.6.1 Inspection of product for delivery. Inspection of product for delivery shall consist of group A inspection.

4.6.1.1 Inspection lot. An inspection lot shall consist of all carriers covered by this specification produced under essentially the same conditions and offered for inspection at one time.

4.6.1.2 Group A inspection. Group A inspection shall consist of the inspections specified in Table IV, in the order shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

4.6.1.2.1 Sampling plan. Unless otherwise authorized, the sampling plan and sample size shall be in accordance with ANSI/ASQC Z1.4 for general inspection level II. The Acceptance Quality Level (AQL) shall be as specified in Table IV. Major and minor defects shall be as defined in ANSI/ASQC Z1.4.

4.6.1.2.2 Rejected lots. If an inspection lot is rejected, the contractor may rework it to correct the defects, or screen out the defective units, and may resubmit for reinspection. Resubmitted lots shall be reinspected. The level and type of inspection are to be determined by the contracting officer. Such lots shall be separate from new lots and shall be clearly identified as reinspected lots.

4.7 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of Medical Marking Standard No. 1.

4.8 Method of inspection.

4.8.1 Examination. Examine representative samples of materials, components, and product to determine compliance with the requirements, specifications, and standards listed in the appropriate paragraph of sections 3 and 5. See section 4.5.2, Table IV, for inspections that require examination.

4.8.2 Testing. Test representative samples of materials, components, and product to determine compliance with requirements, specifications and standards listed in sections 3 and 5. See section 4.5.2, Table IV, for inspections that require testing.

4.8.2.1 Plug weld tests. Plug welds (opposing 3/8 inch diameter) used in the litter holder assembly shall be tested to determine compliance in paragraph 3.6.1.2.

4.8.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material, may be furnished in lieu of actual testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government and provided also that the test methods and results obtained comply with those described in this specification. The certificate shall include the name of the contractor, contract number, name of the manufacturer or supplier, manufacturer's identification of the component or the material, name of the

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

component or material, lot number, lot size, sample size, date of testing, test method, and individual test results.

4.8.4 Handwheel. Handwheel shall turn free. Six pound-inches, maximum, of torque shall rotate handwheel through full range of movement.

5. PACKAGING

5.1 Preservation. Preservation shall be military, as specified in paragraph 6.1.

5.1.1 Military.

5.1.1.1 Unit. Each (EA.). One carrier, as specified, constitutes one unit. Each carrier shall be wrapped in suitable noncorrosive barrier material; wrap shall be secured with suitable tape. Each wrapped carrier shall be packaged in a box conforming to ASTM D5118, class domestic. Closure shall be as specified in ASTM D1974, closure method 1F.

5.2 Packing. Packing shall be level A or B, as specified in paragraph 6.1.

5.2.1 Level A. Each carrier, packaged as specified in 5.1.1.1, shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-621, class 2 or ASTM D5118, class weather-resistant. Closure shall be as specified in the applicable specification, except closure of containers conforming to ASTM D5118 shall be as specified in ASTM D1974, closure method 1D.

5.2.1.1 Case Liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be water-resistant as specified in ASTM D1974.

5.2.2 Level B. Each carrier packaged as specified in paragraph 5.1.1.1 shall require no further overpacking.

5.2.3 Unitization. Carriers shall be unitized on pallets as specified in the contract or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Carrier. Each carrier shall be marked as specified in paragraphs 3.8.1 and 3.8.2.

5.3.2 Unit container, shipping container and unitized load. Each unit container, shipping container and unitized load shall be marked as specified in Medical Marking Standard No. 1.

6. NOTES.

6.1 National Stock Number Data. The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6530-01-220-7186	Carrier, Litter, Wheeled

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS
GSA-FSS-FCGC

Project Number: 6530-2477

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.