|INCH-POUND|
MIL-C-42068
25 June 1989

MILITARY SPECIFICATION

CHEST, MEDICAL INSTRUMENT AND SUPPLY SET

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers three types of chests. Chests are equipped with two cover assemblies which convert to tables through the use of folding legs.

1.2 Classification. Chests covered by this specification shall be of the following types:

Type I - Patient treatment and linen, with three compartments.

Type II - Medicine and instrument, with six drawers.

Type III - Medical instrument and supply set, with 9 fixed compartments.

|Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.|

AMSC/NA FSC 6545

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2).

SPECIFICATIONS

Federal

QQ-A-200/9	Aluminum Alloy 6063, Bar, Rod, Shapes, Tube, and Wire Extruded
QQ-A-250/8	Aluminum Alloy 5052, Plate and Sheet
QQ-A-250/11	Aluminum Alloy 6061, Plate and Sheet
TT-E-491	Enamel: Gloss, Synthetic (For Metal and Wood Furniture)
TT-P-659	Primer Coating and Surfacer; Synthetic, Tints and White (For Metal and Wood Surfaces)
PPP-B-636	Boxes, Shipping, Fiberboard

Military

MIL-P-116	Preservation, Methods of
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible
MIL-P-19834	Plate, Identification, Aluminum Foil, Adhesive Backed, Modification Applied

STANDARDS

Federal

FED-STD-66	Steel: Chemical Composition and Hardenability
FED-STD-151	Metals: Test Methods
FED-STD-595	Colors

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MIL-STD-147	Palletized Unit Loads
MIL-STD-794	Parts and Equipment, Procedures for Packaging of

(Unless otherwise indicated copies of federal and military specifications, standards, and handbooks are available from the Naval Publications and Forms Center, (ATTN: NPODS 5801 Tabor Avenue, Philadelphia, PA 19120-5099.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

DRAWINGS

Defense Personnel Support Center (DPSC)

25031	Chest Assembly, Medicine/Instrument
25032	Chest Assembly, Treatment/Linen
25033	Insert Sub-Assembly
25034	Table Top, Cover, Sub-Assembly
25035	Chest Assembly, Utensil

(Copies of drawings are available from the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 So. 20th Street, Philadelphia, PA 19101-8419.)

2.2 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawings (with bills of material) shown in Section 2 of this specification, and shall be capable of meeting the requirements shown therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in 4.4.2. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and suitable for the purpose intended.

3.1.1. Aluminum.

3.1.1.1 Sheet. Aluminum sheet of alloy 6061-T4 shall conform to QQ-A-250/11. Aluminum sheet of alloy 5052-H32 shall conform to QQ-A-250/8.

3.1.1.2 Tubing and angle. Aluminum alloy 6063-T5 shall be used for tubing and angles, and shall conform to QQ-A-200/9.

3.1.2 Corrosion-resisting steel. Corrosion-resisting steel shall be AISI 301 and 303 and shall comply with the AISI 300 series of FED-STD-66.

3.2 Style, design and construction. Style, design and construction shall be in accordance with the Defense Personnel Support Center (DPSC) Drawings referenced for each component. Each chest shall be supplied with two cover assemblies that can be used as tables through the use of folding legs. The completely assembled units shall be lightweight, and suitable for use with, and functionally compatible with the MUST system. The units shall be readily movable, simple to maintain and easy to clean. The drawers shall slide easily and smoothly.

3.2.1 Type I (Patient Treatment/Linen). The type I, Patient treatment/linen chest shall be furnished with 2 recessed carrying handles, 2 non-adjustable shelves, and 2 removable cover assemblies, and shall be in accordance with DPSC drawings No. 25032, 25033, 25034, and applicable Bill of Materials. The chest weight with shelves and covers shall not exceed 67 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.2.2 Type II (Medicine and Instrument). The type II, Medicine and Instrument Chest shall be furnished with two recessed carrying handles, six drawers, two removable cover assemblies, and shall be in accordance with DPSC drawings No. 25031, 25033, and 25034, and applicable Bill of Materials.

The second drawer from the top shall have a built-in device for securing narcotics and other sensitive items. The chest weight with drawers and cover shall not exceed eighty-eight pounds.

3.2.3 Type III (Medical Instrument and Supply Set). The type III, Medicine Instrument and Supply Set Chest shall be furnished with two recessed carrying handles, nine fixed compartments and two removable cover assemblies, and shall be in accordance with DPSC drawings No. 25033, 25034, and 25035 and applicable Bill of Materials. The chest weight with compartments and covers shall not exceed seventy-five pounds.

3.3 Chest outer covers. Chest outer covers (parts #3) shall be fabricated of either one or two pieces of 0.080 inch thick 6061-T4 aluminum welded with a continuous square-groove butt weld. A back-up strip may be used to facilitate welding. Welds to be ground and finished smooth to blend with adjacent surfaces.

3.4 Drawer glides (Type II chest). The drawer glides on the type II chest shall be coated with nylon, 0.10 inch thick, applied by the fluidized bed process, as indicated on sheet 2 of drawing No. 25031. Necessary holes shall be drilled after the coating process.

3.5 Finish. After welding has been completed, but before final assembly, all exposed metal surfaces of the chest except the drawer glide bottoms shall be furnished in hospital green color No. 14241 of Federal Standard No. 595. All surfaces shall be thoroughly cleaned of dirt, grease, welding flux and loose scale. Before painting, chests shall be primed with one coat of primer per TT-P-659, and two coats of enamel, class B, per TT-E-491. Brackets for the table legs shall be zinc plated prior to painting. The resultant finish shall be hard, smooth and highly resistant to abrasion, chipping, cracking, acid, alkali, and solvents.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

3.6 Welding. Welding shall be done by qualified welders using appropriate equipment in a qualified process. Spot welds shall be capable of developing the tensile strength of the parent metal. All exposed spot welds shall be smoothed sufficiently to blend with the adjacent surfaces. All welds shall be finished in a clean and workman-like manner. All welds shall be clean, nonporous, sound, ground and finished smooth to match adjacent surfaces. All welds shall be free from pitmarks and imperfection. All surfaces shall be free from sharp edges. Other welding shall be continuous on all surfaces, using welding rod of composition similar to parent metal, and suitable for the purpose intended.

3.7 Watertightness. The finished chest shall be tested as specified in 4.4.1. There shall be no entry of water into the chest.

3.8 Identification marking. Identification marking shall be permanently stamped on the chest or a metal nameplate shall be firmly attached to the chest as shown in applicable drawing. The following shall be permanently and legibly marked on the nameplate:

National Stock Number
Item Identification
Contract or Purchase Order Number
Manufacturer's Name or Registered Trademark

A nameplate conforming to type I of MIL-P-19834 shall be considered acceptable. Other types of metal nameplate shall be spotwelded to chest.

3.9 Workmanship. Workmanship shall be first class throughout. The complete chest and component parts shall be clean, well made, and free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.4 Certificates of quality. Certificates of quality supplied by the manufacturer of the component or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for examination purposes shall be one chest.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

Type of Defect	Inspection Level*	AQL (Percent Defective)
For visual examination		
Major defects	S-3	1.0
Minor defects	S-3	2.5
For dimensional examination	S-2	4.0

*Minimum sample size shall be three (3).

4.2.2 For test. Sampling for precipitation test (see 4.4.1) shall be conducted in accordance with MIL-STD-105. The unit of product for examination purposes shall be one chest per lot and the acceptance number shall be zero.

4.3 Examination. The chests shall be examined for defects including, but not limited to, those listed in table II.

TABLE II. Classification of defects.

Catagories	Defects
Major	
101	Covers do not seat or latch properly.
102	Drawers do not move smoothly in and out.
103	Covers not able to fold into tables as required.
104	Drawer stops not positive so each drawer must be lifted up to remove it from the frame.
105	Does not rest evenly on flat surface.
106	Lock drawer and lock not operating correctly.
107	Any component missing or not as specified.
108	Materials non-conforming to specification.
Minor	
201	Finish not free of burrs, scratches, dents.
202	Chest not free of sharp edges.
203	Identification marking not present, complete, legible, correct.
204	Shelves not level.
205	Welds not ground smooth.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

4.3.1 Dimensional examination. The chests and components shall be examined for defects in dimensions. Any dimension not within the specified tolerances shall be classified as a defect.

4.4 Tests.

4.4.1 Precipitation test. The chest shall be placed in a shower stall and subjected to simulated driving rain from a shower head. The chest shall be rotated so that the shower hits the chest from all sides. Water temperature shall be warm and continue for a period of at least fifteen minutes. The chest shall be dried on the exterior and opened. The inside of the cabinets shall be entirely free of water after exposure to the shower.

4.4.2 Chemical analysis of the metals. Four ounces of each lot of each metal used in the construction of the chests shall be analyzed in conformance with FED.STD.No. 151.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Each chest shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservation applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each chest shall be assembled, with the pressure equalizer valve remaining open. The assembled chest shall be packaged in accordance with MIL-P-116, method III, in a box conforming to PPP-B-636, type CF, class weather-resistant. Bursting strength of box shall be in accordance with special requirements of table II of PPP-B-636. Chest shall be cushioned within the box to prevent damage during shipment and storage. Closure, strapping and waterproofing shall be as specified in the box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for each chest shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Each chest packaged as specified in 5.1.1.3 shall require no further overpacking.

5.2.2 Level B. Each chest packaged as specified in 5.1.1.3 shall require no further overpacking, except box shall conform to PPP-B-636, type CF, class domestic, and bursting strength of box shall be in accordance with special requirements of table I of PPP-B-636.

5.2.3 Level C. Each chest shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantity for shipment to one destination exceeds 250 lbs (excluding the pallet) or 20 cubic feet. Loads shall be unitized as specified in MIL-STD-147, on type IV or type V 4-way entry pallets. Pallet shall have a length of 40 inches and a width of 48 inches. Pallet loads, including the pallet, shall not exceed 54 inches in height, 43 inches in length and 52 inches in width. Quantity for shipment to one destination of less than 250 lbs or 20 cubic feet need not be palletized.

5.2.5.1 Levels A and B. Chests packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. Chests packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

5.3.1 Unit. Each chest shall be permanently marked as specified in 3.8.

5.3.2 Levels A, B, and C. Each exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2 and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6.1 Intended use. The medical instrument and supply chests are suitable for use with the Medical Unit, Self-contained, Transportable (MUST). This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specsification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.2 Acquisition requirements.

a. Title, number, and date of the specification.

b. Issue of DODISS to be cited in the solicitation and if required, the specific issue of individual documents referenced (see 2.1).

c. Type of chest required (see 1.2).

d. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

e. National Stock Number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

6.3 National Stock Number (NSN) coverage. This specification covers the following items shown in the Federal Supply Catalog:

NSN	Item Identification
6545-00-118-6248	Chest, Medical Instrument and Supply Set, Type I.
6545-00-118-6249	Chest, Medical Instrument and Supply Set, Type II.
6545-00-131-6977	Chest, Medical Instrument and Supply Set, Type III.

6.4 Subject term (key word) listing.

Care kit
Compartmented table
Cupboard
Kit
MUST chest
Treatment kit

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

DOD - MB

Agent:

DLA-DM

Civil Agency Coordinating Activities:

FDA-MPQAS
USPHS
VA

Project No. 6545-0062

Location:
MILSP/CHEST/MED/INST/H13-34

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
b ADDRESS *(Street, City, State, ZIP Code)*	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE) and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:19Z
Check the source to verify that this is the current version before use.