Inch/Pound

MIL-C-42074
15 February 1991

MILITARY SPECIFICATION

CASE, SPECTACLE FITTING AND ADJUSTMENT SET

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE.

1.1 This specification covers a carrying case designed to contain the components of the spectacle fitting and adjustment set for the cellulose acetate frames.

2. APPLICABLE DOCUMENTS.

2.1 Government Documents.

2.1.1 Specifications, Standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation (see 6.2)

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel DPSC-MST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

L-P-390	Plastic Molding Material, Polyethylene, Low and Medium Density.
MMM-A-181	Adhesive, Phenol, Resorcinol, or Melamine Base.
MMM-A-250	Adhesive, Water-resistant, (For Sealing Fiberboard Boxes).
PPP-B-566	Boxes, Folding Paperboard.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liner, Case, and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible.
MIL-P-26514	Polyurethane Foam, Rigid or Plastic, For Packaging.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MIL-STD-147	Palletized Unit Loads.
MIL-STD-794	Parts and Equipment, Procedures for Packaging of.

(Unless otherwise indicated copies of federal and military specifications, standards, and handbooks are available from the Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia, PA 19111-5094.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

24062 Case, Spectacle Fitting and Adjustment Set: Plastic

(Application for copies should be addressed to the Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-MST, 2800 S. 20th Street, Philadelphia, PA 19101.)

2.2 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted are those listed in the issue of the DoDISS cited in the solicitation. Unless otherwise specified, the issues of documents cited in the solicitation (see 6.2).

ASTM STANDARDS

ASTM D1788 - Rigid Acrylonitrile-Butadiene-Styrene (ABS) Plastics

(Application for copies should be addressed to the ASTM, 1916 Race St., Philadelphia, PA 19103).

(Non-Government standards and other publications are normally available from the organizations that prepare or distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.2 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Materials. Except as specified herein, material shall conform to the applicable portions of the specifications, and standards specified herein, and shall be capable of meeting the requirements shown therein. When materials are not specifically designated, those used shall be of the highest commercial quality, and entirely suitable for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

3.1.1 Plastic for case. Material for the case shall be black acrylonitrile butadiene styrene in accordance with ASTM Standard D1788, ABS 5-2-2. Plastic shall be specially formulated for the manufacturing process used, such as vacuum forming or molding. The finished plastic shall be resistant to the deterioration and damage during use.

3.1.1.1 Plastic inserts. Plastic inserts for compartments and filler shall be composed of polyurethane foam, 1.5 to 2.0 pounds per cubic foot density, in accordance with MIL-P-26514, Type I, Class 2. The foam shall be sufficiently soft and resilient to permit insertion, retention and easy removal of the spectacle components. Color shall be dark gray.

3.1.2 Hinge (with case). Hinge shall be a piano hinge, steel or brass, of a size suitable for the purpose and shall have a bright chromium plated finish.

3.1.3 Rivets. Rivets shall be brass or steel or a size suitable for the purpose intended. They shall have a bright chromium plated finish. As an alternate, rivets for the hinges, latches and feet shall be cadmium or zinc plated.

3.1.4 Handles and feet. The handles and feet for the plastic case shall be injection molded polyethylene, low density, in accordance with specification L-P-390. Hardness shall be shore durometer D50 to D53.

3.1.5 Case hardware. Latches and handle fixtures shall be nickel plated carbon steel. Valances at the mating edges of the lid and body and hinges shall be 6063-T5 aluminum. Rivets shall be nickel plated steel or aluminum. The latches shall be of the single key lockable type.

3.2 Design and construction of the case. The design and construction of the case shall conform to Defense Personnel Support Center, Directorate of Medical Materiel, Drawing No. 24062, and to the requirements specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

3.2.1 Plastic case construction. The plastic case shall be vacuum formed or molded to specified size and shape. Cover and body shall have tongue and groove aluminum valances at mating edges. Two hinges equidistant from each end shall be attached to valances by rivets. Two positive latches with locks shall be attached to the front of the valances equidistant from sides of the case to provide positive closure. All rivets shall be backed with washers to prevent pulling out of the case. The case shall be provided with double handles one each side of valance in order to prevent dropping of items when carried while unlatched. Four plastic feet shall be attached as shown on the drawing. The polyurethane foam insert shall be adhered to the inside surfaces of the case. The insert shall be constructed to retain components in designated position without damage. There shall be twenty four slots to hold the frames erect; three slots to hold pliers; a single slot to hold the screwdriver, rule, and screws.

3.2.2 Design. The design of the interior fittings to maintain components in their individual compartments shall be in such a manner as to insure against shifting and breakage, and shall be so arranged that the components are easily visible and accessible upon opening the lid. The case shall have a smooth grained finish.

3.3 Identification marking. The exterior of the top lid of the case shall be permanently and legibly marked in white paint as follows:

US
6540-01-275-6065
SPECTACLE FITTING AND ADJUSTMENT SET
CELLULOSE ACETATE FRAME
MANUFACTURER'S NAME
CONTRACT NO.

Position and height of the letters shall be as shown on the drawing. The lettering shall be centered on the lid, except for the manufacturers' name and the contract number which shall be placed near the bottom. Height of all letters and spacing between the NSN and Item Identification shall be as shown on the drawing.

3.4 Workmanship. Workmanship shall be first class in all respects. The finished cases shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contractor purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become apart of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.3 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, inspection level II.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Mil-STD-105, inspection level S-1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

4.2.3 Component materials. Sufficient samples of the component materials shall be selected from each lot of raw materials to perform tests as required by the applicable subsidiary specifications and by tests specified herein. No test result shall fall below the minimum or exceed the maximum values specified. The sample unit for testing shall be the quantity necessary to perform all required tests one time each.

4.3 Examination. The case shall be examined to determine compliance with all requirements contained in this specification. Fit of components listed in 3.2.5 shall be determined by the procuring activity. Improper fit-space, too small, or too large for snug fit, shall be cause for rejection of the lot. The examination shall be conducted on items selected in accordance with 4.2.1. Acceptable Quality Levels shall be 1.0 for major defects and 2.5 minor.

4.3.1 Classification of defects. Examination shall be conducted in accordance with but not restricted to the categories and defects in table I. Additional defects shall be classified in accordance with MIL-STD-105.

Table I. Categories and defects

Major	
101	Case is not style and design specified.
102	Dimensions not within tolerance.
103	Mating edges of valance and elements of latches improperly aligned.
104	Any cut, tear, blister or hole.
105	Identification marking missing or illegible.
Minor	
201	Hardware improperly attached.
202	Hardware not plated.
203	Foam insert not securely adhered to inside surface of case.
204	Handles or feet insecurely fastened.
205	Color not as specified.
206	Identification marking wrong size or improperly spaced.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

4.5 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING.

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each case shall be clean, dry and wrapped in draft paper or other suitable material. Each wrapped case shall be further packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for case shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Six unit packages containing case with spectacle fitting and adjustment set, shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Six unit packages containing case with spectacle fitting and adjustment set, shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. Six unit packages containing case with spectacle fitting and adjustment set, shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 54 inches in height, 52 inches in width, and 43 inches in length. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Case with spectacle fitting and adjustment set, shall be packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2. Level C. Case with spectacle fitting and adjustment set, shall be packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

6.1 Intended use. The case is used to contain the components of the spectacle fitting and adjustment set for the cellulose acetate frames.

6.2 Acquisition requirements.

a. Title, number, and date of the specification.

b. Issue of DoDISS to be cited in the solicitation and if required, the specific issue of individual documents referenced (see 2.1 and 2.2).

c. Applicable National Stock Number.

d. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 National Stock Number (NSN) Coverage. This specification covers the following National Stock Number:

National Stock No.	Item Identification
6540-00-455-6827	CASE, SPECTACLE FITTING AND ADJUSTMENT SET

6.4 Subject term (key word) listing.

Case
Adjustment Set
Spectacle Fitting

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6540-1200

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: MIL-P-42074	2. DOCUMENT DATE (YYMMDD): 15 February 1991

3. DOCUMENT TITLE: CASE, SPECTACLE FITTING AND ADJUSTMENT SET

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)

7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:20Z
Check the source to verify that this is the current version before use.