MILITARY SPECIFICATION

SCISSORS, TENOTOMY, SUB-MINIATURE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a scissors, tenotomy, right, Westcott, sub-miniature, 3-3/8 inch (8.6 cm), for use in microsurgical procedures.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-601 - Boxes, Wood, Cleated-plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup

MILITARY

MIL-P-116 - Preservation, Methods of.
MIL-L-10547 - Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

FEDERAL

Federal Test Method Standard No. 151 - Metals; Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.

2.1.2 Drawing. The following drawing forms a part of this specification to the extent specified herein. Unless otherwise specified, the issues shall be those in effect on the date of the solicitation.

DRAWING

Defense Personnel Support Center (DPSC)

22304 - SCISSORS, TENOTOMY, Right, Westcott, Sub-Miniature, 3-3/8 Inch (8.59 cm)

(Copies of specifications, standards and drawing required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM) STANDARDS

E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

E 92 - Vickers Hardness of Metallic Materials, Tests for.

E 140 - Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness and Knoop Hardness).

F 899 - Stainless Steel Billet, Bar and Wire for Surgical Instruments.

F 1089 - Standard Test Method for Corrosion of Surgical Instruments.

F 1078 - Standard Specification for Instruments for Surgery-Terminology Surgical Scissors. Inserted and Noninserted.

F 1079 - Standard Specification for Inserted and Noninserted Surgical Scissors-General Workmanship and Corresponding Test Methods.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race St., Philadelphia, PA 19103.)

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 General. The terms used to describe the components and portions of the scissors described in this specification shall be in accordance with ASTM F1078.

3.2 Material. The scissors shall be manufactured from Class 4, martensitic stainless steel in accordance with ASTM F899 for the class and type used. Composition shall be tested as specified.

3.3 Hardness. The scissors shall have a Rockwell hardness of C-50 to C-55 when tested as specified in 4.5.2.

3.4 Style, design and dimensions. The scissors shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22304.

3.5 Finish. Finish shall be entirely mirror, or satin finish.

3.6 Identification markings The scissors shall be permanently marked with the name or registered trademark of the contractor, and the country or origin. Instruments of foreign manufacture shall also be marked with the word "stainless".

3.7 Resistance to corrosion. Scissors shall show no signs of corrosion or plating of copper sulfate when tested as specified in 4.5.3 and 4.5.4.

3.8 Performance. The scissors shall not bind or snag the test material when tested as specified in 4.5.5.

3.9 Workmanship. Workmanship shall be first class throughout. The scissors shall be free from burrs, cracks, pits, sharp edges, scale, corrosion and other defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become part of the contractor's overall inspection system or quality program The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of assuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling in quality conformance does not authorize submission of known defective material either indicated or actual, nor does it commit the Government to acceptance of defective material.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.2.2 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion-resistant steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

4.3 Sampling.

4.3.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. Unit of product for examination purposes shall be one scissors.

TABLE I. Sampling Criteria for Inspection

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0

4.3.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one scissors.

TABLE II. Sampling Criteria for Tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Composition	3.2	4.5.1	**	*
Hardness	3.3	4.5.2	S-2	1.0
Boil Test	3.7	4.5.3	S-2	1.0
Copper Sulfate	3.7	4.5.4	S-2	1.0
Performance	3.8	4.5.5	S-2	1.0

* Acceptance number shall be zero for all sample sizes.

** Sufficient amount from each lot of each metal used in the construction of the scissors shall be selected for composition analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

4.4 Examination.

4.4.1 Examination of scissors. The scissors shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any scissor in the sample containing one or more defects shall be rejected and if the number of defective scissors in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Scissors not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.
102	Finish of scissors not as specified.
103	Style and design not as specified.
104	Jaw design not as specified.
Minor	
201	Identification markings not present, not correct, not complete, or not legible.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Composition. Raw material shall be tested for composition with Federal Test Method 111 or 112 of Federal Test Method Standard Number 151 (see 4.2.2.)

4.5.2 Hardness test. Hardness shall be determined in accordance with ASTM Standards E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Standard E140.

4.5.3. Boil test. The scissors shall be tested in accordance with the Resistance to Boil Test of ASTM F1089.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

4.5.4. Copper sulfate test. The scissors shall be tested in accordance with the Resistance to Copper Sulfate test of ASTM F1089.

4.5.5 Performance test. The scissors shall be tested in accordance with the scissors performance test of ASTM F1079.

4.6 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. The tenotomy scissors shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each pair of scissors shall be packaged in accordance with MIL-P-116, Method 1C-2.

5.1.1.4 Intermediate package. Twelve unit packages shall be packaged in an intermediate package conforming to PPP-B-566, PPP-B-636, class domestic, or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for tenotomy scissors shall conform to the requirements of MIL-STD-794 for level C.

5.1.2.2 Intermediate package. Twelve unit packages shall be packaged in accordance with the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

5.2.1 Level A. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621 class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Twenty-four intermediate packages (288 units) shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in a exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. The tenotomy scissors packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specififed in 5.2.5.

5.2.5.2 Level C. The tenotomy scissors packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.6.

6. NOTES

6.1 Intended use. The scissors shall be used in microsurgical procedures.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Applicable National Stock Number.

6.3 National Stock Number Coverage. This specification covers the following National Stock Number:

National Stock No.	Item Identification
6515-01-139-8288	SCISSORS, TENOTOMY, RIGHT, WESTCOTT, Sub-Miniature, 3-3/8 Inch (8.6 cm).

6.4 Subject item (key word) listing).

Scissors
Tenotomy scissors
Westcott scissors

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3936

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 - 505-033/92127

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-S-42058	SCISSORS, TENOTOMY, SUB -MINIATURE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
DPSC	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*:
b. ADDRESS *(Street, City, State, ZIP Code)* 2800 S. 20th Street Philadelphia, PA. 19101	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:18Z
Check the source to verify that this is the current version before use.