INCH-POUND

A-A-30085E

June 21, 1993

SUPERSEDING
A-A-30085D
June 9, 1989

COMMERCIAL ITEM DESCRIPTION

SPECIMEN KIT, URINE (MIDSTREAM)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a kit to be used for collecting a sterile urine specimen from male or female without catheterization or attendant.

Salient characteristics. Specimen Kit, Urine, Midstream shall be sterile, disposable, and consists of: translucent, plastic (polypropylene), 118 ml to 133 ml (4 to 4-1/2 oz.) capacity specimen container, with dual graduations of oz. and ml/cc in 1/4 oz. and 10 ml/cc increments molded into the cup; with leakproof screw cap. The cup shall be stable on a flat surface, and shall be free from objectionable odor, sharp edges, deformations, cracks, burrs, holes, or thin areas. Provisions for patient identification either a label or write-on surface that accepts, at a minimum, the patient's name and room number. Each kit includes clearly visible instructions for use.

TYPE I - The kit shall contain two stong and soft tissues, three individually wrapped nonwoven wipes each 3.17 cm x 6.35 cm (1-1/4 inches x 2-1/2 inches) ± 0.32 cm (1/8 inch), single fold, impregnated with moist Povidone Iodine, USP. Shall be packaged in an easy open package which shall maintain sterility until open or damaged.

TYPE II - The kit shall contain sufficient pads or swabs for cleaning and drying the area involved, with ample iodine cleansing solution. Iodophor soap towelettes shall be acceptable. Shall be packaged in an easy open package which shall maintain sterility until open or damaged.

Workmanship. The midstream specimen kit shall be free from defects which detract from its appearance or impair its serviceability.

UNIT. KIT - TYPE I - One specimen kit, as specified, constitues one unit.

PACKAGE - TYPE II - Fifty specimen kits, as specified, constitutes one unit.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Leakage test. The cap shall have a leakproof seal when tested as follows: The cup is to be filled with 4 oz. (118 mL) of water. With the cap placed firmly on the cup, the cup will be held upside down and then placed on its side on a flat surface for 5 minutes. Any evidence of leakage will constitute failure.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Sampling for tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by it subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

NOTE. The following National Stock Numbers are covered by this document.

NATIONAL STOCK NUMBER	NOMENCLATURE	TYPE
6530-00-409-8667	Specimen Kit, Urine, Midstream, Disposable	I
6530-00-835-0941	Specimen Kit, Urine Midstream, Disposable	II

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and Federal Test Standard 191 are available from the General Services Administration Business Service Centers in Boston, MA; New York, NY; Philadelphia, PA; Washington, DC; Atlanta, GA; Chicago, IL; Kansas City, MO; Fort Worth, TX; Denver, CO; San Francisco, CA; Los Angeles, CA; and Seattle, WA.

Copies of Specifications and Standards are also available from Department of Defense Single Stock Point. Special Assistance Desk, telephone (215) 697-2667 or 2179, will provide you adequate information by mail or TeleSpecs mode.

Copies of United States Pharmacopedia (USP) regulations may be obtained from: USPC, Inc., 12601 Twinbrook Parkway, Rockville, MD 20852 USA.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A):

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2405

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.