A-A-30107B
April 23, 1985
SUPERSEDING
A-A-30107A
August 20, 1984

COMMERCIAL ITEM DESCRIPTION

BEDS, ADJUSTABLE (Electric)

The General Services Administration has authorized the use of this commercial item description in preference to VA Specification X554h.

This commercial item description describes a general ward use electric bed.

Salient characteristics: Beds, Adjustable (Electric) with IV pole meets the following general requirements:

Type I, Fixed Patient Control: Mounted as integral part of bed frame or side rails.

Type II, Pendant Patient Control: Operated on isolated low voltage or pneumatically sealed to prevent fluid penetration with connections on each side.

Construction: All steel frame mounted on 5 inch swivel casters, two with tread lock brakes (one at head and one at foot end) with head and footboards easily removed without the use of tools. Head board to accommodate a swinging arm type trapeze assembly or bed be provided with a trapeze support bracket capable of supporting a 275 pound (123.75 kg) load without permanent distortion. Bed shall have six IV mounts (3 on each side) located at head, seat and footend of frame. Bed equipped with chromium plated, epoxy coated or nylon side rails firmly attached to the main frame, easily adjustable and fastened securely in all attainable positions, minimum length of 54 inches (137.16 cm). Perforated or solid steel pan with mattress stop at foot end to accommodate an 80 x 36 inch (203.2 x 91.44 cm) mattress and capable of supporting a 300 pound (135 kg) load placed on any corner without permanent deformity of any part of the bed. Shielded motors and drivetrains; gauge markings to indicate the extent (angle or verticle distance) to which the bed is inclined; frame and other components secured by continuous welds, steel rivets or bolts, spot or tack welds. Bolt type fasteners to include positive type locking features. Metal surfaces not of corrosion resisting steel coated with an epoxy resin or similar protective coatings, baked enamel prohibited. Protective bumpers to prevent damage or marring of walls. Each bed permanently marked with manufacturer's name, model, serial number and packed to completely enclose bed.

Functions: Independent straight-line Trendelenburg, straight-line reverse Trendelenburg, vascular, Fowler, semi-Fowler, and straight-line high/low attainable electronically without the aid of hand manipulation. Cardiac chair attainable with minimal manual assistance. All patient controls

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

require constant pressure type. Nurse controls at foot end of the bed, constant pressure type with patient lockout feature. Trendelenburg positions easily attainable from foot end of bed. Horizontal position attainable with auxillary manual crank in case of power or motor failure. Fowler and semi-Fowler positions attainable from either side of bed.

Labeling: Instructions for: auxillary manual crank operation-warning instructing the user to remove bed from power source when manually operated; electronically sensitive areas, indication of safety for use: patient controls identifying function and direction of movement.

Safety: All electrical components shall be UL544 listed and meet current NFPA 70 and 76B requirements . Electrical motors incorporate over-current/ thermal protection complying with requirements of UL547. Automatic reset devices should not constitute a hazard to persons. Leakage current not to exceed 100 micro-amps, resistance to ground less than 0.15 ohms. No exposed lift screws and/or motors. Wood or particle board protected with laminate and conform to NFPA 255, Class A Fire Retardant Materials. Electrical components shielded to prevent electro-magnetic interference with life-support systems. All exposed surfaces free from burrs, sharp edges or other defects.

Operation: Quiet and smooth and perform all functions with a 300 pound (135 kg) load. With the matress support frame in a horizontal position the bed shall be capable of lifting an evenly distributed 600 pound (270 kg) load twice.

Endurance: Perform all therapeutic positions with a 300 pound (135 kg) load for a minimum of 2100 complete cycles, accomplished in 5 days maximum.

Dimensions: Maximum length with bumpers - 91 inches (231.14 cm); maximum width including side rails (raised or lowered) 42 inches (106.68 cm); height of mattress support surface in low position 16-1/2 inches to 19-1/2 inches (41.91 cm to 49.53 cm); height of mattress support surface in high position - 28 inches to 32 inches (71.12 cm to 81.28 cm). Minimum height of frame to floor - 6-1/2 inches (16.51 cm) (adaptors or extenders not permitted).

Options: When specified the ordering activity may indicate:

a. Color and style of head and footboard.

b. Type of rail configuration (full, three-quarter, half).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/ suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

PREPARING ACTIVITY:

VA-OSS VAMKC (904-A)
Agent VA-OSS (92B)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.