COMMERCIAL ITEM DESCRIPTION
TUBE, INHALER (OXYGEN)

The General Services Administration has authorized the use of this commercial item description as a replacement for VA Specification X-1327c which is cancelled.

This commercial item description describes a tube for making extended connections between oxygen supply to administration equipment.

Salient characteristics: Tube, Inhaler (Oxygen), plastic, disposable, flexible, green transparent tube, provided with molded resilient female connectors at each end.

Tube

Length - 84 inches (or approx. 213.36 cm)

ID - minimum .130 inch (or approx. .3302 cm)
maximum .167 inch (or approx. .4242 cm)

OD - minimum .220 inch (or approx. .5588 cm)
- maximum .240 inch (or approx. .6096 cm)

Wall thickness
- minimum .038 inch (or approx. .0965 cm)
- maximum .045 inch (or approx. .1143 cm)

Connectors are to be permanently attached to tube. One at each end. The open end of the connector to be:

Length - minimum 1.670 inch (or approx. 4.2418 cm)
- maximum 1.710 inch (or approx. 4.3434 cm)

ID - minimum .195 inch (or approx. .4953 cm)
- maximum .235 inch (or approx. .5969 cm)

Tapered end to have an opening of:
- minimum .375 inch (or approx. .9525 cm)
- maximum .400 inch (or approx. 1.016 cm)

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2024-10-05T08:36Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

PREPARING ACTIVITY:

VA-OSS VAMKC (904-A)
Agent VA-OSS (91A)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.