METRIC

A-A-30116B
April 20, 1993
SUPERSEDING
A-A-30116A
September 8, 1989

COMMERCIAL ITEM DESCRIPTION

DRAPE, SURGICAL (Iodophor)

This commercial item description covers an antimicrobial (iodophor) incise drape for use in surgical procedures.

Salient characteristics. Drape, Surgical, (Iodophor), incise, disposable, sterile, individually packaged in a peel open package, roll folded with liner on the inside of fold, antimicrobial agent (iodophor) shall be visibly present in or on the drape, antimicrobial agent (iodophor) area to be uniformly and centrally dispersed in an acrylate pressure sensitive adhesive that shall uniformly and continually release the antimicrobial agent (iodophor) when placed in contact with the skin, the film shall be made of a strong conformable polyester, liner release 60 grams per inch maximum per Test Standard PSTC-4 (8/89).

Size 1 - 90 cm x 45 cm with 60 cm x 45 cm (iodophor) antimicrobial adhesive area.

Size 2 - 90 cm x 85 cm with 60 cm x 85 cm (iodophor) antimicrobial adhesive area.

Plus or minus 2.70 cm tolerance for all dimensions.

Workmanship. The surgical drape shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG) One package containing 10 individually packaged sterile drapes, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105D (Normal), with an AQL of 1.0 (percent defective) and a special inspection level of S-4. The unit of product for examination purposes shall be one surgical drape.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus resnipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

NOTE. The following National Stock Number(s, are covered by this document:

NSN	SIZE	NOMENCLAIURE
6530-01-155-6697	Size 1 -	DRAPE, SURGICAL (Iodophor)
6530-01-199-1989	Size 2 -	DRAPE, SURGICAL (Iodophor)

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Size

MILITARY INTERESTS:

Military Coordinating Activit:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2402

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.