A-A-30117A
November 27, 1991
~~SUPERSEDING~~
~~A-A-0030117~~
~~December 15, 1986~~

COMMERCIAL ITEM DESCRIPTION

CATHETER, URETHRAL (Teflon Coated)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers various sizes of sterile urethral catheters of the Foley design and teflon coated.

Salient characteristics. Catheter, Urethral, (Teflon Coated), sterile, disposable, individually packaged, material latex with teflon coating, Foley design, valve inflation channel, 2 eyes, round tip, tip length not to exceed 1-5/16 inches (33 mm), self retaining hemostatic balloon type, minimum usable length of catheter 11-1/2 inches (29 cm) (measured between the end of the balloon which fits snuggly against the urethral opening at the junction of the urethra and the bladder, and the "Y" take-off of the balloon supply segment near the catheter forming the "Y" at the distal end or drainage end of the catheter).

The catheter shall be in accordance with all design requirements of American Society for Testing and Materials Standards F 623 for Foley Catheter.

Type I 30 cc balloon size (balloon size to be tested using 40 cc of air or water)

Size 1 - 16 Fr (5.33 mm)	Size 4 - 22 Fr (7.33 mm)
Size 2 - 18 Fr (6.00 mm)	Size 5 - 24 Fr (8.00 mm)
Size 3 - 20 Fr (6.67 mm)	Size 6 - 26 Fr (8.67 mm)

Type II 5 cc balloon size (balloon size to be tested using 10 cc of air or water)

Size 1 - 14 Fr (4.67 mm)	
Size 2 - 16 Fr (5.33 mm)	Size 5 - 22 Fr (7.33 mm)
Size 3 - 18 Fr (6.00 mm)	Size 6 - 24 Fr (8.00 mm)
Size 4 - 20 Fr (6.67 mm)	Size 7 - 26 Fr (8.67 mm)

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:36Z
Check the source to verify that this is the current version before use.

Workmanship. The catheter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. The unit of issue is each.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:36Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

NSN	NOMENCLATURE	TYPE	SIZE
6515-00-762-1867	CATHETER, URETHRAL (Teflon Coated)	I	1
6515-00-762-1872	CATHETER, URETHRAL (Teflon Coated)	I	2
6515-00-762-1873	CATHETER, URETHRAL (Teflon Coated)	I	3
6515-00-762-1874	CATHETER, URETHRAL (Teflon Coated)	I	4
6515-00-728-6861	CATHETER, URETHRAL (Teflon Coated)	I	5
6515-00-762-1877	CATHETER, URETHRAL (Teflon Coated)	I	6
6515-00-828-2464	CATHETER, URETHRAL (Teflon Coated)	II	1
6515-00-762-1850	CATHETER, URETHRAL (Teflon Coated)	II	2
6515-00-762-1851	CATHETER, URETHRAL (Teflon Coated)	II	3
6515-00-762-1852	CATHETER, URETHRAL (Teflon Coated)	II	4
6515-00-728-1857	CATHETER, URETHRAL (Teflon Coated)	II	5
6515-00-762-1858	CATHETER, URETHRAL (Teflon Coated)	II	6
6515-00-762-2463	CATHETER, URETHRAL (Teflon Coated)	II	7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type and size

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6515-5168

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:36Z
Check the source to verify that this is the current version before use.