A-A- 30125A
May 28, 1992
SUPERSEDING
A-A-0030125
January 28, 1987

COMMERCIAL ITEM DESCRIPTION

TUBE, FEEDING, SURGICAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a disposable feeding tube with a weighted tungsten tip.

Salient characteristics. The disposable feeding tube shall be nonsterile and individually packaged to protects its cleanliness. The feeding tube shall be medical grade polyurethane which is biocompatible with body tissue, noncollapsible during aspiration, and lubricated when activated with water. The tube shall be radiopaque with placement markings at 50 and 75 cm (+ or - 2 cm) intervals from the feeding exit port at the distal end. Feeding tube shall be 8 French (2.7 mm), 43 to 45 inches (109.2 to 114.3 cm) long (overall length) terminating with a single feeding exit port and tungsten weight tip. The distal end of the tube shall have a 2.4 to 3.4 gram tungsten weighted bolus which shall be a continuous part of the tubing.

The proximal end of the tube shall be provided with a "Y" irrigation port. One side of the "Y" shall be provided with a universal feeding set connector with a tab for stopping. The other side of the "Y" shall accept a catheter tip syringe and be provided with a tab for stopping.

A pre-inserted flow-through design stylet of stainless spring steel or stainless steel braided material shall be provided. The stylet shall have a blunt end which is designed to prevent the stylet from exiting the feeding port.

Manufacturer's instruction for use of the feeding tube shall be provided with each tube.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Workmanship. The feeding tube shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 10 or 12 individually packaged feeding tubes, as specified, constitutes one unit. The unit shall be stated in the solicitation or order and shall be that normally provided by the manufacturer as commercial practice.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one feeding tube as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

6515-01-232-5138

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Unit and number of feeding tubes in the unit.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

Project No. 6515-5385

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.