INCH-POUND

A-A-30134C
January 8, 1993
SUPERSEDING
A-A-30134B
October 24, 1989

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE (SKIN PREPARATION TYPE)

The General Services Administration has authorized the use of this commercial item description.

This draft commercial item description has not been approved and is subject to change. DO NOT USE FOR ACQUISITION PURPOSES.

This commercial item description describes a surgical pack that is used as a scrub prior to surgery.

Salient characteristics: SURGICAL PACK, DISPOSABLE (SKIN PREPARATION TYPE), shall be sterile. Surgical pack shall consist of two cotton tipped, wooden or plastic rod applicators, each a minimum of 5-3/4 inches (14.6 cm), two winged grip sponges, each a minimum of 3-1/2 inches by 3-1/2 inches (8.9 cm by 8.9 cm), presaturated with povidone iodine scrub solution (minimum .75% USP), sealed in a leakproof plastic/foil bag (leakproof - no solution to spill out of bag); (packaged either one in a bag or both in a bag); two stick sponges, each a minimum of 7-7/8 inches (20 cm) long, presaturated with povidone iodine paint solution (1% USP), sealed in leakproof plastic bag (two stick sponges packaged in one leakproof plastic/foil bag), leakproof - no solution to spill out of bag), two blue blotting towels a minimum of 22 inches by 19 inches; (56 cm by 48 cm) two white absorbent towels a minimum of 17 inches by 35 inches (43 cm by 89 cm); one pair medium (7 to 7-1/2) latex surgeons'/examination gloves. All components packaged in a minimum 26 inches by 26 inches (66 cm by 66 cm) CSR wrap. CSR wrap product provided in a sealed plastic bag.

Workmanship. The surgical packs must be free from defects which detract from their appearance or impair their serviceability

Unit. Each (EA)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one disposable surgical pack as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug, and Cosmetic Act - If the product covered by this document has been determined by the U S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document

NSN	NOMENCLATURE
6530-00-207-6425	SURGICAL PACK, DISPOSABLE SKIN PREPARATION TYPE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420

Copies of Federal Test Method Standard 191 may be obtained from your local General Services Administration.

Copies of United States Pharmacopedia (USP) regulations may be obtained from: USPC, Inc., 12601 Twinbrook Parkway, Rockville, MD 20852 USA

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents.

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

PROJECT NO. 6530-2394

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.