A-A-30136
March 20, 1989

COMMERCIAL ITEM DESCRIPTION

BED, ADJUSTABLE (Manual)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description describes a manually operated home health care type bed for use by beneficiaries in their residence.

Salient characteristics. Bed, Adjustable (Manual) shall be in accordance with the following requirements. Bed frame and other components secured by welds, steel rivets or bolts (hollow pop rivets not permitted). Bolt type fasteners to include positive type locking features. Metal surfaces must be coated with an epoxy resin or similar protective coatings, baked enamel prohibited. All exposed surfaces free from burrs, sharp edges or other defects. Each bed must be permanently marked with manufacturer's name, model and serial number. If bed is shipped in more than one box, the boxes must be banded together to assure the end user will receive a complete bed. Assembly/operation manuals must be included. Bed components to be enclosed in cardboard containers.

Bed weight - maximum weight 350 lbs
Load weight - bed must be capable of lifting an evenly distributed 500 lbs without distortion in any way
Overall bed dimensions - maximum 42" wide by 88" long

Mattress Support style - solid pan, perforated pan or link
Mattress Support dimensions - maximum 36" wide by 80" long

Height of mattress support surface in:
Hi position - 17" to 33"
Low position - 12" to 18"

Functions:
Hi/Low
Fowlers (head raised, knees elevated)

Hand crank - must be provided for manual crank functions of Fowler's position and hi/low position.

Siderail length - two (2) full length, epoxy/nylon coated or chrome plated siderails, firmly attached to main frame with release mechanism

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Casters - four [illegible] swivel casters, two with brake [illegible] brakes (one at head and one at foot end)

Headboard & footboard - high pressure walnut finish laminate must conform to NFPA 255 Class A Fire Retardant Materiels.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 20261O-26T08:36Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking - Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

NSN	NOMENCLATURE
6530-01-292-5628	BED, ADJUSTABLE (Manual)

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government, may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Copies of NFPA 255 may be obtained from the National Fire Protection Association, Batterymarch Park, Quincy, MA 02269.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MD

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPCAS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DY

Custodians:

Army-MD
Navy-MS
Air Force-01

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.