A-A-30142M
January 13, 1992

SUPERSEDING
A-A-30142
August 1, 1989

COMMERCIAL ITEM DESCRIPTION

LANCET, FINGER BLEEDING

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers disposable sterile lancets suitable for administering finger punctures to obtain blood for hematologic studies.

General. The finger bleeding lancets shall be furnished in either Type I or Type II as specified herein.

Salient characteristics.

Type I. Lancet, Finger Bleeding, shall be sterile, disposable, and have a plastic ribbed body with corrosion resisting steel tri-beveled point. The body shall be 0.87 inches long plus or minus 0.05 inches (22.10 mm, plus or minus 1.27 mm) with a nominal diameter of 0.250 inches (6.35 mm). The piercing tip wire shank shall be molded into the body of the lancet, having a maximum wire diameter of 0.031 inches (0.79 mm) and shall extend 0.125 to 0.140 inches (3.18 mm to 3.56 mm) from the lancet body and be tapered to a sharp point. The point shall have a protective twist-off cover. The overall length including the tip (with protective cover removed) shall be 1.00 inch plus or minus 0.05 inch (25.4 mm plus or minus 1.27 mm). Packed 200 per package, 25 packages per case.

Type II. Lancet, Finger Bleeding, shall be fabricated from corrosion resisting steel flat stock and shall be 1-1/4 to 1-3/4 inches long (31.75 mm to 44.45 mm) with a blade tip length not to exceed 0.160 inch (4.06 mm). Design shall provide sufficient surface area and thickness to allow for rigidity and a secure grip. The blade tip may be flat, beveled or tri-beveled to provide a sharp point. Packed 100 per package.

Sterility. The sterility assurance level of the lancet piercing tip shall be $1x10^{-6}$ in accordance with AAMI Guidelines.

Workmanship. The finger bleeding lancet shall be free from defects which detract from it's appearance or impair it's serviceability. The lancet tip shall be straight, sharp and free from burrs or deformation. Plastic shall be smooth and free from flash.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by it subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its contractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

NOTE. The following National Stock Numbers are covered by this document:

NSN	NOMENCLATURE	TYPE
6515-01-135-8497	LANCET, FINGER BLEEDING 5000's	I
6515-01-225-4757	LANCET, FINGER BLEEDING 200's	I
6515-00-431-2890	LANCET, FINGER BLEEDING 100's	I
6515-00-431-2890	LANCET, FINGER BLEEDING	II

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government, may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DOD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS

PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A):

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

PROJECT NO. 6515-5160

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.