A-A-30153
February 28, 1992

COMMERCIAL ITEM DESCRIPTION

MASK, SURGICAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a disposable, nonwoven, 4 pleat, rectangular shape surgical mask.

Salient characteristics.

The mask shall be disposable, fluid resistant, nonwoven glass-free filtration media, blue or green, 4 pleats (2 up and 2 down) with binding tape margins (stitching, welding or gluing to continue along entire length of each opposing edge in which tucks are anchored), rectangular shape 7-1/2" ± 1/8" (19 cm ± 3 mm) long, center width with pleats extended 7" ± 1/2" (18 cm ± 13 mm), 3-1/4" ± 1/4" (83 mm ± 6 mm) wide, 4 tie tapes 7/16" ± 1/16" (11 mm ± 2 mm) wide, 14" to 17" (35 cm to 43 cm) long, no fraying or loose threads, adjustable malleable nose piece inserted inside, centrally located, minimum 1/8" (3 mm) wide, minimum 4-1/4" (11 cm) long, filtering efficiency to be no less than 95% when tested by the method of Greene & Vesley (J. Bacteriol. Vol. 83. no. 3, p. 663 March 1982) or a comparable method. The Delta P shall be no more than 2mm of water pressure when tested by any commercially accepted method. Must meet the flammability standards ASTM D1230-85, Class 2; NFPA, Class 3, and Consumer Product Safety Commission Class 2. The mask shall not cause irritation to the skin, nose or eyes, and absent of objectionable odor, 50's.

Workmanship. The mask, surgical, shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty (50) surgical masks, as specified, constitutes one unit.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	NOMENCLATURE
6515-00-448-5179	MASK, SURGICAL

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6515-5109

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:37Z
Check the source to verify that this is the current version before use.