A-A-30155
February 28, 1992

COMMERCIAL ITEM DESCRIPTION

DRESSING, OCCLUSIVE, ADHESIVE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a transparent adhesive dressing for use as a dressing for peripheral and central I.V catheters or as a wound dressing.

Salient characteristics.

Type I - Dressing, Occlusive, adhesive, sterile, transparent (elongation 500%), consists of a thin (1 mil), flexible, polyurethane backing (provided between two paper liners, an outer frame and inner window), coated with hypoallergenic, water resistant, acrylate adhesive that allows the adhesive surface to be touched together briefly and then pulled apart, impenetrable to water and bacteria, but permeable to moisture vapor and oxygen, packaged in tyvek paper and polyester EVA Laminate File.

Size 1	2-3/8 inches wide, plus or minus 1/32 inch x 2-13/16 inches long, plus or minus 1/16 inch, frame 1/4 inch, plus or minus 1/16 inch (6cm, plus or minus .08 cm x 7 cm, plus or minus .16 cm; frame .64 cm (plus or minus .16 cm) .
Size 2	4 inches wide, plus or minus 1/16 inch x 4-3/4 inches long, plus or minus 1/16 inch, frame 1/2 inch, plus or minus 1/16 inch (10 cm wide, plus or minus .16 cm) by 12 cm long, plus or minus .16 cm; frame 1.27 cm, plus or minus .16 cm.
Size 3	5-15/16 inch wide, plus or minus 1/16 inch by 8 inches long, plus or minus 1/16 inch, frame 5/8 inch, plus or minus 1/8 inch (15 cm wide, plus or minus .16 cm by 20 cm, plus or minus .32 cm long; frame 1.59 cm, plus or minus .32 cm.

Type II - Dressing, Occlusive, shall be a sterile adhesive dressing which shall adhere to skin and shall be non-adherent to wounds. The adhesive dressing shall be transparent, permeable, nonporous, waterproof, elastic and hypoallergenic and shall be suitable for dressing wounds without the need for an additional gauze bandage. The adhesive dressing shall be permeable to gases and to water vapor and shall be impermeable to liquid water and shall retain serous wound exudate fluid. The adhesive dressing shall be a barrier to bacteria and viruses.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

The material shall be polyurethane or polyethylene covered on one side by pressure sensitive adhesive. The adhesive shall be covered by an easily removable facing. The adhesive dressing shall have removable tabs to aid in aseptic application of the dressing.

The adhesive dressing shall have a minimum average permeability of 450 grams of water vapor/square meter/24 hours and shall have a minimum average elasticity of 20 percent in both the length and width directions

Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproduceabile.

Adhesive dressings shall be of the following sizes, once tabs are removed:

Size 4	2 inches minimum width (5.1 Cm), 2-1/2 inches minimum length (6.3 cm)
Size 5	3-7/8 inches minimum width (9.8 cm), 3-7/8 inches minimum length (9.8 cm)

Workmanship. The dressing shall be free from defects which detract from its appearance or impairs its serviceability.

Unit. Package (PG). For 6510-01-101-8936, one package containing 100 individually packaged dressing, as specified, constitutes one unit. For 6510-01-153-4655, one package containing 50 individually packaged dressings, as specified, constitute one unit. Packaging quantities for the VA items shall appear in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one dressing. Sampling shall be inspection level S-4 with an acceptable quality level (AQL) of 2.5 (percent defective).

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordnace with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one dressing. Sampling shall be inspection level S-1 with an AQL of 2.5 (percent defective).

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract. -A-30155

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

TYPE	SIZE	NSN	NOMENCLATURE
I	1	6510-01-171-8905	DRESSING, OCCLUSIVE, ADHESIVE (2 3/8 inches x 2 13/16 inches)
I	2	6510-01-135-4267	DRESSING, OCCLUSIVE, ADHESIVE (4 inches x 4 3/4 inches)
I	3	6510-01-135-4268	DRESSING, OCCLUSIVE, ADHESIVE (5 15/16 inches x 8 inches)
II	4	6510-01-101-8936	DRESSING, OCCLUSIVE, ADHESIVE (2 inches x 2 1/2 inches)
II	5	6510-01-153-4655	DRESSING, OCCLUSIVE, ADHESIVE (3 7/8 inches x 3 7/8 inches)

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type and size

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQA
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6510-0812

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.