A-A-30160
February 28, 1992

COMMERCIAL ITEM DESCRIPTION

BAG, URINE COLLECTION (LEG TYPE)

The General Services Administration has authorized the use of this commercial item description in preference to VA Specification X-1306J.

This commercial item description describes a urinary drainage leg bag to be worn by the patient and used for the collection of urine.

Salient characteristics. The plastic urinary leg bag shall consist of a leakproof bag, individual package must be marked/labeled "Sterile or Sterile Fluid Path" (no strike through or bleeding of the marking/labeling); fitted with a tapered adapter at each end, 2-1/2 inches (6 cm) flutter valve at top; cap for adapter at lower end of bag shall be made of latex to insure proper fit, withstanding shock effect of walking motion and accompaning friction of trouser leg; and provided with two straps (straps are means for attaching to the upper or lower section of leg), two connecting buttons (smooth and nonirritating). Each strap a minimum of 22 inches (56 cm) long, 3/4 inch (2 cm) wide (plus zero, minus 1/16 inch (0.2 cm)), 5 holes at each end - holes starting 1-1/4 inches (3 cm) from each end and 1-1/2 inches (4 cm) apart, on center. Each hole 1/8 inch (0.3 cm) diameter (minus 1/64 inch (0.03 cm) plus 0). Holes to be free and clear. The inside diameter of the fitting shall accept the insertion of a 3/16 inch (0.5 cm) lumen tube a distance of 2-1/2 inches (6 cm) into the bag, without destroying function of the flutter valve. Package/product must be clean and without holes or tears.

SIZE	CAPACITY
Medium	19 to 26 oz. (539 to 737 g)
Large	32 to 40 oz. (907 to 1,134 g)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Workmanship. The urine collection bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. The unit of issue shall appear in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Leakage test. Will be performed as follows:

Fill the bag to capacity and hang for 24 hours. If bag leaks at any time during this 24 hour period, bag is to be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be 5 samples from each shipment as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

SIZE	CAPACITY	NATIONAL STOCK NUMBER
Medium	19 to 26 oz. (539 to 737 g)	6530-00-761-0936
Large	32 to 40 oz. (907 to 1,134 g)	6530-00-761-0932

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type, class, style and size

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

Project No. 6530-2323

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.