A-A-30163

April 30, 1992

COMMERCIAL ITEM DESCRIPTION

STERILIZATION TEST PACK, BIOLOGICAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes disposable biological sterilization test packs suitable for monitoring the performance of steam and ethylene oxide sterilizers.

Salient characteristics. Manufacturers shall certify that the disposable biological sterilization test packs provided by use of this document have been tested and shown to be equivalent to the Association for the Advancemnet of Medical Instrumentation (AAMI) standard biological test pack. These test packs shall be disposable and consist of a self-contained biological indicator preassembled in a package ready for placement in the appropriate sterilizer. A chemical process indicator shall be provided each biological indicator and used to provide immediate proof of processing in steam or ethylene oxide sterilizers as appropriate. Test packs shall be designed for use in monitoring gravity, prevacuum and pulse vacuum steam sterilizers and ethylene oxide sterilizers. Test packs covered by this document are not designed to monitor the performance of liquid sterilization or washer-sterilizer cycles.

Complete and detailed instructions for use of the test pack, biological indicator and incubators shall be supplied with each unit.

After sterilization cycle and cooling, biological indicators shall be removed from the test pack, instructions on handling and placement in incubators are followed, indicators are observed for specific times per instructions for color change, and results are interpreted and recorded as per instructions.

Not less than 9 months of expiration dating shall remain at time of delivery to the Government.

Test packs shall be of the following style and types

[illegible]

Style A - Biological indicator with Bacillus subtilis for testing of ethylene oxide cycles.

Style B - Biological indicator with Bacillus stearothermophilus for testing of gravity, prevacuum and pulse vacuum steam sterilization cycles.

Style C - Biological indicator with Bacillus subtilis and Bacillus stearothermophilus for testing of ethylene oxide and gravity, prevacuum and pulse vacuum steam sterilization cycles.

Type 1 A control is provided with every four test packs, case of 64 test packs.

Type 2 A control is provided with each test pack, case of 50 tests (ethylene oxide) and case of 25 tests (steam).

Type 3 Five controls are provided with case of 20 tests

Type 4 Twenty controls are provided with each case of 40 test packs.

Workmanship The sterilization test packs shall be free from defects which detract from its appearance or impair its serviceability.

Unit Package (PG) One package containing the number of controls and test packs as stated in Types 1 through 5 above, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of zero (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one sterilization test pack as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

NOTE The following National Stock Numbers are covered by this document

Style A, Type 1 - NSN 6530-01-345-3325 (64 tests)
Style A, Type 2 - NSN 6530-01-277-2743 (50 tests)
Style A, Type 3 - NSN 6530-01-345-3326 (20 tests)
Style B, Type 1 - NSN 6530-01-326-6550 (64 tests)
Style B, Type 2 - NSN 6530-01-276-6147 (25 tests)
Style B, Type 3 - NSN 6530-01-290-8958 (20 tests)
Style C, Type 4 - NSN 6530-01-271-9124 (40 tests)

Copies of Federal Standard No 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U. S. Government Printing Office Washington DC 20420

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description.
(b) Style, Type and Unit.

MILITARY INTERESTS

Medical Coordinating Activity

DoD-MB

AGENT

DLA-DM

Custodians

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES

FDA-MPQAS
PHS

PREPARING ACTIVITY

VA-OSS VAMKC (904)
Agent VA-OSS VACO (924)

Project No. [illegible]